

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** POWE02  
**Lab Code:** ACE **SDG NO.:** Q3155  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 09/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113948.D **Time Analyzed:** 05:16

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.742 | 4.643     | 4.843 | 507.230            | 500.000           | 1.4  |
| Aroclor-1016-2       | 4.760 | 4.661     | 4.861 | 525.810            | 500.000           | 5.2  |
| Aroclor-1016-3       | 4.817 | 4.718     | 4.918 | 520.560            | 500.000           | 4.1  |
| Aroclor-1016-4       | 4.936 | 4.837     | 5.037 | 512.640            | 500.000           | 2.5  |
| Aroclor-1016-5       | 5.192 | 5.093     | 5.293 | 540.270            | 500.000           | 8.1  |
| Aroclor-1260-1       | 6.226 | 6.129     | 6.329 | 556.490            | 500.000           | 11.3 |
| Aroclor-1260-2       | 6.417 | 6.319     | 6.519 | 532.460            | 500.000           | 6.5  |
| Aroclor-1260-3       | 6.782 | 6.683     | 6.883 | 530.960            | 500.000           | 6.2  |
| Aroclor-1260-4       | 7.039 | 6.941     | 7.141 | 593.180            | 500.000           | 18.6 |
| Aroclor-1260-5       | 7.283 | 7.185     | 7.385 | 593.790            | 500.000           | 18.8 |
| Decachlorobiphenyl   | 8.668 | 8.570     | 8.770 | 48.670             | 50.000            | -2.7 |
| Tetrachloro-m-xylene | 3.659 | 3.560     | 3.760 | 53.560             | 50.000            | 7.1  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** POWE02  
**Lab Code:** ACE **SDG NO.:** Q3155  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 09/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113948.D **Time Analyzed:** 05:16

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.721 | 4.623     | 4.823 | 496.850            | 500.000           | -0.6 |
| Aroclor-1016-2       | 4.738 | 4.641     | 4.841 | 497.320            | 500.000           | -0.5 |
| Aroclor-1016-3       | 4.913 | 4.816     | 5.016 | 515.340            | 500.000           | 3.1  |
| Aroclor-1016-4       | 4.956 | 4.858     | 5.058 | 504.390            | 500.000           | 0.9  |
| Aroclor-1016-5       | 5.167 | 5.070     | 5.270 | 534.610            | 500.000           | 6.9  |
| Aroclor-1260-1       | 6.193 | 6.096     | 6.296 | 492.560            | 500.000           | -1.5 |
| Aroclor-1260-2       | 6.382 | 6.284     | 6.484 | 485.240            | 500.000           | -3.0 |
| Aroclor-1260-3       | 6.532 | 6.435     | 6.635 | 460.360            | 500.000           | -7.9 |
| Aroclor-1260-4       | 7.001 | 6.903     | 7.103 | 476.490            | 500.000           | -4.7 |
| Aroclor-1260-5       | 7.243 | 7.146     | 7.346 | 472.800            | 500.000           | -5.4 |
| Decachlorobiphenyl   | 8.611 | 8.514     | 8.714 | 51.530             | 50.000            | 3.1  |
| Tetrachloro-m-xylene | 3.650 | 3.551     | 3.751 | 51.760             | 50.000            | 3.5  |