

Surrogate Summary

SDG No.: **Q3155**

Client: **Kleinfelder**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113719.D	PIBLK-PO113719.D	Tetrachloro-m-xyl	1	20	17.0	85		60	140
		Decachlorobiphen	1	20	17.3	86		60	140
		Tetrachloro-m-xyl	2	20	17.0	85		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO113908.D	PIBLK-PO113908.D	Tetrachloro-m-xyl	1	20	18.0	90		60	140
		Decachlorobiphen	1	20	15.8	79		60	140
		Tetrachloro-m-xyl	2	20	17.5	88		60	140
		Decachlorobiphen	2	20	18.0	90		60	140
Q3155-01	COMP-1	Tetrachloro-m-xyl	1	20	19.0	95		32	144
		Decachlorobiphen	1	20	12.7	64		32	175
		Tetrachloro-m-xyl	2	20	18.3	92		32	144
		Decachlorobiphen	2	20	15.0	75		32	175
Q3155-02	COMP-2	Tetrachloro-m-xyl	1	20	21.6	108		32	144
		Decachlorobiphen	1	20	16.7	83		32	175
		Tetrachloro-m-xyl	2	20	20.1	101		32	144
		Decachlorobiphen	2	20	19.6	98		32	175
Q3155-03	COMP-3	Tetrachloro-m-xyl	1	20	21.4	107		32	144
		Decachlorobiphen	1	20	16.3	82		32	175
		Tetrachloro-m-xyl	2	20	20.1	101		32	144
		Decachlorobiphen	2	20	18.7	94		32	175
I.BLK-PO113923.D	PIBLK-PO113923.D	Tetrachloro-m-xyl	1	20	18.2	91		60	140
		Decachlorobiphen	1	20	16.1	81		60	140
		Tetrachloro-m-xyl	2	20	17.7	88		60	140
		Decachlorobiphen	2	20	18.3	92		60	140
I.BLK-PO113952.D	PIBLK-PO113952.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	18.2	91		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
		Decachlorobiphen	2	20	18.9	95		60	140
PB169771BL	PB169771BL	Tetrachloro-m-xyl	1	20	19.7	99		32	144
		Decachlorobiphen	1	20	20.2	101		32	175
		Tetrachloro-m-xyl	2	20	19.1	95		32	144
		Decachlorobiphen	2	20	20.8	104		32	175
PB169771BS	PB169771BS	Tetrachloro-m-xyl	1	20	19.4	97		32	144
		Decachlorobiphen	1	20	20.6	103		32	175
		Tetrachloro-m-xyl	2	20	18.5	93		32	144
		Decachlorobiphen	2	20	21.6	108		32	175
I.BLK-PO113958.D	PIBLK-PO113958.D	Tetrachloro-m-xyl	1	20	18.2	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
		Decachlorobiphen	2	20	18.9	95		60	140
I.BLK-PQ070788.D	PIBLK-PQ070788.D	Tetrachloro-m-xyl	1	20	17.2	86		60	140

Surrogate Summary

SDG No.: Q3155

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PQ070788.D	PIBLK-PQ070788.D	Decachlorobiphen	1	20	18.9	94		60	140
		Tetrachloro-m-xyl	2	20	17.6	88		60	140
		Decachlorobiphen	2	20	18.5	93		60	140
I.BLK-PQ070934.D	PIBLK-PQ070934.D	Tetrachloro-m-xyl	1	20	16.9	85		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	15.8	79		60	140
		Decachlorobiphen	2	20	14.9	74		60	140
Q3152-01MS	EO-02-091925MS	Tetrachloro-m-xyl	1	20	14.4	72		32	144
		Decachlorobiphen	1	20	13.3	67		32	175
		Tetrachloro-m-xyl	2	20	12.7	64		32	144
		Decachlorobiphen	2	20	9.14	46		32	175
Q3152-01MSD	EO-02-091925MSD	Tetrachloro-m-xyl	1	20	15.0	75		32	144
		Decachlorobiphen	1	20	12.9	64		32	175
		Tetrachloro-m-xyl	2	20	13.1	66		32	144
		Decachlorobiphen	2	20	8.88	44		32	175
I.BLK-PQ070945.D	PIBLK-PQ070945.D	Tetrachloro-m-xyl	1	20	17.3	86		60	140
		Decachlorobiphen	1	20	16.0	80		60	140
		Tetrachloro-m-xyl	2	20	15.4	77		60	140
		Decachlorobiphen	2	20	12.7	64		60	140