



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3155	Analytical Method:	8082A
Client:	Kleinfelder	Datafile :	PO113954.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169771BS (Column 1)	AR1016	166.5	156	ug/kg	94				71	120	
	AR1260	166.5	163	ug/kg	98				65	130	
PB169771BS (Column 2)	AR1016	166.5	146	ug/kg	88				71	120	
	AR1260	166.5	139	ug/kg	83				65	130	