

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3155

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 00:46

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
Aldrin	5.23	5.23	5.13	5.33	0.00
Dieldrin	6.31	6.31	6.21	6.41	0.01
4,4'-DDE	6.16	6.16	6.06	6.26	0.00
4,4'-DDD	6.66	6.67	6.57	6.77	0.01
4,4'-DDT	6.98	6.98	6.88	7.08	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3155

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 00:46

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
Aldrin	4.29	4.29	4.19	4.39	0.00
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
4,4'-DDD	5.85	5.85	5.75	5.95	0.00
4,4'-DDT	6.10	6.10	6.00	6.20	0.00