

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097317.D **Time Analyzed:** 11:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.664	6.566	6.766	51.270	50.000	2.5
4,4'-DDE	6.156	6.056	6.256	50.930	50.000	1.9
4,4'-DDT	6.978	6.880	7.080	47.230	50.000	-5.5
Aldrin	5.232	5.132	5.332	49.650	50.000	-0.7
Decachlorobiphenyl	8.997	8.902	9.102	49.650	50.000	-0.7
Dieldrin	6.305	6.206	6.406	49.240	50.000	-1.5
Tetrachloro-m-xylene	3.527	3.428	3.628	49.820	50.000	-0.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097317.D **Time Analyzed:** 11:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.848	5.750	5.950	49.680	50.000	-0.6
4,4'-DDE	5.296	5.197	5.397	49.380	50.000	-1.2
4,4'-DDT	6.101	6.002	6.202	47.130	50.000	-5.7
Aldrin	4.290	4.191	4.391	49.760	50.000	-0.5
Decachlorobiphenyl	7.982	7.884	8.084	44.610	50.000	-10.8
Dieldrin	5.427	5.328	5.528	49.310	50.000	-1.4
Tetrachloro-m-xylene	2.824	2.724	2.924	49.770	50.000	-0.5