

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.664	6.566	6.766	52.070	50.000	4.1
4,4'-DDE	6.156	6.056	6.256	50.410	50.000	0.8
4,4'-DDT	6.979	6.880	7.080	45.630	50.000	-8.7
Aldrin	5.232	5.132	5.332	50.300	50.000	0.6
Decachlorobiphenyl	9.001	8.902	9.102	48.730	50.000	-2.5
Dieldrin	6.305	6.206	6.406	48.860	50.000	-2.3
Tetrachloro-m-xylene	3.527	3.428	3.628	51.250	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.849	5.750	5.950	48.330	50.000	-3.3
4,4'-DDE	5.297	5.197	5.397	48.430	50.000	-3.1
4,4'-DDT	6.102	6.002	6.202	43.840	50.000	-12.3
Aldrin	4.291	4.191	4.391	49.350	50.000	-1.3
Decachlorobiphenyl	7.984	7.884	8.084	40.500	50.000	-19.0
Dieldrin	5.427	5.328	5.528	48.310	50.000	-3.4
Tetrachloro-m-xylene	2.824	2.724	2.924	49.030	50.000	-1.9