

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.664	6.566	6.766	57.380	50.000	14.8
4,4'-DDE	6.155	6.056	6.256	56.060	50.000	12.1
4,4'-DDT	6.978	6.880	7.080	50.060	50.000	0.1
Aldrin	5.232	5.132	5.332	54.320	50.000	8.6
Decachlorobiphenyl	8.999	8.902	9.102	53.620	50.000	7.2
Dieldrin	6.305	6.206	6.406	53.760	50.000	7.5
Tetrachloro-m-xylene	3.527	3.428	3.628	54.690	50.000	9.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3155
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.849	5.750	5.950	55.530	50.000	11.1
4,4'-DDE	5.295	5.197	5.397	54.070	50.000	8.1
4,4'-DDT	6.101	6.002	6.202	49.450	50.000	-1.1
Aldrin	4.291	4.191	4.391	54.450	50.000	8.9
Decachlorobiphenyl	7.983	7.884	8.084	48.470	50.000	-3.1
Dieldrin	5.428	5.328	5.528	54.040	50.000	8.1
Tetrachloro-m-xylene	2.824	2.724	2.924	53.880	50.000	7.8