

Surrogate Summary

SDG No.: **Q3155**

Client: **Kleinfelder**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097268.D	PIBLK-PL097268.D	Decachlorobiphen	1	20	17.3	87		57	171
		Tetrachloro-m-xyl	1	20	15.6	78		61	148
		Decachlorobiphen	2	20	17.1	86		57	171
		Tetrachloro-m-xyl	2	20	16.0	80		61	148
I.BLK-PL097315.D	PIBLK-PL097315.D	Decachlorobiphen	1	20	20.8	104		57	171
		Tetrachloro-m-xyl	1	20	17.6	88		61	148
		Decachlorobiphen	2	20	18.6	93		57	171
		Tetrachloro-m-xyl	2	20	18.7	94		61	148
PB169772BL	PB169772BL	Decachlorobiphen	1	20	16.8	84		20	144
		Tetrachloro-m-xyl	1	20	14.9	75		19	148
		Decachlorobiphen	2	20	14.7	73		20	144
		Tetrachloro-m-xyl	2	20	15.3	76		19	148
PB169772BS	PB169772BS	Decachlorobiphen	1	20	25.4	127		20	144
		Tetrachloro-m-xyl	1	20	23.5	117		19	148
		Decachlorobiphen	2	20	22.5	112		20	144
		Tetrachloro-m-xyl	2	20	23.6	118		19	148
I.BLK-PL097327.D	PIBLK-PL097327.D	Decachlorobiphen	1	20	20.2	101		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	17.3	87		57	171
		Tetrachloro-m-xyl	2	20	18.6	93		61	148
I.BLK-PL097339.D	PIBLK-PL097339.D	Decachlorobiphen	1	20	21.0	105		57	171
		Tetrachloro-m-xyl	1	20	18.3	91		61	148
		Decachlorobiphen	2	20	19.2	96		57	171
		Tetrachloro-m-xyl	2	20	18.9	95		61	148
Q3153-01MS	OR-500-COMP-39MS	Decachlorobiphen	1	20	13.1	65		20	144
		Tetrachloro-m-xyl	1	20	15.0	75		19	148
		Decachlorobiphen	2	20	11.5	58		20	144
		Tetrachloro-m-xyl	2	20	15.4	77		19	148
Q3153-01MSD	OR-500-COMP-39MSD	Decachlorobiphen	1	20	12.9	64		20	144
		Tetrachloro-m-xyl	1	20	14.9	74		19	148
		Decachlorobiphen	2	20	11.5	57		20	144
		Tetrachloro-m-xyl	2	20	15.4	77		19	148
Q3155-01	COMP-1	Decachlorobiphen	1	20	13.0	65		20	144
		Tetrachloro-m-xyl	1	20	14.1	71		19	148
		Decachlorobiphen	2	20	12.2	61		20	144
		Tetrachloro-m-xyl	2	20	14.3	72		19	148
Q3155-02	COMP-2	Decachlorobiphen	1	20	17.2	86		20	144
		Tetrachloro-m-xyl	1	20	16.5	83		19	148
		Decachlorobiphen	2	20	17.4	87		20	144
		Tetrachloro-m-xyl	2	20	16.1	80		19	148
Q3155-03	COMP-3	Decachlorobiphen	1	20	11.6	58		20	144

Surrogate Summary

SDG No.: Q3155

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3155-03	COMP-3	Tetrachloro-m-xyl	1	20	11.4	57		19	148
		Decachlorobiphen	2	20	10.6	53		20	144
		Tetrachloro-m-xyl	2	20	11.1	56		19	148
I.BLK-PL097354.D	PIBLK-PL097354.D	Decachlorobiphen	1	20	20.8	104		57	171
		Tetrachloro-m-xyl	1	20	18.1	91		61	148
		Decachlorobiphen	2	20	17.4	87		57	171
		Tetrachloro-m-xyl	2	20	18.6	93		61	148