

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3155</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PL097320.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169772BS (Column 1)	Aldrin	16.66	15.6	ug/kg	94				82	124		
	Dieldrin	16.66	15.8	ug/kg	95				85	121		
	4,4'-DDE	16.66	16.3	ug/kg	98				81	123		
	4,4'-DDD	16.66	16.6	ug/kg	100				80	131		
	4,4'-DDT	16.66	14.8	ug/kg	89				70	129		
PB169772BS (Column 2)	Aldrin	16.66	15.5	ug/kg	93				82	124		
	Dieldrin	16.66	15.4	ug/kg	92				85	121		
	4,4'-DDE	16.66	15.6	ug/kg	94				81	123		
	4,4'-DDD	16.66	16.0	ug/kg	96				80	131		
	4,4'-DDT	16.66	14.3	ug/kg	86				70	129		