

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3155

Client ID : SSTDCCC040

Date Analyzed: 09/23/2025

Lab File ID: BG064435.D

Time Analyzed: 11:28

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	19841	7.843	85807	10.63	61310	14.47
UPPER LIMIT	39682	8.343	171614	11.128	122620	14.965
LOWER LIMIT	9920.5	7.343	42903.5	10.128	30655	13.965
EPA SAMPLE NO.						
01 COMP-1	19558	7.84	83888	10.62	59255	14.46
02 COMP-2	22529	7.84	95164	10.62	68543	14.46
03 COMP-3	22469	7.84	93751	10.63	66454	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.