

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025724.D
Instrument ID: BNA_P

Contract: POWE02
SDG NO.: Q3155
DFTPP Injection Date: 09/11/2025
DFTPP Injection Time: 08:09

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	36.5
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	10.8
442	Greater than 50% of mass 198	69.4
443	15.0 - 24.0% of mass 442	13.7 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP025725.D	09/11/2025	08:56
SSTDICC005	SSTDICC005	BP025726.D	09/11/2025	09:37
SSTDICC010	SSTDICC010	BP025727.D	09/11/2025	10:18
SSTDICC020	SSTDICC020	BP025728.D	09/11/2025	10:59
SSTDICCC040	SSTDICCC040	BP025729.D	09/11/2025	12:21
SSTDICC060	SSTDICC060	BP025731.D	09/11/2025	13:43
SSTDICC080	SSTDICC080	BP025732.D	09/11/2025	14:24
SSTDICC050	SSTDICC050	BP025733.D	09/11/2025	15:20