

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN092325

Review By	John Carlone	Review On	9/24/2025 8:10:38 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:41:42 AM
SubDirectory	VN092325	HP Acquire Method	HP Processing Method 82N092325W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135592		
Initial Calibration Stds	VP135674,VP135675,VP135676,VP135677,VP135678		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP135679		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087922.D	23 Sep 2025 08:31		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN087923.D	23 Sep 2025 08:51	Not used	JC\MD	Not Ok
3	VSTDIC005	VSTDIC005	VN087924.D	23 Sep 2025 09:33		JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN087925.D	23 Sep 2025 09:54	Spiked Low	JC\MD	Not Ok
5	VSTDIC050	VSTDIC050	VN087926.D	23 Sep 2025 10:15		JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN087927.D	23 Sep 2025 10:36		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN087928.D	23 Sep 2025 10:56		JC\MD	Ok,M
8	IBLK	IBLK	VN087929.D	23 Sep 2025 11:17		JC\MD	Ok
9	VSTDIC020	VSTDIC020	VN087930.D	23 Sep 2025 11:47		JC\MD	Ok,M
10	VSTDICV050	ICVVN092325	VN087931.D	23 Sep 2025 15:08		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN092425

Review By	John Carlone	Review On	9/25/2025 10:14:21 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:46:09 AM
SubDirectory	VN092425	HP Acquire Method	HP Processing Method 82N092325W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135601		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135602,VP135603		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087932.D	24 Sep 2025 07:41		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087933.D	24 Sep 2025 08:27	pH#Lot#V12668	JC\MD	Ok,M
3	VN0924MBL01	VN0924MBL01	VN087934.D	24 Sep 2025 09:26		JC\MD	Ok
4	VN0924WBL01	VN0924WBL01	VN087935.D	24 Sep 2025 09:47		JC\MD	Ok
5	VN0924WBS01	VN0924WBS01	VN087936.D	24 Sep 2025 10:08		JC\MD	Ok,M
6	VN0924WBSD01	VN0924WBSD01	VN087937.D	24 Sep 2025 10:42		JC\MD	Ok,M
7	IBLK	IBLK	VN087938.D	24 Sep 2025 11:04		JC\MD	Ok
8	Q3155-04	TB	VN087939.D	24 Sep 2025 11:25	vial A pH<2 TB	JC\MD	Ok
9	Q3157-03	TB	VN087940.D	24 Sep 2025 11:46	vial A pH<2 TB	JC\MD	Ok
10	PB169792TB	PB169792TB	VN087941.D	24 Sep 2025 12:07		JC\MD	Ok,M
11	Q3159-02	VNJ-245	VN087942.D	24 Sep 2025 12:28	vial A pH#5.0	JC\MD	Ok
12	Q3159-04	VNJ-255	VN087943.D	24 Sep 2025 12:50	vial A pH#5.0	JC\MD	Ok
13	Q3167-02	RT4784	VN087944.D	24 Sep 2025 13:11	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
14	Q3167-02	RT4784	VN087945.D	24 Sep 2025 13:47	vial B pH#5.0	JC\MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VN087946.D	24 Sep 2025 15:37		JC\MD	Ok,M

M : Manual Integration