

Surrogate Summary

SDG No.: Q3155

Client: Kleinfelder

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3155-04	TB	1,2-Dichloroethane-d4	50	57.4	115		74	125
		Dibromofluoromethane	50	52.7	105		75	124
		Toluene-d8	50	49.7	99		86	113
		4-Bromofluorobenzene	50	45.3	91		77	121
VN0924WBL01	VN0924WBL01	1,2-Dichloroethane-d4	50	54.7	109		74	125
		Dibromofluoromethane	50	50.0	100		75	124
		Toluene-d8	50	49.3	99		86	113
		4-Bromofluorobenzene	50	46.5	93		77	121
VN0924WBS01	VN0924WBS01	1,2-Dichloroethane-d4	50	53.6	107		74	125
		Dibromofluoromethane	50	49.8	100		75	124
		Toluene-d8	50	49.1	98		86	113
		4-Bromofluorobenzene	50	50.0	100		77	121
VN0924WBSD0	VN0924WBSD01	1,2-Dichloroethane-d4	50	49.3	99		74	125
		Dibromofluoromethane	50	50.4	101		75	124
		Toluene-d8	50	50.0	100		86	113
		4-Bromofluorobenzene	50	49.6	99		77	121