

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3155
 Lab File ID: VN087933.D Date Analyzed: 09/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:27
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	248830	8.21	449654	9.08	427717	11.85
UPPER LIMIT	497660	8.706	899308	9.582	855434	12.347
LOWER LIMIT	124415	7.706	224827	8.582	213859	11.347
EPA SAMPLE NO.						
TB	174992	8.21	380143	9.08	371356	11.85
VN0924WBL01	194391	8.21	423678	9.08	397836	11.84
VN0924WBS01	214330	8.21	404775	9.08	389038	11.85
VN0924WBSD01	242602	8.21	429972	9.08	407713	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.