

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>POWE02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3155</u>
Instrument ID:	<u>MSVOA_W</u>	Calibration Date/Time:	<u>09/19/2025</u> <u>09:38</u>
Lab File ID:	<u>VW032204.D</u>	Init. Calib. Date(s):	<u>08/26/2025</u> <u>08/26/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:39</u> <u>12:34</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.718	0.725		0.98	20
1,1,1-Trichloroethane	0.891	0.911		2.24	20
Benzene	1.406	1.384		-1.57	20
Trichloroethene	0.354	0.347		-1.98	20
Toluene	0.919	0.917		-0.22	20
Ethyl Benzene	1.865	1.878		0.7	20
m/p-Xylenes	0.718	0.729		1.53	20
o-Xylene	0.679	0.677		-0.29	20
Isopropylbenzene	3.397	3.509		3.3	20
1,2-Dichloroethane-d4	0.715	0.725		1.4	20
Dibromofluoromethane	0.351	0.334		-4.84	20
Toluene-d8	1.231	1.189		-3.41	20
4-Bromofluorobenzene	0.455	0.445		-2.2	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.