

Surrogate Summary

SDG No.: **Q3155**

Client: **Kleinfelder**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3155-01	COMP-1	1,2-Dichloroethane-d4	50	52.7	105		63	155
		Dibromofluoromethane	50	45.1	90		70	134
		Toluene-d8	50	42.3	85		74	123
		4-Bromofluorobenzene	50	48.0	96		17	146
Q3155-02	COMP-2	1,2-Dichloroethane-d4	50	55.0	110		63	155
		Dibromofluoromethane	50	45.8	92		70	134
		Toluene-d8	50	42.5	85		74	123
		4-Bromofluorobenzene	50	49.0	98		17	146
Q3155-03	COMP-3	1,2-Dichloroethane-d4	50	55.6	111		63	155
		Dibromofluoromethane	50	46.9	94		70	134
		Toluene-d8	50	43.6	87		74	123
		4-Bromofluorobenzene	50	48.8	98		17	146
VW0919SBL01	VW0919SBL01	1,2-Dichloroethane-d4	50	58.7	117		63	155
		Dibromofluoromethane	50	48.5	97		70	134
		Toluene-d8	50	44.7	89		74	123
		4-Bromofluorobenzene	50	47.4	95		17	146
VW0919SBS01	VW0919SBS01	1,2-Dichloroethane-d4	50	52.9	106		63	155
		Dibromofluoromethane	50	52.9	106		70	134
		Toluene-d8	50	52.6	105		74	123
		4-Bromofluorobenzene	50	50.3	101		17	146
VW0919SBSD0	VW0919SBSD01	1,2-Dichloroethane-d4	50	53.3	107		63	155
		Dibromofluoromethane	50	51.3	103		70	134
		Toluene-d8	50	51.5	103		74	123
		4-Bromofluorobenzene	50	49.5	99		17	146