

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3155
 Lab File ID: VW032204.D Date Analyzed: 09/19/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193605	7.95	337702	8.85	309889	11.63
UPPER LIMIT	387210	8.453	675404	9.349	619778	12.129
LOWER LIMIT	96802.5	7.453	168851	8.349	154945	11.129
EPA SAMPLE NO.						
COMP-1	176619	7.97	381216	8.86	368580	11.63
COMP-2	149567	7.97	332025	8.86	337360	11.63
COMP-3	145040	7.96	318815	8.86	319906	11.63
VW0919SBL01	128568	7.96	281974	8.85	287563	11.63
VW0919SBS01	187746	7.96	322734	8.85	298059	11.63
VW0919SBSD01	180121	7.96	317195	8.86	296209	11.64

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.