

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3155
 Lab File ID: VW032204.D Date Analyzed: 09/19/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	159183	13.556				
UPPER LIMIT	318366	14.056				
LOWER LIMIT	79591.5	13.056				
EPA SAMPLE NO.						
COMP-1	190890	13.56				
COMP-2	169813	13.56				
COMP-3	158309	13.56				
VW0919SBL01	136277	13.56				
VW0919SBS01	153629	13.56				
VW0919SBSD01	152766	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.