

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG No.: Q3156
 Instrument ID: BNA_N Calibration Date/Time: 09/30/2025 23:14
 Lab File ID: BN037959.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.479		-13.8	50.0
Fluoranthene-d10	1.006	0.916		-8.9	50.0
2-Fluorophenol	0.913	0.835		-8.5	50.0
Phenol-d6	1.199	1.111		-7.3	50.0
Nitrobenzene-d5	0.305	0.260		-14.8	50.0
2-Fluorobiphenyl	1.638	1.550		-5.4	50.0
2,4,6-Tribromophenol	0.152	0.135		-11.2	50.0
Terphenyl-d14	0.797	0.784		-1.6	50.0
1,4-Dioxane	0.406	0.400		-1.5	50.0

All other compounds must meet a minimum RRF of 0.010.