

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN093025\
 Data File : BN037959.D
 Acq On : 30 Sep 2025 23:14
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 01 01:24:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	-0.01
2	1,4-Dioxane	0.406	0.400	1.5	122	0.00
3	n-Nitrosodimethylamine	0.540	0.501	7.2	121	0.00
4 S	2-Fluorophenol	0.913	0.835	8.5	122	0.00
5 S	Phenol-d6	1.199	1.111	7.3	124	0.00
6	bis(2-Chloroethyl)ether	1.113	1.095	1.6	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	141	-0.01
8 S	Nitrobenzene-d5	0.305	0.260	14.8	127	-0.01
9	Naphthalene	1.102	1.054	4.4	142	0.00
10	Hexachlorobutadiene	0.188	0.185	1.6	144	-0.01
11 SURR	2-Methylnaphthalene-d10	0.556	0.479	13.8	135	0.00
12	2-Methylnaphthalene	0.720	0.636	11.7	134	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	-0.01
14 S	2,4,6-Tribromophenol	0.152	0.135	11.2	137	-0.01
15 S	2-Fluorobiphenyl	1.638	1.550	5.4	136	0.00
16	Acenaphthylene	1.816	1.710	5.8	141	-0.01
17	Acenaphthene	1.294	1.199	7.3	137	-0.01
18	Fluorene	1.565	1.445	7.7	136	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	143	-0.01
20	4,6-Dinitro-2-methylphenol	0.057	0.037	35.1#	122	-0.01
21	4-Bromophenyl-phenylether	0.261	0.241	7.7	141	-0.01
22	Hexachlorobenzene	0.316	0.305	3.5	143	-0.01
23	Atrazine	0.199	0.165	17.1	138	-0.01
24	Pentachlorophenol	0.108	0.089	17.6	139	-0.01
25	Phenanthrene	1.316	1.228	6.7	143	0.00
26	Anthracene	1.142	1.028	10.0	144	-0.01
27 SURR	Fluoranthene-d10	1.006	0.916	8.9	148	0.00
28	Fluoranthene	1.450	1.329	8.3	145	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	143	0.00
30	Pyrene	1.579	1.544	2.2	145	0.00
31 S	Terphenyl-d14	0.797	0.784	1.6	149	0.00
32	Benzo(a)anthracene	1.398	1.282	8.3	141	0.00
33	Chrysene	1.512	1.488	1.6	145	0.00
34	Bis(2-ethylhexyl)phthalate	0.553	0.546	1.3	160#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	137	0.00
36	Indeno(1,2,3-cd)pyrene	1.828	1.667	8.8	138	-0.01
37	Benzo(b)fluoranthene	1.725	1.661	3.7	139	0.00
38	Benzo(k)fluoranthene	1.736	1.644	5.3	141	0.00
39 C	Benzo(a)pyrene	1.364	1.277	6.4	142	0.00
40	Dibenzo(a,h)anthracene	1.427	1.303	8.7	137	0.00
41	Benzo(g,h,i)perylene	1.499	1.312	12.5	128	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4