

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BN092925

SequenceID : BN092925

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	<u>✓</u>
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	<u>✓</u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
5. GC/MS Calibration Met:	_____	_____	<u>✓</u>
a. Initial calibration Meet Criteria	_____	_____	<u>✓</u>
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____
b. Continuous Calibration(CCC) Meet Criteria	_____	<u>✓</u>	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____
Compound Atrazine,Bis(2-ethylhexyl)phthalate fail Marginally from lowside & Pentachlorophenol fail from highside in the SSTDCCC040 (BN037931.D)	_____	_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____	<u>✓</u>
a. B/N Fraction	_____	_____	_____
d. Acid Fraction	_____	_____	_____

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

____ ✓

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

✓ _____

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

____ ✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

____ ✓

ADDITIONAL COMMENTS:

Surrogates recoveries failed in the water sample Q3165-02 due to matrix interference as seen in the chromatogram. Re-analysis of the sample would give the same result. Hence it was not re-analyzed.

dhruv

Analyst

09/29/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BN092925

SequenceID : BN092925 NA NO YES