

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3156

Client ID : SSTDCCC0.4

Date Analyzed: 09/26/2025

Lab File ID: BN037903.D

Time Analyzed: 00:27

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	8644	7.818	24903	10.63	13049	14.47
	UPPER LIMIT	17288	8.318	49806	11.126	26098	14.973
	LOWER LIMIT	4322	7.318	12451.5	10.126	6524.5	13.973
	EPA SAMPLE NO.						
01	PB169794BL	7888	7.82	21247	10.63	10244	14.47
02	PB169794BS	8269	7.82	22708	10.63	11283	14.47
03	PB169794BSD	7872	7.82	21986	10.63	10679	14.47
04	TT149I1-20250918	6896	7.82	19086	10.63	9820	14.47
05	RE132D3-20250918	6764	7.82	19035	10.63	9869	14.47
06	TT180D1-20250918	6637	7.82	20394	10.63	10463	14.47
07	RE132D1-20250918	6685	7.82	20112	10.63	10348	14.47
08	TT190D2-20250918	8047	7.82	22852	10.62	11726	14.47
09	EB02-20250918	7190	7.82	20705	10.62	11165	14.47
10	EB03-20250918	6448	7.82	20010	10.62	10991	14.47
11	RW9-MW01D1-20250919	5754	7.82	16872	10.62	8712	14.47
12	TT191D2-20250918	7032	7.82	20474	10.62	10961	14.47
13	RW9-MW01S-20250918	6549	7.82	18878	10.62	9985	14.47
14	TT191D1-20250918	6652	7.82	19099	10.62	9587	14.47
15	DUP09-20250918	6521	7.82	18178	10.62	9526	14.47
16	RW9-MW01D3-20250918	6947	7.82	19224	10.63	9834	14.47
17	RW9-MW01D2-20250919	7742	7.82	21958	10.62	11286	14.47

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3156

Client ID : SSTDCCC0.4

Date Analyzed: 09/26/2025

Lab File ID: BN037903.D

Time Analyzed: 00:27

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8644	7.818	24903	10.63	13049	14.47
UPPER LIMIT	17288	8.318	49806	11.126	26098	14.973
LOWER LIMIT	4322	7.318	12451.5	10.126	6524.5	13.973
EPA SAMPLE NO.						

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.