

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3156

Client ID : SSTDCCC0.4

Date Analyzed: 09/29/2025

Lab File ID: BN037931.D

Time Analyzed: 11:30

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9694	7.818	28657	10.62	15093	14.47
UPPER LIMIT	19388	8.318	57314	11.116	30186	14.973
LOWER LIMIT	4847	7.318	14328.5	10.116	7546.5	13.973
EPA SAMPLE NO.						
01 RE132D2-20250918DL	5575	7.82	16174	10.62	8394	14.47
02 TT189D2-20250918DL	6024	7.82	17127	10.62	8846	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.