

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3156

Client ID : SSTDCCC0.4

Date Analyzed: 09/30/2025

Lab File ID: BN037943.D

Time Analyzed: 12:16

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9644	7.811	27222	10.62	14738	14.46
UPPER LIMIT	19288	8.311	54444	11.116	29476	14.962
LOWER LIMIT	4822	7.311	13611	10.116	7369	13.962
EPA SAMPLE NO.						
01 RE132D4-20250918DL	8143	7.81	24296	10.62	12338	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.