

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047894.D
 Acq On : 29 Sep 2025 18:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 30 01:26:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.518	0.585	-12.9	96	0.00
3 P	Chloromethane	0.680	0.697	-2.5	86	0.00
4 C	Vinyl Chloride	0.666	0.724	-8.7#	89	0.00
5 T	Bromomethane	0.422	0.458	-8.5	88	0.00
6 T	Chloroethane	0.445	0.479	-7.6	89	0.00
7 T	Trichlorofluoromethane	0.989	1.060	-7.2	87	0.00
8 T	Diethyl Ether	0.405	0.427	-5.4	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.578	0.619	-7.1	87	0.00
10 T	Methyl Iodide	0.902	0.832	7.8	75	0.00
11 T	Tert butyl alcohol	0.089	0.087	2.2	79	0.00
12 CM	1,1-Dichloroethene	0.594	0.627	-5.6#	85	0.00
13 T	Acrolein	0.082	0.082	0.0	82	0.00
14 T	Allyl chloride	1.226	1.200	2.1	81	0.00
15 T	Acrylonitrile	0.328	0.382	-16.5	90	0.00
16 T	Acetone	0.346	0.297	14.2	77	0.00
17 T	Carbon Disulfide	1.626	1.653	-1.7	86	0.00
18 T	Methyl Acetate	0.770	0.960	-24.7	90	0.00
19 T	Methyl tert-butyl Ether	2.169	2.290	-5.6	83	0.00
20 T	Methylene Chloride	0.732	0.780	-6.6	87	0.00
21 T	trans-1,2-Dichloroethene	0.640	0.676	-5.6	85	0.00
22 T	Diisopropyl ether	2.376	2.543	-7.0	84	0.00
23 T	Vinyl Acetate	1.901	2.026	-6.6	83	0.00
24 P	1,1-Dichloroethane	1.263	1.352	-7.0	85	0.00
25 T	2-Butanone	0.432	0.454	-5.1	85	0.00
26 T	2,2-Dichloropropane	1.050	0.803	23.5	61	0.00
27 T	cis-1,2-Dichloroethene	0.783	0.836	-6.8	85	0.00
28 T	Bromochloromethane	0.551	0.601	-9.1	85	0.00
29 T	Tetrahydrofuran	0.260	0.289	-11.2	86	0.00
30 C	Chloroform	1.276	1.393	-9.2#	86	0.00
31 T	Cyclohexane	1.052	1.001	4.8	79	0.00
32 T	1,1,1-Trichloroethane	1.082	1.138	-5.2	83	0.00
33 S	1,2-Dichloroethane-d4	0.796	0.803	-0.9	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.335	0.348	-3.9	88	0.00
36 T	1,1-Dichloropropene	0.490	0.478	2.4	82	0.00
37 T	Ethyl Acetate	0.525	0.543	-3.4	85	0.00
38 T	Carbon Tetrachloride	0.532	0.535	-0.6	83	0.00
39 T	Methylcyclohexane	0.556	0.525	5.6	78	0.00
40 TM	Benzene	1.488	1.545	-3.8	85	0.00
41 T	Methacrylonitrile	0.271	0.290	-7.0	85	0.00
42 TM	1,2-Dichloroethane	0.566	0.573	-1.2	82	0.00
43 T	Isopropyl Acetate	0.860	0.873	-1.5	81	0.00
44 TM	Trichloroethene	0.360	0.370	-2.8	84	0.00
45 C	1,2-Dichloropropane	0.381	0.404	-6.0#	85	0.00
46 T	Dibromomethane	0.277	0.292	-5.4	85	0.00
47 T	Bromodichloromethane	0.572	0.605	-5.8	83	0.00
48 T	Methyl methacrylate	0.429	0.452	-5.4	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047894.D
 Acq On : 29 Sep 2025 18:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 30 01:26:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.005	-25.0	92	0.00
50 S	Toluene-d8	1.160	1.168	-0.7	85	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.529	-10.9	87	0.00
52 CM	Toluene	0.917	0.940	-2.5#	84	0.00
53 T	t-1,3-Dichloropropene	0.584	0.573	1.9	77	0.00
54 T	cis-1,3-Dichloropropene	0.624	0.619	0.8	79	0.00
55 T	1,1,2-Trichloroethane	0.354	0.380	-7.3	87	0.00
56 T	Ethyl methacrylate	0.555	0.593	-6.8	83	0.00
57 T	1,3-Dichloropropane	0.607	0.656	-8.1	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.293	-29.1#	90	0.00
59 T	2-Hexanone	0.343	0.362	-5.5	84	0.00
60 T	Dibromochloromethane	0.407	0.447	-9.8	86	0.00
61 T	1,2-Dibromoethane	0.360	0.395	-9.7	87	0.00
62 S	4-Bromofluorobenzene	0.440	0.447	-1.6	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.326	0.328	-0.6	86	0.00
65 PM	Chlorobenzene	1.136	1.164	-2.5	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.413	-5.4	85	0.00
67 C	Ethyl Benzene	1.960	1.962	-0.1#	82	0.00
68 T	m/p-Xylenes	0.729	0.744	-2.1	83	0.00
69 T	o-Xylene	0.706	0.720	-2.0	82	0.00
70 T	Styrene	1.236	1.282	-3.7	84	0.00
71 P	Bromoform	0.293	0.316	-7.8	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.801	3.606	5.1	82	0.00
74 T	N-amyl acetate	1.733	1.665	3.9	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.192	-3.7	89	0.00
76 T	1,2,3-Trichloropropane	1.012	1.046	-3.4	85	0.00
77 T	Bromobenzene	0.936	0.910	2.8	84	0.00
78 T	n-propylbenzene	4.494	4.282	4.7	82	0.00
79 T	2-Chlorotoluene	2.752	2.594	5.7	81	0.00
80 T	1,3,5-Trimethylbenzene	3.123	2.981	4.5	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.344	13.4	76	0.00
82 T	4-Chlorotoluene	3.251	3.087	5.0	82	0.00
83 T	tert-Butylbenzene	3.197	2.976	6.9	81	0.00
84 T	1,2,4-Trimethylbenzene	3.166	3.024	4.5	83	0.00
85 T	sec-Butylbenzene	3.797	3.561	6.2	81	0.00
86 T	p-Isopropyltoluene	3.215	2.998	6.7	80	0.00
87 T	1,3-Dichlorobenzene	1.739	1.659	4.6	84	0.00
88 T	1,4-Dichlorobenzene	1.785	1.686	5.5	84	0.00
89 T	n-Butylbenzene	2.975	2.733	8.1	80	0.00
90 T	Hexachloroethane	0.564	0.537	4.8	83	0.00
91 T	1,2-Dichlorobenzene	1.657	1.607	3.0	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.224	3.9	81	0.00
93 T	1,2,4-Trichlorobenzene	1.077	0.969	10.0	77	0.00
94 T	Hexachlorobutadiene	0.366	0.320	12.6	76	0.00
95 T	Naphthalene	3.279	3.173	3.2	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047894.D
Acq On : 29 Sep 2025 18:49
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 30 01:26:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	0.936	6.6	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6