



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047824.D
Acq On : 26 Sep 2025 00:46
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Title : SW846 8260
Last Update : Wed Sep 17 06:39:58 2025



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.8	11063	PASS
75	95	30	60	51.3	30160	PASS
95	95	100	100	100.0	58787	PASS
96	95	5	9	6.4	3761	PASS
173	174	0.00	2	0.5	218	PASS
174	95	50	100	76.6	45032	PASS
175	174	5	9	7.7	3475	PASS
176	174	95	101	95.9	43203	PASS
177	176	5	9	6.6	2870	PASS