

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3156
 Lab File ID: VX047868.D Date Analyzed: 09/29/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	185893	5.53	309770	6.74	268975	10.04
UPPER LIMIT	371786	6.031	619540	7.238	537950	10.537
LOWER LIMIT	92946.5	5.031	154885	6.238	134488	9.537
EPA SAMPLE NO.						
EB03-20250918	191712	5.54	390271	6.75	393970	10.04
TB03-20250912	121586	5.54	250408	6.75	258535	10.04
RW9-MW01D1-20250919	128268	5.54	266783	6.75	263857	10.04
DUP09-20250918	162647	5.54	350912	6.75	362589	10.04
RW9-MW01D3-20250918	118564	5.54	244768	6.75	250385	10.04
RW9-MW01D2-20250919	142790	5.54	301628	6.75	311543	10.04
VX0929WBL01	151039	5.54	304108	6.75	305847	10.04
VX0929WBS01	165805	5.54	286774	6.75	261167	10.04
VX0929WBSD01	159091	5.54	281426	6.75	258238	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.