

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3156
 Lab File ID: VX047896.D Date Analyzed: 09/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:20
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	159265	5.53	274684	6.74	247069	10.04
UPPER LIMIT	318530	6.031	549368	7.239	494138	10.537
LOWER LIMIT	79632.5	5.031	137342	6.239	123535	9.537
EPA SAMPLE NO.						
TT180D1-20250918	132223	5.54	271727	6.75	277527	10.04
EB02-20250918	125275	5.54	259422	6.75	266390	10.04
TT191D2-20250918	125420	5.54	258118	6.75	260626	10.04
RW9-MW01S-20250918	131219	5.54	272909	6.75	274678	10.04
TT191D1-20250918	129020	5.54	267176	6.75	271468	10.04
VX0930WBL01	126165	5.54	254131	6.75	265110	10.04
VX0930WBS01	139945	5.54	252162	6.74	231700	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.