

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3156</u>
Lab File ID:	<u>VX047825.D</u>	Date Analyzed:	<u>09/26/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>01:28</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	110381	12.00€				
UPPER LIMIT	220762	12.50€				
LOWER LIMIT	55190.5	11.50€				
EPA SAMPLE NO.						
RE10-MW01S-20250918	123234	12.01				
RE131D3-20250918	140996	12.01				
TT190D1-20250918	125121	12.01				
RE132D2-20250918	157952	12.01				
RE132D4-20250918	126267	12.01				
TT189D2-20250918	124851	12.01				
TT149I1-20250918	153439	12.01				
RE132D3-20250918	126419	12.01				
RE132D1-20250918	127602	12.01				
TT190D2-20250918	131382	12.01				
VX0925WBL02	138899	12.01				
VX0925WBS02	122300	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.