

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3156</u>
Lab File ID:	<u>VX047868.D</u>	Date Analyzed:	<u>09/29/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:59</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	131495	12.00€				
UPPER LIMIT	262990	12.50€				
LOWER LIMIT	65747.5	11.50€				
EPA SAMPLE NO.						
EB03-20250918	200279	12.01				
TB03-20250912	125044	12.01				
RW9-MW01D1-20250919	123079	12.01				
DUP09-20250918	185913	12.01				
RW9-MW01D3-20250918	126372	12.01				
RW9-MW01D2-20250919	162191	12.01				
VX0929WBL01	151711	12.01				
VX0929WBS01	131097	12.01				
VX0929WBSD01	129173	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.