



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Always Available Construction LLC.

SDG No.: Q3157

Project: Yard Soil Cleanup

RunNo.: LB137277

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.097	0.099	98	90-110	09/23/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	09/23/2025
Sample ID: Cyanide	CCV2	mg/L	0.26	0.25	104	90-110	09/23/2025