

Initial and Continuing Calibration Blank Summary

Client: Always Available Construction LLC.

SDG No.: Q3157

Project: Yard Soil Cleanup

RunNo.: LB137277

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	09/23/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	09/23/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	09/23/2025