

Duplicate Sample Summary

Client:	Always Available Construction LLC.	SDG No.:	Q3157
Project:	Yard Soil Cleanup	Sample ID:	Q3153-01
Client ID:	OR-500-COMP-39DUP	Percent Solids for Spike Sample:	93

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.043	U	0.043	U	1	0		09/23/2025

Duplicate Sample Summary

Client:	Always Available Construction LLC.	SDG No.:	Q3157
Project:	Yard Soil Cleanup	Sample ID:	Q3153-01
Client ID:	OR-500-COMP-39MSD	Percent Solids for Spike Sample:	93

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	1.90		2.00		1	5		09/23/2025