

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Always Available Construction LLC.

SDG No.: Q3157

Contract: ALWA01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	247000	255000	97	216000	294000	09/24/2025	12:04	LB137295
	Antimony	-1.57			-50	50	09/24/2025	12:04	LB137295
	Arsenic	7.04			-20	20	09/24/2025	12:04	LB137295
	Barium	1.38	6.0	23	-94	106	09/24/2025	12:04	LB137295
	Beryllium	1.90			-6	6	09/24/2025	12:04	LB137295
	Cadmium	1.05	1.0	105	-5	7	09/24/2025	12:04	LB137295
	Calcium	234000	245000	96	208000	282000	09/24/2025	12:04	LB137295
	Chromium	49.2	52.0	95	42	62	09/24/2025	12:04	LB137295
	Cobalt	1.52			-30	30	09/24/2025	12:04	LB137295
	Copper	15.3	2.0	765	-18	22	09/24/2025	12:04	LB137295
	Iron	103000	101000	102	85600	116500	09/24/2025	12:04	LB137295
	Lead	0.29			-12	12	09/24/2025	12:04	LB137295
	Magnesium	248000	255000	97	216000	294000	09/24/2025	12:04	LB137295
	Manganese	7.06	7.0	101	-13	27	09/24/2025	12:04	LB137295
	Nickel	4.00	2.0	200	-38	42	09/24/2025	12:04	LB137295
	Potassium	87.4			0	0	09/24/2025	12:04	LB137295
	Selenium	5.58			-20	20	09/24/2025	12:04	LB137295
	Silver	3.73			-10	10	09/24/2025	12:04	LB137295
	Sodium	57.7			0	0	09/24/2025	12:04	LB137295
	Thallium	-0.065			-40	40	09/24/2025	12:04	LB137295
	Vanadium	6.87			-40	40	09/24/2025	12:04	LB137295
	Zinc	5.23			-40	40	09/24/2025	12:04	LB137295
ICSAB01	Aluminum	248000	247000	100	209000	285000	09/24/2025	12:08	LB137295
	Antimony	618	618	100	525	711	09/24/2025	12:08	LB137295
	Arsenic	109	104	105	88.4	120	09/24/2025	12:08	LB137295
	Barium	477	537	89	437	637	09/24/2025	12:08	LB137295
	Beryllium	483	495	98	420	570	09/24/2025	12:08	LB137295
	Cadmium	959	972	99	826	1120	09/24/2025	12:08	LB137295
	Calcium	235000	235000	100	199000	271000	09/24/2025	12:08	LB137295
	Chromium	549	542	101	460	624	09/24/2025	12:08	LB137295
	Cobalt	482	476	101	404	548	09/24/2025	12:08	LB137295
	Copper	484	511	95	434	588	09/24/2025	12:08	LB137295
	Iron	103000	99300	104	84400	114500	09/24/2025	12:08	LB137295
	Lead	46.7	49.0	95	37	61	09/24/2025	12:08	LB137295
	Magnesium	249000	248000	100	210000	286000	09/24/2025	12:08	LB137295
	Manganese	477	507	94	430	584	09/24/2025	12:08	LB137295
	Nickel	961	954	101	810	1100	09/24/2025	12:08	LB137295
	Potassium	15.6			0	0	09/24/2025	12:08	LB137295
	Selenium	55.6	46.0	121	26	66	09/24/2025	12:08	LB137295
	Silver	214	201	106	170	232	09/24/2025	12:08	LB137295
	Sodium	70.9			0	0	09/24/2025	12:08	LB137295
	Thallium	87.8	108	81	68	148	09/24/2025	12:08	LB137295

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Always Available Construction LLC.

SDG No.: Q3157

Contract: ALWA01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	479	491	98	417	565	09/24/2025	12:08	LB137295
	Zinc	1040	952	109	809	1095	09/24/2025	12:08	LB137295