

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Always Available Construction LLC.

SDG No.: Q3157

Contract: ALWA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169761BS							
Aluminum	mg/Kg	100	101		101	80 - 120	P
Antimony	mg/Kg	40.0	42.1		105	80 - 120	P
Arsenic	mg/Kg	40.0	40.5		101	80 - 120	P
Barium	mg/Kg	10.0	9.35		94	80 - 120	P
Beryllium	mg/Kg	10.0	10.8		108	80 - 120	P
Cadmium	mg/Kg	10.0	9.81		98	80 - 120	P
Calcium	mg/Kg	50.0	51.6	J	103	80 - 120	P
Chromium	mg/Kg	20.0	21.0		105	80 - 120	P
Cobalt	mg/Kg	10.0	9.89		99	80 - 120	P
Copper	mg/Kg	15.0	15.6		104	80 - 120	P
Iron	mg/Kg	150	147		98	80 - 120	P
Lead	mg/Kg	50.0	48.7		97	80 - 120	P
Magnesium	mg/Kg	100	100		100	80 - 120	P
Manganese	mg/Kg	10.0	9.93		99	80 - 120	P
Nickel	mg/Kg	25.0	24.9		100	80 - 120	P
Potassium	mg/Kg	500	492		98	80 - 120	P
Selenium	mg/Kg	100	104		104	80 - 120	P
Silver	mg/Kg	3.8	3.93		103	80 - 120	P
Sodium	mg/Kg	150	152		101	80 - 120	P
Thallium	mg/Kg	100	94.1		94	80 - 120	P
Vanadium	mg/Kg	15.0	15.3		102	80 - 120	P
Zinc	mg/Kg	10.0	10.2		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Always Available Construction LLC.

SDG No.: Q3157

Contract: ALWA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169787BS Mercury	mg/Kg	0.26	0.29		113	80 - 120	CV