



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| ICB40     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 09/23/2025       | 09:06            | LB137272      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|----|------------------|------------------|---------------|
| CCB20     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 09/23/2025       | 09:11            | LB137272      |
| CCB21     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 09/23/2025       | 09:43            | LB137272      |
| CCB22     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 09/23/2025       | 10:14            | LB137272      |
| CCB23     | Mercury | 0.076          | +/-0.2              | U            | 0.20 | CV | 09/23/2025       | 10:24            | LB137272      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID    | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|--------------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| <b>ICB01</b> | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/23/2025       | 14:07            | LB137283      |
|              | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/23/2025       | 14:07            | LB137283      |
|              | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Cadmium   | 0.51           | +/-3                | J            | 6.00 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/23/2025       | 14:07            | LB137283      |
|              | Lead      | 3.01           | +/-6                | J            | 12.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/23/2025       | 14:07            | LB137283      |
|              | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/23/2025       | 14:07            | LB137283      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/23/2025       | 15:23            | LB137283      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/23/2025       | 15:23            | LB137283      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/23/2025       | 15:23            | LB137283      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/23/2025       | 15:23            | LB137283      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/23/2025       | 15:23            | LB137283      |
| CCB02     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/23/2025       | 17:09            | LB137283      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/23/2025       | 17:09            | LB137283      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/23/2025       | 17:09            | LB137283      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/23/2025       | 17:09            | LB137283      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB02     | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/23/2025       | 17:09            | LB137283      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/23/2025       | 17:09            | LB137283      |
| CCB03     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/23/2025       | 18:12            | LB137283      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/23/2025       | 18:12            | LB137283      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/23/2025       | 18:12            | LB137283      |
|           | Lead      | 2.66           | +/-6                | J            | 12.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/23/2025       | 18:12            | LB137283      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/23/2025       | 18:12            | LB137283      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 11:56            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 11:56            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 11:56            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 11:56            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 11:56            | LB137295      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 12:26            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 12:26            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 12:26            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 12:26            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 12:26            | LB137295      |
| CCB02     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 13:44            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 13:44            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 13:44            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 13:44            | LB137295      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB02     | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 13:44            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 13:44            | LB137295      |
| CCB03     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 14:46            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 14:46            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 14:46            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 14:46            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 14:46            | LB137295      |
| CCB04     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 15:46            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 15:46            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 15:46            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 15:46            | LB137295      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB04     | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 15:46            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 15:46            | LB137295      |
| CCB05     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 16:35            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 16:35            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 16:35            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 16:35            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 16:35            | LB137295      |
| CCB06     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 17:37            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 17:37            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 17:37            | LB137295      |

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB06     | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 17:37            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 17:37            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 17:37            | LB137295      |
| CCB07     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 18:42            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 18:42            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 18:42            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 18:42            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 18:42            | LB137295      |
| CCB08     | Aluminum  | 11.3           | +/-50               | U            | 100  | P | 09/24/2025       | 19:07            | LB137295      |
|           | Antimony  | 6.76           | +/-25               | U            | 50.0 | P | 09/24/2025       | 19:07            | LB137295      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB08     | Arsenic   | 5.12           | +/-10               | U            | 20.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Barium    | 14.6           | +/-50               | U            | 100  | P | 09/24/2025       | 19:07            | LB137295      |
|           | Beryllium | 0.56           | +/-3                | U            | 6.00 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Cadmium   | 0.50           | +/-3                | U            | 6.00 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Calcium   | 234            | +/-1000             | U            | 2000 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Chromium  | 2.12           | +/-5                | U            | 10.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Cobalt    | 2.26           | +/-15               | U            | 30.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Copper    | 4.60           | +/-10               | U            | 20.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Iron      | 23.4           | +/-50               | U            | 100  | P | 09/24/2025       | 19:07            | LB137295      |
|           | Lead      | 2.30           | +/-6                | U            | 12.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Magnesium | 244            | +/-1000             | U            | 2000 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Manganese | 5.94           | +/-10               | U            | 20.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Nickel    | 3.06           | +/-20               | U            | 40.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Potassium | 918            | +/-1000             | U            | 2000 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Selenium  | 9.64           | +/-10               | U            | 20.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Silver    | 1.62           | +/-5                | U            | 10.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Sodium    | 868            | +/-1000             | U            | 2000 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Thallium  | 4.38           | +/-20               | U            | 40.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Vanadium  | 6.26           | +/-20               | U            | 40.0 | P | 09/24/2025       | 19:07            | LB137295      |
|           | Zinc      | 16.7           | +/-20               | U            | 40.0 | P | 09/24/2025       | 19:07            | LB137295      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Always Available Construction LLC.

**SDG No.:** Q3157

**Contract:** ALWA01

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|