



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Always Available Construction LLC.

SDG No.: Q3157

Contract: ALWA01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	22.1	20.0	110	65 - 135	P	09/22/2025	13:05	LB137271
	Barium	102	100	102	65 - 135	P	09/22/2025	13:05	LB137271
	Cadmium	6.04	6.0	101	65 - 135	P	09/22/2025	13:05	LB137271
	Chromium	10.3	10.0	103	65 - 135	P	09/22/2025	13:05	LB137271
	Lead	13.9	12.0	116	65 - 135	P	09/22/2025	13:05	LB137271
	Selenium	25.3	20.0	126	65 - 135	P	09/22/2025	13:05	LB137271
	Silver	10.1	10.0	101	65 - 135	P	09/22/2025	13:05	LB137271
CRA	Mercury	0.22	0.2	108	70 - 130	CV	09/23/2025	11:28	LB137274
CRI01	Arsenic	20.9	20.0	105	65 - 135	P	09/24/2025	12:00	LB137295
	Barium	97.2	100	97	65 - 135	P	09/24/2025	12:00	LB137295
	Cadmium	5.84	6.0	97	65 - 135	P	09/24/2025	12:00	LB137295
	Chromium	9.93	10.0	99	65 - 135	P	09/24/2025	12:00	LB137295
	Lead	11.1	12.0	92	65 - 135	P	09/24/2025	12:00	LB137295
	Selenium	21.7	20.0	109	65 - 135	P	09/24/2025	12:00	LB137295
	Silver	11.5	10.0	115	65 - 135	P	09/24/2025	12:00	LB137295
CRI01	Arsenic	21.8	20.0	109	65 - 135	P	09/25/2025	11:38	LB137314
	Barium	105	100	105	65 - 135	P	09/25/2025	11:38	LB137314
	Cadmium	6.34	6.0	106	65 - 135	P	09/25/2025	11:38	LB137314
	Chromium	10.7	10.0	107	65 - 135	P	09/25/2025	11:38	LB137314
	Lead	11.4	12.0	95	65 - 135	P	09/25/2025	11:38	LB137314
	Selenium	25.2	20.0	126	65 - 135	P	09/25/2025	11:38	LB137314
	Silver	11.0	10.0	110	65 - 135	P	09/25/2025	11:38	LB137314