

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Always Available Construction LLC.

SDG No.: Q3157

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169754TB		WATER		Batch Number:	PB169781		Prep Date:	09/22/2025	
	Mercury	0.76	<2	U	2.00	CV	09/23/2025	12:36	LB137274
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169781BL		WATER		Batch Number:	PB169781		Prep Date:	09/22/2025	
	Mercury	0.076	<0.2	U	0.20	CV	09/23/2025	11:35	LB137274

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Always Available Construction LLC.

SDG No.: Q3157

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169754TB		WATER		Batch Number:	PB169773		Prep Date:	09/22/2025	
	Arsenic	25.6	<50	U	100	P	09/25/2025	16:47	LB137314
	Barium	72.8	<250	U	500	P	09/25/2025	16:47	LB137314
	Cadmium	2.50	<15	U	30.0	P	09/25/2025	16:47	LB137314
	Chromium	10.6	<25	U	50.0	P	09/25/2025	16:47	LB137314
	Lead	11.5	<30	U	60.0	P	09/25/2025	16:47	LB137314
	Selenium	48.2	<50	U	100	P	09/25/2025	16:47	LB137314
	Silver	8.10	<25	U	50.0	P	09/25/2025	16:47	LB137314
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169773BL		WATER		Batch Number:	PB169773		Prep Date:	09/22/2025	
	Arsenic	25.6	<50	U	100	P	09/22/2025	17:37	LB137271
	Barium	72.8	<250	U	500	P	09/22/2025	17:37	LB137271
	Cadmium	2.50	<15	U	30.0	P	09/22/2025	17:37	LB137271
	Chromium	10.6	<25	U	50.0	P	09/22/2025	17:37	LB137271
	Lead	11.5	<30	U	60.0	P	09/22/2025	17:37	LB137271
	Selenium	48.2	<50	U	100	P	09/22/2025	17:37	LB137271
	Silver	8.10	<25	U	50.0	P	09/22/2025	17:37	LB137271