

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Always Available Construction LLC.

SDG No.: Q3157

Contract: ALWA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169773BS							
Arsenic	ug/L	4000	3900		98	80 - 120	P
Barium	ug/L	1000	953		95	80 - 120	P
Cadmium	ug/L	1000	941		94	80 - 120	P
Chromium	ug/L	2000	1970		98	80 - 120	P
Lead	ug/L	5000	4730		95	80 - 120	P
Selenium	ug/L	10000	9970		100	80 - 120	P
Silver	ug/L	380	373		98	80 - 120	P



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Always Available Construction LLC.

SDG No.: Q3157

Contract: ALWA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169781BS Mercury	ug/L	4.0	4.67		117	80 - 120	CV