

Analytical Sequence

Client: Always Available Construction LLC.

SDG No.: Q3157

Project: Yard Soil Cleanup

Instrument ID: FID_F

GC Column: RXI-1MS **ID:** 0.18 (mm)

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION 15.071					
CLIENT SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
PIBLK01	LBLK01	01 Oct 2025 09:28	FF016482.D	15.063	
50 PPM TRPH STD	50 PPM TRPH STD	01 Oct 2025 10:43	FF016483.D	15.066	
PB169930BL	PB169930BL	01 Oct 2025 11:58	FF016485.D	15.062	
PB169930BS	PB169930BS	01 Oct 2025 12:27	FF016486.D	15.063	
20-DEAD-1	Q3157-01	01 Oct 2025 12:56	FF016487.D	15.064	
20-DEAD-1MS	Q3157-01MS	01 Oct 2025 13:26	FF016488.D	15.066	
20-DEAD-1MSD	Q3157-01MSD	01 Oct 2025 13:55	FF016489.D	15.068	
PIBLK02	LBLK02	01 Oct 2025 14:25	FF016490.D	15.068	
50 PPM TRPH STD	50 PPM TRPH STD	01 Oct 2025 14:55	FF016491.D	15.071	