

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113957.D **Time Analyzed:** 08:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.643	4.843	501.600	500.000	0.3
Aroclor-1016-2	4.759	4.661	4.861	511.580	500.000	2.3
Aroclor-1016-3	4.816	4.718	4.918	504.390	500.000	0.9
Aroclor-1016-4	4.934	4.837	5.037	501.090	500.000	0.2
Aroclor-1016-5	5.191	5.093	5.293	506.890	500.000	1.4
Aroclor-1260-1	6.226	6.129	6.329	547.250	500.000	9.5
Aroclor-1260-2	6.416	6.319	6.519	537.780	500.000	7.6
Aroclor-1260-3	6.781	6.683	6.883	523.420	500.000	4.7
Aroclor-1260-4	7.039	6.941	7.141	592.790	500.000	18.6
Aroclor-1260-5	7.282	7.185	7.385	596.750	500.000	19.4
Decachlorobiphenyl	8.668	8.570	8.770	51.980	50.000	4.0
Tetrachloro-m-xylene	3.659	3.560	3.760	53.050	50.000	6.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113957.D **Time Analyzed:** 08:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.623	4.823	479.330	500.000	-4.1
Aroclor-1016-2	4.738	4.641	4.841	482.560	500.000	-3.5
Aroclor-1016-3	4.913	4.816	5.016	512.650	500.000	2.5
Aroclor-1016-4	4.956	4.858	5.058	494.990	500.000	-1.0
Aroclor-1016-5	5.167	5.070	5.270	543.180	500.000	8.6
Aroclor-1260-1	6.192	6.096	6.296	476.790	500.000	-4.6
Aroclor-1260-2	6.381	6.284	6.484	481.400	500.000	-3.7
Aroclor-1260-3	6.532	6.435	6.635	443.950	500.000	-11.2
Aroclor-1260-4	7.000	6.903	7.103	457.200	500.000	-8.6
Aroclor-1260-5	7.243	7.146	7.346	456.250	500.000	-8.8
Decachlorobiphenyl	8.611	8.514	8.714	52.880	50.000	5.8
Tetrachloro-m-xylene	3.649	3.551	3.751	50.440	50.000	0.9