

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Always Available Construction LLC.</u>	DataFile :	<u>PL097347.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Low	Limits		RPD
			Result				Qual	RPD	Qual		High		
Lab Sample ID:	Q3153-01MS (Column 1)	Client Sample ID:		OR-500-COMP-39MS									
	alpha-BHC	17.89	0	13.9	ug/kg	78				60		144	
	beta-BHC	17.89	0	13.8	ug/kg	77				54		143	
	delta-BHC	17.89	0	14.1	ug/kg	79				29		151	
	gamma-BHC (Lindane)	17.89	0	13.8	ug/kg	77				61		140	
	Heptachlor	17.89	0	13.5	ug/kg	75				63		135	
	Aldrin	17.89	0	13.7	ug/kg	77				49		139	
	Heptachlor epoxide	17.89	0	14.1	ug/kg	79				41		156	
	Endosulfan I	17.89	0	13.9	ug/kg	78				56		142	
	Dieldrin	17.89	0	13.7	ug/kg	77				47		161	
	4,4'-DDE	17.89	0	13.9	ug/kg	78				55		136	
	Endrin	17.89	0	13.2	ug/kg	74				57		139	
	Endosulfan II	17.89	0	12.4	ug/kg	69				40		163	
	4,4'-DDD	17.89	0	14.8	ug/kg	83				47		163	
	Endosulfan sulfate	17.89	0	13.7	ug/kg	77				62		139	
	4,4'-DDT	17.89	0	12.7	ug/kg	71				51		146	
	Methoxychlor	17.89	0	12.5	ug/kg	70				54		136	
	Endrin ketone	17.89	0	13.7	ug/kg	77				60		129	
	Endrin aldehyde	17.89	0	15.0	ug/kg	84				59		132	
	alpha-Chlordane	17.89	0	13.8	ug/kg	77				39		166	
	gamma-Chlordane	17.89	0	14.0	ug/kg	78				44		175	
Lab Sample ID:	Q3153-01MS (Column 2)	Client Sample ID:		OR-500-COMP-39MS									
	alpha-BHC	17.89	0	14.0	ug/kg	78				60		144	
	beta-BHC	17.89	0	13.4	ug/kg	75				54		143	
	delta-BHC	17.89	0	13.4	ug/kg	75				29		151	
	gamma-BHC (Lindane)	17.89	0	13.8	ug/kg	77				61		140	
	Heptachlor	17.89	0	13.4	ug/kg	75				63		135	
	Aldrin	17.89	0	13.7	ug/kg	77				49		139	
	Heptachlor epoxide	17.89	0	13.6	ug/kg	76				41		156	
	Endosulfan I	17.89	0	13.0	ug/kg	73				56		142	
	Dieldrin	17.89	0	13.5	ug/kg	75				47		161	
	4,4'-DDE	17.89	0	13.5	ug/kg	75				55		136	
	Endrin	17.89	0	12.9	ug/kg	72				57		139	
	Endosulfan II	17.89	0	13.1	ug/kg	73				40		163	
	4,4'-DDD	17.89	0	13.0	ug/kg	73				47		163	
	Endosulfan sulfate	17.89	0	12.7	ug/kg	71				62		139	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Always Available Construction LLC.</u>	DataFile :	<u>PL097347.D</u>

Parameter	Spike	Sample		Units	Rec	Rec	Qual	RPD	Qual	Low	Limits	
		Result	Result								High	RPD
4,4'-DDT	17.89	0	11.9	ug/kg	67					51	146	
Methoxychlor	17.89	0	12.0	ug/kg	67					54	136	
Endrin ketone	17.89	0	12.6	ug/kg	70					60	129	
Endrin aldehyde	17.89	0	13.5	ug/kg	75					59	132	
alpha-Chlordane	17.89	0	13.6	ug/kg	76					39	166	
gamma-Chlordane	17.89	0	13.4	ug/kg	75					44	175	