

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097355.D **Time Analyzed:** 00:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.999	8.902	9.102	51.350	50.000	2.7
Endrin	6.531	6.432	6.632	48.870	50.000	-2.3
gamma-BHC (Lindane)	4.301	4.202	4.402	54.300	50.000	8.6
Heptachlor	4.892	4.793	4.993	51.040	50.000	2.1
Heptachlor epoxide	5.651	5.552	5.752	52.630	50.000	5.3
Methoxychlor	7.451	7.352	7.552	44.570	50.000	-10.9
Tetrachloro-m-xylene	3.527	3.428	3.628	54.430	50.000	8.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ALWA01
Lab Code: ACE
SDG NO.: Q3157
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No.: CCAL03
Date Analyzed: 09/23/2025
Lab Sample No.: PSTDCCC050
Data File : PL097355.D
Time Analyzed: 00:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	45.360	50.000	-9.3
Endrin	5.702	5.603	5.803	47.950	50.000	-4.1
gamma-BHC (Lindane)	3.661	3.561	3.761	54.250	50.000	8.5
Heptachlor	4.008	3.909	4.109	51.270	50.000	2.5
Heptachlor epoxide	4.792	4.693	4.893	52.550	50.000	5.1
Methoxychlor	6.673	6.574	6.774	41.840	50.000	-16.3
Tetrachloro-m-xylene	2.824	2.724	2.924	53.670	50.000	7.3