

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097359.D **Time Analyzed:** 10:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.003	8.902	9.102	53.250	50.000	6.5
Endrin	6.533	6.432	6.632	49.330	50.000	-1.3
gamma-BHC (Lindane)	4.302	4.202	4.402	54.410	50.000	8.8
Heptachlor	4.893	4.793	4.993	50.970	50.000	1.9
Heptachlor epoxide	5.652	5.552	5.752	51.910	50.000	3.8
Methoxychlor	7.453	7.352	7.552	46.240	50.000	-7.5
Tetrachloro-m-xylene	3.527	3.428	3.628	52.820	50.000	5.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097359.D **Time Analyzed:** 10:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.884	8.084	49.070	50.000	-1.9
Endrin	5.702	5.603	5.803	47.520	50.000	-5.0
gamma-BHC (Lindane)	3.659	3.561	3.761	52.620	50.000	5.2
Heptachlor	4.007	3.909	4.109	50.390	50.000	0.8
Heptachlor epoxide	4.791	4.693	4.893	51.290	50.000	2.6
Methoxychlor	6.674	6.574	6.774	43.870	50.000	-12.3
Tetrachloro-m-xylene	2.822	2.724	2.924	52.390	50.000	4.8