

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097364.D **Time Analyzed:** 12:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.004	8.902	9.102	52.350	50.000	4.7
Endrin	6.533	6.432	6.632	49.760	50.000	-0.5
gamma-BHC (Lindane)	4.301	4.202	4.402	53.420	50.000	6.8
Heptachlor	4.892	4.793	4.993	50.770	50.000	1.5
Heptachlor epoxide	5.652	5.552	5.752	51.580	50.000	3.2
Methoxychlor	7.454	7.352	7.552	45.610	50.000	-8.8
Tetrachloro-m-xylene	3.526	3.428	3.628	53.100	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097364.D **Time Analyzed:** 12:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	48.450	50.000	-3.1
Endrin	5.700	5.603	5.803	47.320	50.000	-5.4
gamma-BHC (Lindane)	3.657	3.561	3.761	52.520	50.000	5.0
Heptachlor	4.005	3.909	4.109	49.830	50.000	-0.3
Heptachlor epoxide	4.790	4.693	4.893	51.900	50.000	3.8
Methoxychlor	6.673	6.574	6.774	43.640	50.000	-12.7
Tetrachloro-m-xylene	2.820	2.724	2.924	52.450	50.000	4.9