

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097269.D Date Analyzed: 09/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.000	8.900	9.100	15.660	20.000	-21.7
Tetrachloro-m-xylene	3.527	3.480	3.580	14.770	20.000	-26.2
alpha-BHC	3.974	3.920	4.020	6.930	10.000	-30.7
beta-BHC	4.489	4.440	4.540	7.340	10.000	-26.6
gamma-BHC (Lindane)	4.302	4.250	4.350	7.210	10.000	-27.9
Endrin	6.532	6.460	6.600	37.780	50.000	-24.4
4,4'-DDT	6.979	6.910	7.050	73.290	100.000	-26.7
Methoxychlor	7.452	7.380	7.520	180.320	250.000	-27.9

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097269.D Date Analyzed: 09/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.880	8.080	15.470	20.000	-22.7
Tetrachloro-m-xylene	2.824	2.770	2.870	14.550	20.000	-27.3
alpha-BHC	3.329	3.280	3.380	6.890	10.000	-31.1
beta-BHC	3.957	3.910	4.010	7.490	10.000	-25.1
gamma-BHC (Lindane)	3.660	3.610	3.710	7.000	10.000	-30.0
Endrin	5.702	5.630	5.770	36.670	50.000	-26.7
4,4'-DDT	6.102	6.030	6.170	76.460	100.000	-23.5
Methoxychlor	6.674	6.600	6.740	166.760	250.000	-33.3