

Cover Page

Order ID : Q3157

Project ID : Yard Soil Cleanup

Client : Always Available Construction LLC.

Lab Sample Number

Q3157-01
Q3157-02
Q3157-03

Client Sample Number

20-DEAD-1
20-DEAD-1
TB

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 9/25/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3157

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: KETAN PATEL

Date: 09/25/2025

LAB CHRONICLE

OrderID:	Q3157	OrderDate:	9/19/2025 2:22:59 PM
Client:	Always Available Construction LLC.	Project:	Yard Soil Cleanup
Contact:	Michael Roth	Location:	D31,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3157-01	20-DEAD-1	SOIL	Gasoline Range Organics	8015D	09/19/25		09/23/25	09/19/25
Q3157-02	20-DEAD-1	TCLP	TCLP Pesticide	8081B	09/19/25	09/22/25	09/22/25	09/19/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3157

Order ID: Q3157

Client: Always Available Construction LLC.

Project ID: Yard Soil Cleanup

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: **Q3157**

Client: **Always Available Construction LLC.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097268.D	PIBLK-PL097268.D	Decachlorobiphen	1	20	17.3	87		57	171
		Tetrachloro-m-xyl	1	20	15.6	78		61	148
		Decachlorobiphen	2	20	17.1	86		57	171
		Tetrachloro-m-xyl	2	20	16.0	80		61	148
I.BLK-PL097327.D	PIBLK-PL097327.D	Decachlorobiphen	1	20	20.2	101		57	171
		Tetrachloro-m-xyl	1	20	18.6	93		61	148
		Decachlorobiphen	2	20	17.3	87		57	171
		Tetrachloro-m-xyl	2	20	18.6	93		61	148
PB169782BL	PB169782BL	Decachlorobiphen	1	20	17.5	88		57	171
		Tetrachloro-m-xyl	1	20	16.6	83		61	148
		Decachlorobiphen	2	20	15.5	77		57	171
		Tetrachloro-m-xyl	2	20	16.9	85		61	148
Q3092-02MS	291376MS	Decachlorobiphen	1	20	16.9	84		57	171
		Tetrachloro-m-xyl	1	20	16.4	82		61	148
		Decachlorobiphen	2	20	15.3	77		57	171
		Tetrachloro-m-xyl	2	20	17.4	87		61	148
Q3092-02MSD	291376MSD	Decachlorobiphen	1	20	16.9	85		57	171
		Tetrachloro-m-xyl	1	20	16.5	82		61	148
		Decachlorobiphen	2	20	15.1	76		57	171
		Tetrachloro-m-xyl	2	20	17.1	86		61	148
PB169754TB	PB169754TB	Decachlorobiphen	1	20	17.8	89		57	171
		Tetrachloro-m-xyl	1	20	16.2	81		61	148
		Decachlorobiphen	2	20	16.2	81		57	171
		Tetrachloro-m-xyl	2	20	14.5	73		61	148
I.BLK-PL097339.D	PIBLK-PL097339.D	Decachlorobiphen	1	20	21.0	105		57	171
		Tetrachloro-m-xyl	1	20	18.3	91		61	148
		Decachlorobiphen	2	20	19.2	96		57	171
		Tetrachloro-m-xyl	2	20	18.9	95		61	148
Q3157-02	20-DEAD-1	Decachlorobiphen	1	20	17.5	87		57	171
		Tetrachloro-m-xyl	1	20	16.8	84		61	148
		Decachlorobiphen	2	20	16.6	83		57	171
		Tetrachloro-m-xyl	2	20	17.4	87		61	148
I.BLK-PL097354.D	PIBLK-PL097354.D	Decachlorobiphen	1	20	20.8	104		57	171
		Tetrachloro-m-xyl	1	20	18.1	91		61	148
		Decachlorobiphen	2	20	17.4	87		57	171
		Tetrachloro-m-xyl	2	20	18.6	93		61	148
I.BLK-PL097357.D	PIBLK-PL097357.D	Decachlorobiphen	1	20	22.2	111		57	171
		Tetrachloro-m-xyl	1	20	19.7	98		61	148
		Decachlorobiphen	2	20	20.1	100		57	171
		Tetrachloro-m-xyl	2	20	19.3	97		61	148
PB169782BS	PB169782BS	Decachlorobiphen	1	20	18.1	91		57	171

Surrogate Summary

SDG No.: Q3157

Client: Always Available Construction LLC.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB169782BS	PB169782BS	Tetrachloro-m-xyl	1	20	17.0	85		61	148
		Decachlorobiphen	2	20	17.4	87		57	171
		Tetrachloro-m-xyl	2	20	17.4	87		61	148
I.BLK-PL097363.D	PIBLK-PL097363.D	Decachlorobiphen	1	20	21.4	107		57	171
		Tetrachloro-m-xyl	1	20	18.8	94		61	148
		Decachlorobiphen	2	20	20.4	102		57	171
		Tetrachloro-m-xyl	2	20	19.6	98		61	148

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Always Available Construction LLC.</u>	DataFile :	<u>PL097335.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits	
			Qual	RPD			Qual	Low	High	RPD	
Lab Sample ID:	Q3092-02MS	Client Sample ID:		291376MS							
	(Column 1)										
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92				60	152
	Heptachlor	5	0	4.50	ug/L	90				56	147
	Heptachlor epoxide	5	0	4.70	ug/L	94				77	143
	Endrin	5	0	4.40	ug/L	88				76	144
	Methoxychlor	5	0	4.20	ug/L	84				70	142
Lab Sample ID:	Q3092-02MS	Client Sample ID:		291376MS							
	(Column 2)										
	gamma-BHC (Lindane)	5	0	4.70	ug/L	94				60	152
	Heptachlor	5	0	4.50	ug/L	90				56	147
	Heptachlor epoxide	5	0	4.70	ug/L	94				77	143
	Endrin	5	0	4.50	ug/L	90				76	144
	Methoxychlor	5	0	4.00	ug/L	80				70	142

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Always Available Construction LLC.</u>	DataFile :	<u>PL097336.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits		
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3092-02MSD	Client Sample ID:		291376MSD								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92		0		60	152	20
	Heptachlor	5	0	4.50	ug/L	90		0		56	147	20
	Heptachlor epoxide	5	0	4.70	ug/L	94		0		77	143	20
	Endrin	5	0	4.50	ug/L	90		2		76	144	20
	Methoxychlor	5	0	4.20	ug/L	84		0		70	142	20
Lab Sample ID:	Q3092-02MSD	Client Sample ID:		291376MSD								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92		2		60	152	20
	Heptachlor	5	0	4.40	ug/L	88		2		56	147	20
	Heptachlor epoxide	5	0	4.60	ug/L	92		2		77	143	20
	Endrin	5	0	4.50	ug/L	90		0		76	144	20
	Methoxychlor	5	0	4.00	ug/L	80		0		70	142	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3157

Analytical Method: 8081B

Client: Always Available Construction LLC.

Datafile : PL097362.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169782BS (Column 1)	gamma-BHC (Lindane)	0.5	0.51	ug/L	102				82	129		
	Heptachlor	0.5	0.48	ug/L	97				79	127		
	Heptachlor epoxide	0.5	0.52	ug/L	104				81	124		
	Endrin	0.5	0.45	ug/L	91				81	128		
	Methoxychlor	0.5	0.42	ug/L	84				78	108		
PB169782BS (Column 2)	gamma-BHC (Lindane)	0.5	0.52	ug/L	104				82	129		
	Heptachlor	0.5	0.49	ug/L	97				79	127		
	Heptachlor epoxide	0.5	0.51	ug/L	102				81	124		
	Endrin	0.5	0.45	ug/L	91				81	128		
	Methoxychlor	0.5	0.41	ug/L	82				78	108		

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB169782BL

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Lab Sample ID: PB169782BL

Lab File ID: PL097331.D

Matrix: (soil/water) water

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 09/22/2025

Date Analyzed (1): 09/22/2025

Date Analyzed (2): 09/22/2025

Time Analyzed (1): 18:52

Time Analyzed (2): 18:52

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
291376MS	Q3092-02MS	PL097335.D	09/22/2025	09/22/2025
291376MSD	Q3092-02MSD	PL097336.D	09/22/2025	09/22/2025
PB169754TB	PB169754TB	PL097338.D	09/22/2025	09/22/2025
20-DEAD-1	Q3157-02	PL097343.D	09/22/2025	09/22/2025
PB169782BS	PB169782BS	PL097362.D	09/23/2025	09/23/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	
Project:	Yard Soil Cleanup	Date Received:	09/22/25
Client Sample ID:	PB169754TB	SDG No.:	Q3157
Lab Sample ID:	PB169754TB	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	100 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097338.D	1	09/22/25 10:05	09/22/25 20:28	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.8		57 - 171	89%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.2		61 - 148	81%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097338.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 20:28
 Operator : AR\AJ
 Sample : PB169754TB
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169754TB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:48:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	73237704	95483252	16.177	14.521
28)	SA Decachlor...	9.000	7.983	58357317	92253200	17.834	16.209

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097338.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:28
Operator : AR\AJ
Sample : PB169754TB
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169754TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 73237704
Conc: 16.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169754TB




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 95483252
Conc: 14.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.000 min
Delta R.T.: -0.001 min
Response: 58357317
Conc: 17.83 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 92253200
Conc: 16.21 ng/ml

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	09/19/25
Project:	Yard Soil Cleanup	Date Received:	09/19/25
Client Sample ID:	20-DEAD-1	SDG No.:	Q3157
Lab Sample ID:	Q3157-02	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097343.D	1	09/22/25 10:05	09/22/25 22:03	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.037	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.027	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.096	U	0.096	0.50	ug/L
72-20-8	Endrin	0.032	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.11	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.5		57 - 171	87%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.4		61 - 148	87%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097343.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 22:03
Operator : AR\AJ
Sample : Q3157-02
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
20-DEAD-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.825	75812758	114.8E6	16.746	17.455
28)	SA Decachlor...	8.999	7.983	57171406	94429346	17.472	16.591

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097343.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 22:03
Operator : AR\AJ
Sample : Q3157-02
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
20-DEAD-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.528 min
Delta R.T.: 0.001 min
Response: 75812758
Conc: 16.75 ng/ml

Instrument :
ECD_L
ClientSampleId :
20-DEAD-1




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: 0.001 min
Response: 114774758
Conc: 17.45 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: -0.002 min
Response: 57171406
Conc: 17.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 94429346
Conc: 16.59 ng/ml



CALIBRATION SUMMARY

Calibration Times: **17:06** **18:00**

LAB FILE ID:	RT 100 =	<u>PL097271.D</u>	RT 075 =	<u>PL097272.D</u>
	RT 050 =	PL097273.D	RT 025 =	PL097274.D
			RT 005 =	PL097275.D

[illegible]

Calibration Times: **17:06** **18:00**

LAB FILE ID:	RT 100 =	<u>PL097271.D</u>	RT 075 =	<u>PL097272.D</u>
	RT 050 =	PL097273.D	RT 025 =	PL097274.D
			RT 005 =	PL097275.D

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Instrument ID: ECD_L

Calibration Date(s): 09/18/2025 09/18/2025

Calibration Times: 17:06 18:00

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:								CF 100 =	<u>PL097271.D</u>	CF 075 =	<u>PL097272.D</u>	
CF 050 =								<u>PL097273.D</u>	CF 025 =	<u>PL097274.D</u>	CF 005 =	<u>PL097275.D</u>
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD					
Decachlorobiphenyl	3151670000	3121880000	3187620000	3242730000	3657150000	3272210000	7					
Endrin	4777540000	4640970000	4618310000	4356250000	4672620000	4613140000	3					
gamma-BHC (Lindane)	6940420000	6699850000	6593990000	6320490000	6510990000	6613150000	3					
Heptachlor	6809100000	6621120000	6608040000	6426970000	6885270000	6670100000	3					
Heptachlor epoxide	5764160000	5655320000	5756800000	5829540000	5728960000	5746960000	1					
Methoxychlor	2052960000	2027370000	2062620000	2055200000	2228740000	2085380000	4					
Tetrachloro-m-xylene	4554070000	4459430000	4465020000	4402740000	4754560000	4527160000	3					



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Instrument ID: ECD_L

Calibration Date(s): 09/18/2025 09/18/2025

Calibration Times: 17:06 18:00

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID:								CF 100 =	<u>PL097271.D</u>	CF 075 =	<u>PL097272.D</u>	
CF 050 =								<u>PL097273.D</u>	CF 025 =	<u>PL097274.D</u>	CF 005 =	<u>PL097275.D</u>
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD					
Decachlorobiphenyl	5360550000	5307930000	5459740000	5623180000	6706450000	5691570000	10					
Endrin	7450260000	7335750000	7338920000	7289700000	8137930000	7510510000	5					
gamma-BHC (Lindane)	9632950000	9373760000	9300450000	8991300000	9274840000	9314660000	2					
Heptachlor	9299100000	9162550000	9216060000	9078210000	9769790000	9305140000	3					
Heptachlor epoxide	7825700000	7749420000	7831840000	7763530000	8487220000	7931540000	4					
Methoxychlor	3580590000	3582010000	3698660000	3768400000	4217720000	3769480000	7					
Tetrachloro-m-xylene	6546800000	6457210000	6449880000	6384190000	7039690000	6575560000	4					



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
Instrument ID: ECD_L **Date(s) Analyzed:** 09/18/2025 09/18/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.68	4.58	4.78	206851000
		2	5.20	5.10	5.30	216096000
		3	5.91	5.81	6.01	813754000
		4	5.99	5.89	6.09	992045000
		5	6.83	6.73	6.93	162879000
Toxaphene	500	1	6.20	6.10	6.30	42075500
		2	6.60	6.50	6.70	33847100
		3	7.01	6.91	7.11	141900000
		4	7.11	7.01	7.21	105792000
		5	7.88	7.78	7.98	74166200



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
Instrument ID: ECD_L **Date(s) Analyzed:** 09/18/2025 09/18/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.83	3.73	3.93	252157000
		2	4.41	4.31	4.51	284305000
		3	5.04	4.94	5.14	877635000
		4	5.11	5.01	5.21	766472000
		5	6.00	5.90	6.10	292534000
Toxaphene	500	1	5.07	4.97	5.17	42482700
		2	5.75	5.65	5.85	54918500
		3	6.03	5.93	6.13	57955900
		4	6.67	6.57	6.77	177984000
		5	7.11	7.01	7.21	109140000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097271.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:06
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.823	455.4E6	654.7E6	101.994	101.503
28)	SA Decachlor...	9.001	7.984	315.2E6	536.1E6	98.872	98.183
Target Compounds							
2)	A alpha-BHC	3.974	3.329	743.3E6	1044.5E6	105.888	104.303
3)	MA gamma-BHC...	4.302	3.660	694.0E6	963.3E6	105.254	103.575
4)	MA Heptachlor	4.893	4.008	680.9E6	929.9E6	103.043	100.901
5)	MB Aldrin	5.232	4.290	655.2E6	886.1E6	104.089	102.301
6)	B beta-BHC	4.489	3.956	269.2E6	397.0E6	100.455	100.557
7)	B delta-BHC	4.735	4.189	640.9E6	958.6E6	104.552	103.832
8)	B Heptachlo...	5.652	4.793	576.4E6	782.6E6	100.128	99.922
9)	A Endosulfan I	6.033	5.163	524.9E6	738.7E6	102.490	99.419
10)	B gamma-Chl...	5.905	5.044	579.3E6	855.4E6	103.252	99.860
11)	B alpha-Chl...	5.986	5.108	572.8E6	827.5E6	102.408	100.361
12)	B 4,4'-DDE	6.156	5.297	496.4E6	791.1E6	104.790	102.753
13)	MA Dieldrin	6.306	5.428	566.8E6	825.1E6	103.988	101.465
14)	MA Endrin	6.532	5.702	477.8E6	745.0E6	103.448	101.517
15)	B Endosulfa...	6.745	5.994	469.5E6	697.8E6	98.323	100.299
16)	A 4,4'-DDD	6.665	5.850	389.1E6	658.9E6	103.408	103.640
17)	MA 4,4'-DDT	6.979	6.102	440.3E6	719.0E6	103.710	103.115
18)	B Endrin al...	6.874	6.172	303.6E6	476.2E6	99.226	96.896
19)	B Endosulfa...	7.108	6.396	408.4E6	666.5E6	101.600	99.231
20)	A Methoxychlor	7.452	6.674	205.3E6	358.1E6	99.532	96.808
21)	B Endrin ke...	7.587	6.900	429.6E6	722.3E6	101.695	97.725
22)	Mirex	8.065	7.089	330.8E6	552.6E6	97.189	96.302



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:06
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097272.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:19
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.824	334.5E6	484.3E6	74.906	75.085
2)	SA Decachlor...	9.002	7.985	234.1E6	398.1E6	73.453	72.915
Target Compounds							
2)	A alpha-BHC	3.975	3.329	536.3E6	757.5E6	76.392	75.640
3)	MA gamma-BHC...	4.302	3.661	502.5E6	703.0E6	76.204	75.591
4)	MA Heptachlor	4.893	4.009	496.6E6	687.2E6	75.148	74.565
5)	MB Aldrin	5.232	4.291	477.7E6	650.5E6	75.880	75.100
6)	B beta-BHC	4.490	3.957	198.1E6	292.5E6	73.934	74.080
7)	B delta-BHC	4.735	4.190	463.0E6	697.0E6	75.527	75.496
8)	B Heptachlo...	5.652	4.793	424.1E6	581.2E6	73.678	74.211
9)	A Endosulfan I	6.034	5.163	383.9E6	550.9E6	74.971	74.138
10)	B gamma-Chl...	5.906	5.045	423.9E6	636.8E6	75.554	74.339
11)	B alpha-Chl...	5.986	5.109	419.8E6	614.2E6	75.053	74.493
12)	B 4,4'-DDE	6.156	5.297	361.5E6	581.0E6	76.313	75.459
13)	MA Dieldrin	6.306	5.428	412.9E6	609.0E6	75.745	74.885
14)	MA Endrin	6.533	5.703	348.1E6	550.2E6	75.368	74.968
15)	B Endosulfa...	6.745	5.995	347.8E6	518.4E6	72.847	74.509
16)	A 4,4'-DDD	6.665	5.850	284.1E6	483.2E6	75.512	76.000
17)	MA 4,4'-DDT	6.980	6.102	320.6E6	526.7E6	75.521	75.530
18)	B Endrin al...	6.875	6.173	224.8E6	356.8E6	73.496	72.597
19)	B Endosulfa...	7.108	6.396	299.3E6	496.8E6	74.448	73.965
20)	A Methoxychlor	7.453	6.674	152.1E6	268.7E6	73.718	72.634
21)	B Endrin ke...	7.588	6.900	314.3E6	536.5E6	74.403	72.588
22)	Mirex	8.066	7.090	246.6E6	415.4E6	72.461	72.394



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:19
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:33
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.824	223.3E6	322.5E6	50.000	50.000
28)	SA Decachlor...	9.002	7.984	159.4E6	273.0E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.975	3.329	351.0E6	500.7E6	50.000	50.000
3)	MA gamma-BHC...	4.302	3.661	329.7E6	465.0E6	50.000	50.000
4)	MA Heptachlor	4.893	4.009	330.4E6	460.8E6	50.000	50.000
5)	MB Aldrin	5.232	4.291	314.7E6	433.1E6	50.000	50.000
6)	B beta-BHC	4.490	3.957	134.0E6	197.4E6	50.000	50.000
7)	B delta-BHC	4.735	4.190	306.5E6	461.6E6	50.000	50.000
8)	B Heptachlo...	5.652	4.793	287.8E6	391.6E6	50.000	50.000
9)	A Endosulfan I	6.034	5.164	256.1E6	371.5E6	50.000	50.000
10)	B gamma-Chl...	5.906	5.045	280.5E6	428.3E6	50.000	50.000
11)	B alpha-Chl...	5.986	5.109	279.7E6	412.3E6	50.000	50.000
12)	B 4,4'-DDE	6.156	5.297	236.9E6	385.0E6	50.000	50.000
13)	MA Dieldrin	6.306	5.428	272.5E6	406.6E6	50.000	50.000
14)	MA Endrin	6.532	5.703	230.9E6	366.9E6	50.000	50.000
15)	B Endosulfa...	6.746	5.995	238.7E6	347.8E6	50.000	50.000
16)	A 4,4'-DDD	6.666	5.850	188.1E6	317.9E6	50.000	50.000
17)	MA 4,4'-DDT	6.980	6.102	212.3E6	348.6E6	50.000	50.000
18)	B Endrin al...	6.874	6.173	153.0E6	245.7E6	50.000	50.000
19)	B Endosulfa...	7.107	6.396	201.0E6	335.8E6	50.000	50.000
20)	A Methoxychlor	7.452	6.674	103.1E6	184.9E6	50.000	50.000
21)	B Endrin ke...	7.587	6.900	211.2E6	369.6E6	50.000	50.000
22)	Mirex	8.065	7.090	170.2E6	286.9E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:33
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097274.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:47
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:45:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	110.1E6	159.6E6	24.651	24.745
28)	SA Decachlor...	9.001	7.984	81068137	140.6E6	25.432	25.748
Target Compounds							
2)	A alpha-BHC	3.974	3.329	165.9E6	238.9E6	23.631	23.856
3)	MA gamma-BHC...	4.302	3.661	158.0E6	224.8E6	23.963	24.169
4)	MA Heptachlor	4.893	4.009	160.7E6	227.0E6	24.315	24.626
5)	MB Aldrin	5.232	4.291	151.8E6	210.1E6	24.117	24.257
6)	B beta-BHC	4.489	3.957	66839868	99596935	24.946	25.226
7)	B delta-BHC	4.735	4.190	146.9E6	220.9E6	23.960	23.922
8)	B Heptachlo...	5.652	4.793	145.7E6	194.1E6	25.316	24.782
9)	A Endosulfan I	6.034	5.164	125.0E6	186.4E6	24.417	25.081
10)	B gamma-Chl...	5.906	5.045	136.6E6	215.6E6	24.338	25.170
11)	B alpha-Chl...	5.986	5.109	137.3E6	206.1E6	24.543	24.992
12)	B 4,4'-DDE	6.156	5.297	112.8E6	185.2E6	23.816	24.048
13)	MA Dieldrin	6.306	5.428	131.4E6	197.7E6	24.115	24.315
14)	MA Endrin	6.532	5.702	108.9E6	182.2E6	23.581	24.832
15)	B Endosulfa...	6.746	5.995	118.2E6	173.5E6	24.764	24.933
16)	A 4,4'-DDD	6.665	5.850	89516591	156.6E6	23.790	24.632
17)	MA 4,4'-DDT	6.979	6.102	101.6E6	167.4E6	23.933	24.009
18)	B Endrin al...	6.875	6.173	75911729	128.9E6	24.814	26.230
19)	B Endosulfa...	7.108	6.396	98950148	168.3E6	24.614	25.062
20)	A Methoxychlor	7.453	6.675	51380116	94210076	24.910	25.471
21)	B Endrin ke...	7.587	6.900	102.9E6	184.8E6	24.367	25.000
22)	Mirex	8.065	7.090	87774467	148.3E6	25.790	25.847



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:47
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:45:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097275.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 18:00
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC005

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/19/2025
 Supervised By : mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:45:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.824	23772791	35198466	5.324	5.457
2) SA Decachlor...	9.001	7.984	18285769	33532264	5.737	6.142
Target Compounds						
2) A alpha-BHC	3.974	3.328	33487008	48361581	4.770	4.829
3) MA gamma-BHC...	4.302	3.660	32554930	46374195	4.937	4.986
4) MA Heptachlor	4.893	4.008	34426346	48848941	5.210	5.300
5) MB Aldrin	5.232	4.290	31901901	43982840	5.068	5.078
6) B beta-BHC	4.490	3.957	14725134	22540181	5.496	5.709
7) B delta-BHC	4.733	4.189	28774806	45070449	4.694m	4.882
8) B Heptachlo...	5.650	4.793	28644801	42436094	4.976m	5.418
9) A Endosulfan I	6.033	5.163	27081618	44002825	5.288	5.922
10) B gamma-Chl...	5.905	5.043	28749631	47175597	5.124	5.507m
11) B alpha-Chl...	5.986	5.107	29943592	42753742	5.353	5.185m
12) B 4,4'-DDE	6.156	5.297	23948455	38300213	5.056	4.974
13) MA Dieldrin	6.305	5.428	27648945	42357479	5.073	5.209
14) MA Endrin	6.532	5.702	23363118	40689632	5.059	5.544
15) B Endosulfa...	6.744	5.994	28776249	38809493	6.027m	5.579
16) A 4,4'-DDD	6.665	5.850	19149707	32954369	5.089	5.183
17) MA 4,4'-DDT	6.979	6.102	20744145	33603734	4.886	4.819
18) B Endrin al...	6.874	6.171	16425107	35487549	5.369	7.221m#
19) B Endosulfa...	7.108	6.396	21803103	37554894	5.424	5.592
20) A Methoxychlor	7.452	6.674	11143676	21088586	5.403	5.702
21) B Endrin ke...	7.587	6.899	21956995	42459706	5.197	5.745m
22) Mirex	8.065	7.090	21205020	35794871	6.230	6.238

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 18:00
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PSTDICC005

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/19/2025
Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:45:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097278.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 18:41
 Operator : AR\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:25:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:24:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.823	181.2E6	350.3E6	50.000	50.000
28)	SA Decachlor...	9.002	7.984	130.8E6	233.3E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.680	3.832	103.4E6	126.1E6	500.000	500.000
24)	Chlordane-2	5.204	4.412	108.0E6	142.2E6	500.000	500.000
25)	Chlordane-3	5.907	5.044	406.9E6	438.8E6	500.000	500.000
26)	Chlordane-4	5.991	5.108	496.0E6	383.2E6	500.000	500.000
27)	Chlordane-5	6.828	6.002	81439283	146.3E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 18:41
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:25:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:24:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097283.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 19:49
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 06:18:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 06:17:16 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
 Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.824	212.8E6	316.1E6	50.000	50.000
7) SA Decachlor...	9.001	7.985	151.2E6	274.7E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	6.202	5.066	21037740	21241348	500.000	500.000
3) Toxaphene-2	6.601	5.752	16923557	27459266	500.000	500.000
4) Toxaphene-3	7.014	6.033	70949766	28977953	500.000	500.000
5) Toxaphene-4	7.105	6.667	52895758	88991768	500.000	500.000
6) Toxaphene-5	7.883	7.107	37083116	54569782	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 19:49
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 06:18:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 06:17:16 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097286.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 08:39
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL092025

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/22/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 09:03:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.820	233.1E6	328.4E6	51.482	49.938
28)	SA Decachlor...	9.006	7.985	165.2E6	255.3E6	50.482	44.851
Target Compounds							
2)	A alpha-BHC	3.974	3.325	368.4E6	500.7E6	52.723	50.281
3)	MA gamma-BHC...	4.302	3.657	344.6E6	461.1E6	52.110	49.498
4)	MA Heptachlor	4.893	4.005	349.0E6	445.3E6	52.326	47.851
5)	MB Aldrin	5.233	4.288	325.9E6	418.1E6	51.460	48.168
6)	B beta-BHC	4.490	3.954	136.9E6	187.9E6	50.230	46.246
7)	B delta-BHC	4.734	4.186	321.2E6	449.4E6	52.930m	48.893
8)	B Heptachlo...	5.651	4.789	306.1E6	376.2E6	53.263m	47.430m
9)	A Endosulfan I	6.034	5.161	264.5E6	345.6E6	51.051m	44.982
10)	B gamma-Chl...	5.905	5.042	288.6E6	406.9E6	51.050m	46.594
11)	B alpha-Chl...	5.986	5.107	285.1E6	401.4E6	50.192m	48.364
12)	B 4,4'-DDE	6.156	5.294	248.9E6	367.5E6	52.230m	47.821m
13)	MA Dieldrin	6.307	5.424	282.4E6	392.3E6	51.517	47.972m
14)	MA Endrin	6.532	5.701	241.9E6	351.7E6	52.445m	46.833
15)	B Endosulfa...	6.745	5.993	240.2E6	329.4E6	48.828m	46.333
16)	A 4,4'-DDD	6.666	5.848	191.1E6	297.9E6	50.687m	46.193
17)	MA 4,4'-DDT	6.982	6.101	229.3E6	328.0E6	54.236	47.389
18)	B Endrin al...	6.876	6.171	175.3E6	229.7E6	56.862	43.033
19)	B Endosulfa...	7.110	6.395	209.3E6	310.2E6	51.264	45.293
20)	A Methoxychlor	7.455	6.673	103.8E6	168.0E6	49.793	44.559
21)	B Endrin ke...	7.590	6.898	220.0E6	395.4E6	51.828	52.514m
22)	Mirex	8.068	7.089	174.4E6	269.2E6	49.109	45.026

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2025 08:39
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

ICVPL092025

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/22/2025

Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 09:03:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097287.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 08:52
 Operator : AR\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL092025

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/22/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 09:12:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.822	199.7E6	365.8E6	44.121	55.633 #
28)	SA Decachlor...	9.001	7.983	139.6E6	237.0E6	42.668	41.636
Target Compounds							
23)	Chlordane-1	4.679	3.831	111.7E6	129.9E6	527.132	511.091
24)	Chlordane-2	5.201	4.410	111.1E6	150.3E6	500.634m	511.589
25)	Chlordane-3	5.905	5.043	424.6E6	455.7E6	520.618m	509.555
26)	Chlordane-4	5.989	5.106	515.0E6	380.7E6	514.915m	499.841m
27)	Chlordane-5	6.826	6.001	87854088	149.8E6	524.084m	502.416

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097287.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2025 08:52
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
ICVPL092025

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/22/2025
Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 09:12:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 09:25
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL092025

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/22/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 09:41:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 06:34:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
 Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.822	206.8E6	305.4E6	49.043	49.125
7) SA Decachlor...	9.004	7.985	146.0E6	259.8E6	48.901	47.512
Target Compounds						
2) Toxaphene-1	6.203	5.064	20584370	20730281	478.691	527.108m
3) Toxaphene-2	6.602	5.752	15850196	25914739	475.278	475.930
4) Toxaphene-3	7.016	6.033	66126700	27574610	472.119	485.585
5) Toxaphene-4	7.106	6.667	50737974	89589604	488.063	520.666
6) Toxaphene-5	7.885	7.107	34706442	51873806	481.757	482.163

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2025 09:25
Operator : AR\AJ
Sample : PTOXICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
ICVPL092025

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/22/2025
Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 09:41:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\LTX092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 06:34:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticide 1 Signal #2 Phase: Rtx-CLPesticide 1
Signal #1 Info : 30M x 0.32mm x0.3 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025 09/18/2025

Continuing Calib Time: 18:11

Initial Calibration Time(s): 17:06 18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 18:11

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.001	8.902	9.102	48.730	50.000	-2.5
Endrin	6.531	6.432	6.632	46.480	50.000	-7.0
gamma-BHC (Lindane)	4.302	4.202	4.402	51.410	50.000	2.8
Heptachlor	4.892	4.793	4.993	49.840	50.000	-0.3
Heptachlor epoxide	5.651	5.552	5.752	50.370	50.000	0.7
Methoxychlor	7.452	7.352	7.552	44.040	50.000	-11.9
Tetrachloro-m-xylene	3.527	3.428	3.628	51.250	50.000	2.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097328.D **Time Analyzed:** 18:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.984	7.884	8.084	40.500	50.000	-19.0
Endrin	5.702	5.603	5.803	44.550	50.000	-10.9
gamma-BHC (Lindane)	3.661	3.561	3.761	49.630	50.000	-0.7
Heptachlor	4.009	3.909	4.109	48.080	50.000	-3.8
Heptachlor epoxide	4.792	4.693	4.893	48.930	50.000	-2.1
Methoxychlor	6.674	6.574	6.774	41.040	50.000	-17.9
Tetrachloro-m-xylene	2.824	2.724	2.924	49.030	50.000	-1.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097328.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 18:11
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:47:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	232.0E6	322.4E6	51.248	49.032
28)	SA Decachlor...	9.001	7.984	159.5E6	230.5E6	48.729	40.496
Target Compounds							
2)	A alpha-BHC	3.974	3.329	362.2E6	499.4E6	51.831	50.157
3)	MA gamma-BHC...	4.302	3.661	340.0E6	462.3E6	51.413	49.634
4)	MA Heptachlor	4.892	4.009	332.4E6	447.4E6	49.840	48.077
5)	MB Aldrin	5.232	4.291	318.6E6	428.3E6	50.303	49.349
6)	B beta-BHC	4.489	3.957	141.1E6	196.9E6	51.762	48.472
7)	B delta-BHC	4.733	4.189	324.6E6	457.6E6	53.487m	49.783
8)	B Heptachlo...	5.651	4.792	289.5E6	388.1E6	50.367	48.925
9)	A Endosulfan I	6.033	5.163	252.6E6	363.4E6	48.750	47.298
10)	B gamma-Chl...	5.905	5.044	282.8E6	425.3E6	50.028	48.697
11)	B alpha-Chl...	5.986	5.108	277.1E6	406.0E6	48.784	48.919
12)	B 4,4'-DDE	6.156	5.297	240.2E6	372.2E6	50.406	48.428
13)	MA Dieldrin	6.305	5.427	267.8E6	395.0E6	48.858	48.307
14)	MA Endrin	6.531	5.702	214.4E6	334.6E6	46.481	44.546
15)	B Endosulfa...	6.745	5.994	230.4E6	328.0E6	46.849	46.131
16)	A 4,4'-DDD	6.664	5.849	196.3E6	311.7E6	52.066	48.328
17)	MA 4,4'-DDT	6.979	6.102	192.9E6	303.4E6	45.633	43.844
18)	B Endrin al...	6.874	6.172	149.5E6	235.9E6	48.491	44.193
19)	B Endosulfa...	7.107	6.396	196.2E6	311.1E6	48.060	45.418
20)	A Methoxychlor	7.452	6.674	91835427	154.7E6	44.038	41.036
21)	B Endrin ke...	7.587	6.899	207.6E6	333.6E6	48.909	44.299
22)	Mirex	8.065	7.089	158.3E6	248.2E6	44.594	41.509

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097328.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 18:11
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:47:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/22/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 21:08

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY**Lab Name:** Alliance**Contract:** ALWA01**Lab Code:** ACE**SDG NO.:** Q3157**Continuing Calib Date:** 09/22/2025**Initial Calibration Date(s):** 09/18/202509/18/2025**Continuing Calib Time:** 21:08**Initial Calibration Time(s):** 17:0618:00**GC Column:** ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.999	8.902	9.102	53.620	50.000	7.2
Endrin	6.531	6.432	6.632	50.770	50.000	1.5
gamma-BHC (Lindane)	4.302	4.202	4.402	55.280	50.000	10.6
Heptachlor	4.893	4.793	4.993	53.040	50.000	6.1
Heptachlor epoxide	5.651	5.552	5.752	54.390	50.000	8.8
Methoxychlor	7.451	7.352	7.552	48.210	50.000	-3.6
Tetrachloro-m-xylene	3.527	3.428	3.628	54.690	50.000	9.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097341.D **Time Analyzed:** 21:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	48.470	50.000	-3.1
Endrin	5.702	5.603	5.803	50.690	50.000	1.4
gamma-BHC (Lindane)	3.661	3.561	3.761	54.800	50.000	9.6
Heptachlor	4.009	3.909	4.109	52.810	50.000	5.6
Heptachlor epoxide	4.792	4.693	4.893	53.620	50.000	7.2
Methoxychlor	6.673	6.574	6.774	46.500	50.000	-7.0
Tetrachloro-m-xylene	2.824	2.724	2.924	53.880	50.000	7.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097341.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 21:08
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:49:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	247.6E6	354.3E6	54.690	53.880
2)	SA Decachlor...	8.999	7.983	175.5E6	275.9E6	53.622	48.471
Target Compounds							
2)	A alpha-BHC	3.974	3.329	389.8E6	551.3E6	55.786	55.367
3)	MA gamma-BHC...	4.302	3.661	365.6E6	510.4E6	55.285	54.799
4)	MA Heptachlor	4.893	4.009	353.8E6	491.4E6	53.038	52.810
5)	MB Aldrin	5.232	4.291	344.0E6	472.6E6	54.317	54.449
6)	B beta-BHC	4.489	3.957	150.7E6	215.8E6	55.262	53.136
7)	B delta-BHC	4.733	4.190	345.8E6	503.1E6	56.983m	54.736
8)	B Heptachlo...	5.651	4.792	312.6E6	425.3E6	54.386	53.618
9)	A Endosulfan I	6.033	5.163	279.5E6	400.2E6	53.953	52.087
10)	B gamma-Chl...	5.905	5.044	310.6E6	467.2E6	54.934	53.497
11)	B alpha-Chl...	5.985	5.108	305.6E6	447.6E6	53.810	53.929
12)	B 4,4'-DDE	6.155	5.295	267.1E6	415.5E6	56.064	54.073m
13)	MA Dieldrin	6.305	5.428	294.7E6	441.9E6	53.759	54.044
14)	MA Endrin	6.531	5.702	234.2E6	380.7E6	50.768	50.690
15)	B Endosulfa...	6.744	5.994	252.8E6	372.7E6	51.394	52.423
16)	A 4,4'-DDD	6.664	5.849	216.4E6	358.1E6	57.382	55.533
17)	MA 4,4'-DDT	6.978	6.101	211.6E6	342.2E6	50.061	49.448
18)	B Endrin al...	6.873	6.172	166.4E6	267.4E6	53.991	50.099
19)	B Endosulfa...	7.106	6.395	216.7E6	352.3E6	53.088	51.427
20)	A Methoxychlor	7.451	6.673	100.5E6	175.3E6	48.210	46.501
21)	B Endrin ke...	7.586	6.899	232.0E6	404.6E6	54.663	53.724
22)	Mirex	8.064	7.089	175.9E6	296.9E6	49.552	49.664

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097341.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 21:08
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025 09/18/2025

Continuing Calib Time: 00:46

Initial Calibration Time(s): 17:06 18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 00:46

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097355.D **Time Analyzed:** 00:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.999	8.902	9.102	51.350	50.000	2.7
Endrin	6.531	6.432	6.632	48.870	50.000	-2.3
gamma-BHC (Lindane)	4.301	4.202	4.402	54.300	50.000	8.6
Heptachlor	4.892	4.793	4.993	51.040	50.000	2.1
Heptachlor epoxide	5.651	5.552	5.752	52.630	50.000	5.3
Methoxychlor	7.451	7.352	7.552	44.570	50.000	-10.9
Tetrachloro-m-xylene	3.527	3.428	3.628	54.430	50.000	8.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097355.D **Time Analyzed:** 00:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	45.360	50.000	-9.3
Endrin	5.702	5.603	5.803	47.950	50.000	-4.1
gamma-BHC (Lindane)	3.661	3.561	3.761	54.250	50.000	8.5
Heptachlor	4.008	3.909	4.109	51.270	50.000	2.5
Heptachlor epoxide	4.792	4.693	4.893	52.550	50.000	5.1
Methoxychlor	6.673	6.574	6.774	41.840	50.000	-16.3
Tetrachloro-m-xylene	2.824	2.724	2.924	53.670	50.000	7.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097355.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 00:46
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:51:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	246.4E6	352.9E6	54.430	53.667
28)	SA Decachlor...	8.999	7.983	168.0E6	258.2E6	51.345	45.360
Target Compounds							
2)	A alpha-BHC	3.974	3.329	385.7E6	547.8E6	55.205	55.011
3)	MA gamma-BHC...	4.301	3.661	359.1E6	505.4E6	54.296	54.255
4)	MA Heptachlor	4.892	4.008	340.5E6	477.0E6	51.042	51.266
5)	MB Aldrin	5.232	4.290	335.1E6	465.2E6	52.912	53.598
6)	B beta-BHC	4.489	3.957	148.2E6	213.5E6	54.367	52.550
7)	B delta-BHC	4.733	4.189	338.2E6	494.7E6	55.740m	53.812
8)	B Heptachlo...	5.651	4.792	302.5E6	416.8E6	52.630	52.555
9)	A Endosulfan I	6.032	5.163	269.4E6	386.9E6	52.002	50.349
10)	B gamma-Chl...	5.905	5.044	300.6E6	453.9E6	53.175	51.965
11)	B alpha-Chl...	5.985	5.108	295.1E6	433.6E6	51.958	52.232
12)	B 4,4'-DDE	6.156	5.297	258.7E6	399.5E6	54.289	51.982
13)	MA Dieldrin	6.305	5.428	286.3E6	424.7E6	52.216	51.938
14)	MA Endrin	6.531	5.702	225.4E6	360.1E6	48.869	47.953
15)	B Endosulfa...	6.744	5.994	244.9E6	354.6E6	49.800	49.877
16)	A 4,4'-DDD	6.664	5.849	214.9E6	344.9E6	56.990	53.475
17)	MA 4,4'-DDT	6.978	6.101	195.4E6	306.4E6	46.235	44.275
18)	B Endrin al...	6.873	6.172	160.8E6	247.7E6	52.174	46.416
19)	B Endosulfa...	7.107	6.395	210.4E6	332.9E6	51.530	48.603
20)	A Methoxychlor	7.451	6.673	92934830	157.7E6	44.565	41.836
21)	B Endrin ke...	7.586	6.899	223.1E6	366.2E6	52.564	48.627
22)	Mirex	8.064	7.088	168.3E6	274.7E6	47.413	45.941

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097355.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 00:46
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:51:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 10:03

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 10:03

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097359.D **Time Analyzed:** 10:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.003	8.902	9.102	53.250	50.000	6.5
Endrin	6.533	6.432	6.632	49.330	50.000	-1.3
gamma-BHC (Lindane)	4.302	4.202	4.402	54.410	50.000	8.8
Heptachlor	4.893	4.793	4.993	50.970	50.000	1.9
Heptachlor epoxide	5.652	5.552	5.752	51.910	50.000	3.8
Methoxychlor	7.453	7.352	7.552	46.240	50.000	-7.5
Tetrachloro-m-xylene	3.527	3.428	3.628	52.820	50.000	5.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097359.D **Time Analyzed:** 10:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.884	8.084	49.070	50.000	-1.9
Endrin	5.702	5.603	5.803	47.520	50.000	-5.0
gamma-BHC (Lindane)	3.659	3.561	3.761	52.620	50.000	5.2
Heptachlor	4.007	3.909	4.109	50.390	50.000	0.8
Heptachlor epoxide	4.791	4.693	4.893	51.290	50.000	2.6
Methoxychlor	6.674	6.574	6.774	43.870	50.000	-12.3
Tetrachloro-m-xylene	2.822	2.724	2.924	52.390	50.000	4.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097359.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 10:03
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/24/2025
 Supervised By : mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.822	239.1E6	344.5E6	52.825	52.394
28)	SA Decachlor...	9.003	7.984	174.3E6	279.3E6	53.255	49.072
Target Compounds							
2)	A alpha-BHC	3.974	3.328	376.7E6	531.4E6	53.909	53.364
3)	MA gamma-BHC...	4.302	3.659	359.9E6	490.1E6	54.415	52.618
4)	MA Heptachlor	4.893	4.007	340.0E6	468.9E6	50.967	50.387
5)	MB Aldrin	5.232	4.289	331.7E6	453.2E6	52.377	52.212
6)	B beta-BHC	4.489	3.956	144.1E6	205.3E6	52.852	50.545
7)	B delta-BHC	4.734	4.188	329.9E6	478.4E6	54.360m	52.039
8)	B Heptachlo...	5.652	4.791	298.4E6	406.8E6	51.915	51.294
9)	A Endosulfan I	6.034	5.162	269.5E6	373.9E6	52.005	48.664
10)	B gamma-Chl...	5.906	5.043	297.1E6	428.9E6	52.544	49.103
11)	B alpha-Chl...	5.986	5.108	294.1E6	418.6E6	51.782	50.427
12)	B 4,4'-DDE	6.157	5.296	254.4E6	397.9E6	53.395	51.775
13)	MA Dieldrin	6.306	5.427	286.3E6	424.3E6	52.223	51.894
14)	MA Endrin	6.533	5.702	227.6E6	356.9E6	49.327	47.520
15)	B Endosulfa...	6.746	5.994	248.8E6	357.6E6	50.576	50.307
16)	A 4,4'-DDD	6.666	5.849	222.0E6	345.0E6	58.889	53.501
17)	MA 4,4'-DDT	6.981	6.102	201.6E6	324.2E6	47.680	46.843
18)	B Endrin al...	6.875	6.172	163.7E6	261.0E6	53.109	48.893
19)	B Endosulfa...	7.109	6.396	212.2E6	338.0E6	51.969	49.339
20)	A Methoxychlor	7.453	6.674	96417789	165.4E6	46.235	43.867
21)	B Endrin ke...	7.589	6.900	229.8E6	386.0E6	54.142	51.254
22)	Mirex	8.066	7.089	172.1E6	289.3E6	48.470	48.388

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097359.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:03
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/24/2025

Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025 09/18/2025

Continuing Calib Time: 12:52

Initial Calibration Time(s): 17:06 18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endrin	6.53	6.53	6.43	6.63	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Continuing Calib Date: 09/23/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 12:52

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097364.D **Time Analyzed:** 12:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.004	8.902	9.102	52.350	50.000	4.7
Endrin	6.533	6.432	6.632	49.760	50.000	-0.5
gamma-BHC (Lindane)	4.301	4.202	4.402	53.420	50.000	6.8
Heptachlor	4.892	4.793	4.993	50.770	50.000	1.5
Heptachlor epoxide	5.652	5.552	5.752	51.580	50.000	3.2
Methoxychlor	7.454	7.352	7.552	45.610	50.000	-8.8
Tetrachloro-m-xylene	3.526	3.428	3.628	53.100	50.000	6.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALWA01
Lab Code: ACE **SDG NO.:** Q3157
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/23/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097364.D **Time Analyzed:** 12:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.884	8.084	48.450	50.000	-3.1
Endrin	5.700	5.603	5.803	47.320	50.000	-5.4
gamma-BHC (Lindane)	3.657	3.561	3.761	52.520	50.000	5.0
Heptachlor	4.005	3.909	4.109	49.830	50.000	-0.3
Heptachlor epoxide	4.790	4.693	4.893	51.900	50.000	3.8
Methoxychlor	6.673	6.574	6.774	43.640	50.000	-12.7
Tetrachloro-m-xylene	2.820	2.724	2.924	52.450	50.000	4.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 12:52
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/24/2025
 Supervised By : mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.820	240.4E6	344.9E6	53.103	52.451
28)	SA Decachlor...	9.004	7.983	171.3E6	275.8E6	52.353	48.452
Target Compounds							
2)	A alpha-BHC	3.973	3.325	378.2E6	528.3E6	54.130	53.056
3)	MA gamma-BHC...	4.301	3.657	353.3E6	489.2E6	53.419	52.520
4)	MA Heptachlor	4.892	4.005	338.7E6	463.7E6	50.774	49.832
5)	MB Aldrin	5.232	4.288	330.7E6	449.2E6	52.206	51.754
6)	B beta-BHC	4.489	3.952	143.9E6	206.1E6	52.768	50.746m
7)	B delta-BHC	4.733	4.185	339.8E6	476.3E6	55.993m	51.817m
8)	B Heptachlo...	5.652	4.790	296.4E6	411.7E6	51.577	51.904
9)	A Endosulfan I	6.034	5.161	272.5E6	376.0E6	52.595	48.932
10)	B gamma-Chl...	5.906	5.042	298.6E6	431.1E6	52.816	49.364
11)	B alpha-Chl...	5.986	5.106	296.6E6	421.0E6	52.217	50.714
12)	B 4,4'-DDE	6.157	5.294	259.9E6	396.2E6	54.543	51.557
13)	MA Dieldrin	6.306	5.425	290.1E6	425.8E6	52.921	52.079
14)	MA Endrin	6.533	5.700	229.6E6	355.4E6	49.764	47.322
15)	B Endosulfa...	6.747	5.992	240.8E6	364.6E6	48.958	51.282
16)	A 4,4'-DDD	6.666	5.848	224.3E6	345.5E6	59.478	53.572
17)	MA 4,4'-DDT	6.981	6.100	199.3E6	318.0E6	47.139	45.954
18)	B Endrin al...	6.875	6.171	162.8E6	255.1E6	52.807	47.798
19)	B Endosulfa...	7.109	6.394	215.0E6	343.7E6	52.665	50.175
20)	A Methoxychlor	7.454	6.673	95109290	164.5E6	45.608	43.639
21)	B Endrin ke...	7.589	6.899	233.9E6	396.5E6	55.108	52.654
22)	Mirex	8.067	7.088	174.8E6	295.6E6	49.227	49.442

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:52
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/24/2025

Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097269.D Date Analyzed: 09/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.000	8.900	9.100	15.660	20.000	-21.7
Tetrachloro-m-xylene	3.527	3.480	3.580	14.770	20.000	-26.2
alpha-BHC	3.974	3.920	4.020	6.930	10.000	-30.7
beta-BHC	4.489	4.440	4.540	7.340	10.000	-26.6
gamma-BHC (Lindane)	4.302	4.250	4.350	7.210	10.000	-27.9
Endrin	6.532	6.460	6.600	37.780	50.000	-24.4
4,4'-DDT	6.979	6.910	7.050	73.290	100.000	-26.7
Methoxychlor	7.452	7.380	7.520	180.320	250.000	-27.9

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097269.D Date Analyzed: 09/18/2025

Lab Sample No.(PEM): PEM Time Analyzed: 16:39

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.880	8.080	15.470	20.000	-22.7
Tetrachloro-m-xylene	2.824	2.770	2.870	14.550	20.000	-27.3
alpha-BHC	3.329	3.280	3.380	6.890	10.000	-31.1
beta-BHC	3.957	3.910	4.010	7.490	10.000	-25.1
gamma-BHC (Lindane)	3.660	3.610	3.710	7.000	10.000	-30.0
Endrin	5.702	5.630	5.770	36.670	50.000	-26.7
4,4'-DDT	6.102	6.030	6.170	76.460	100.000	-23.5
Methoxychlor	6.674	6.600	6.740	166.760	250.000	-33.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097269.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 16:39
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/19/2025
 Supervised By :mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:42:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	66867312	95692858	14.770	14.553
28)	SA Decachlor...	9.000	7.984	51238202	88066168	15.659	15.473
Target Compounds							
2)	A alpha-BHC	3.974	3.329	48421315	68644926	6.930	6.894
3)	MA gamma-BHC...	4.302	3.660	47659315	65230882	7.207	7.003
6)	B beta-BHC	4.489	3.957	20012852	30421608	7.341	7.489
14)	MA Endrin	6.532	5.702	174.3E6	275.4E6	37.781	36.671
16)	A 4,4'-DDD	6.665	5.856	465298	1405401	0.123m	0.218m#
17)	MA 4,4'-DDT	6.979	6.102	309.8E6	529.1E6	73.292	76.458
18)	B Endrin al...	6.879	6.171	1009715	9847104	0.328m	1.845m#
20)	A Methoxychlor	7.452	6.674	376.0E6	628.6E6	180.319	166.763
21)	B Endrin ke...	7.586	6.896	1666712	2086500	0.393m	0.277m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 16:39
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/19/2025
Supervised By : mohammad ahmed 09/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:42:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097316.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:20

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.999	8.900	9.100	18.680	20.000	-6.6
Tetrachloro-m-xylene	3.527	3.480	3.580	17.520	20.000	-12.4
alpha-BHC	3.974	3.920	4.020	8.550	10.000	-14.5
beta-BHC	4.489	4.440	4.540	9.390	10.000	-6.1
gamma-BHC (Lindane)	4.301	4.250	4.350	8.730	10.000	-12.7
Endrin	6.530	6.460	6.600	42.130	50.000	-15.7
4,4'-DDT	6.978	6.910	7.050	83.940	100.000	-16.1
Methoxychlor	7.451	7.380	7.520	204.110	250.000	-18.4

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097316.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:20

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.982	7.880	8.080	16.770	20.000	-16.2
Tetrachloro-m-xylene	2.824	2.770	2.870	17.780	20.000	-11.1
alpha-BHC	3.328	3.280	3.380	8.450	10.000	-15.5
beta-BHC	3.956	3.910	4.010	9.160	10.000	-8.4
gamma-BHC (Lindane)	3.660	3.610	3.710	8.590	10.000	-14.1
Endrin	5.701	5.630	5.770	42.060	50.000	-15.9
4,4'-DDT	6.100	6.030	6.170	85.270	100.000	-14.7
Methoxychlor	6.673	6.600	6.740	176.620	250.000	-29.4

Data File: PEM
 PL097316.D Date Acquired 9/22/2025 11:20
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.53	194354667.4	202880976.3	8526308.84	4.20
Endrin aldehyde	6.87	2149412.225			
Endrin ketone	7.59	6376896.611			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.70	315882513.4	343526859.1	27644345.7	8.05
Endrin aldehyde #2	6.17	14279473.48			
Endrin ketone #2	6.90	13364872.21			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	6.98	354820908.4	364179964.7	9359056.34	2.57
4,4'-DDE	0.00	0			
4,4'-DDD	6.66	9359056.342			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.10	590137095.6	606116706	15979610.4	2.64
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.85	15979610.44			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097316.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 11:20
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:45:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	79301035	116.9E6	17.517	17.780
28)	SA Decachlor...	8.999	7.982	61113780	95442246	18.677	16.769
Target Compounds							
2)	A alpha-BHC	3.974	3.328	59728278	84100216	8.548	8.446
3)	MA gamma-BHC...	4.301	3.660	57729518	80052449	8.730	8.594
6)	B beta-BHC	4.489	3.956	25610468	37189948	9.394	9.155
14)	MA Endrin	6.530	5.701	194.4E6	315.9E6	42.131	42.059
16)	A 4,4'-DDD	6.664	5.849	9359056	15979610	2.482	2.478
17)	MA 4,4'-DDT	6.978	6.100	354.8E6	590.1E6	83.938	85.274
18)	B Endrin al...	6.874	6.171	2149412	14279473	0.697m	2.675 #
20)	A Methoxychlor	7.451	6.673	425.7E6	665.8E6	204.115	176.621
21)	B Endrin ke...	7.585	6.898	6376897	13364872	1.502	1.775

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097316.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 11:20
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/23/2025
Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:45:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097340.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:55

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.000	8.900	9.100	18.730	20.000	-6.4
Tetrachloro-m-xylene	3.528	3.480	3.580	18.350	20.000	-8.3
alpha-BHC	3.974	3.920	4.020	8.690	10.000	-13.1
beta-BHC	4.490	4.440	4.540	9.410	10.000	-5.9
gamma-BHC (Lindane)	4.302	4.250	4.350	8.850	10.000	-11.5
Endrin	6.531	6.460	6.600	42.820	50.000	-14.4
4,4'-DDT	6.979	6.910	7.050	81.390	100.000	-18.6
Methoxychlor	7.452	7.380	7.520	197.170	250.000	-21.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097340.D Date Analyzed: 09/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 20:55

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.984	7.880	8.080	17.340	20.000	-13.3
Tetrachloro-m-xylene	2.824	2.770	2.870	18.060	20.000	-9.7
alpha-BHC	3.329	3.280	3.380	8.600	10.000	-14.0
beta-BHC	3.957	3.910	4.010	9.230	10.000	-7.7
gamma-BHC (Lindane)	3.661	3.610	3.710	8.720	10.000	-12.8
Endrin	5.702	5.630	5.770	42.230	50.000	-15.5
4,4'-DDT	6.102	6.030	6.170	82.920	100.000	-17.1
Methoxychlor	6.674	6.600	6.740	175.530	250.000	-29.8

Data File: PEM
 PL097340.D Date Acquired 9/22/2025 20:55
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.53	197522792.5	210565760.3	13042967.8	Down 6.19
Endrin aldehyde	6.87	2701989.386			
Endrin ketone	7.59	10340978.44			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.70	317134944.9	353282053.4	36147108.4	10.23
Endrin aldehyde #2	6.17	13823351.75			
Endrin ketone #2	6.90	22323756.66			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	6.98	344059391.2	360645154.8	16585763.6	4.60
4,4'-DDE	0.00	0			
4,4'-DDD	6.66	16585763.59			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.10	573838234.1	601488127.4	27649893.3	4.60
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.85	27649893.29			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097340.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 20:55
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:49:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.824	83095318	118.7E6	18.355	18.056
28)	SA Decachlor...	9.000	7.984	61292884	98676908	18.731	17.337
Target Compounds							
2)	A alpha-BHC	3.974	3.329	60697471	85662634	8.687	8.603
3)	MA gamma-BHC...	4.302	3.661	58558042	81230119	8.855	8.721
6)	B beta-BHC	4.490	3.957	25643491	37511854	9.406	9.235
14)	MA Endrin	6.531	5.702	197.5E6	317.1E6	42.817	42.225
16)	A 4,4'-DDD	6.665	5.849	16585764	27649893	4.399	4.287
17)	MA 4,4'-DDT	6.979	6.102	344.1E6	573.8E6	81.392	82.919
18)	B Endrin al...	6.874	6.173	2701989	13823352	0.876m	2.590 #
20)	A Methoxychlor	7.452	6.674	411.2E6	661.6E6	197.174	175.527
21)	B Endrin ke...	7.586	6.899	10340978	22323757	2.436	2.965

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097340.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:55
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/23/2025
Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097358.D Date Analyzed: 09/23/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:29

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.000	8.900	9.100	19.890	20.000	-0.6
Tetrachloro-m-xylene	3.527	3.480	3.580	19.560	20.000	-2.2
alpha-BHC	3.974	3.920	4.020	9.120	10.000	-8.8
beta-BHC	4.489	4.440	4.540	9.560	10.000	-4.4
gamma-BHC (Lindane)	4.302	4.250	4.350	9.380	10.000	-6.2
Endrin	6.531	6.460	6.600	44.210	50.000	-11.6
4,4'-DDT	6.979	6.910	7.050	81.540	100.000	-18.5
Methoxychlor	7.452	7.380	7.520	198.490	250.000	-20.6

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097358.D Date Analyzed: 09/23/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:29

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.983	7.880	8.080	18.140	20.000	-9.3
Tetrachloro-m-xylene	2.824	2.770	2.870	18.590	20.000	-7.1
alpha-BHC	3.328	3.280	3.380	8.860	10.000	-11.4
beta-BHC	3.956	3.910	4.010	9.450	10.000	-5.5
gamma-BHC (Lindane)	3.660	3.610	3.710	8.990	10.000	-10.1
Endrin	5.702	5.630	5.770	41.890	50.000	-16.2
4,4'-DDT	6.101	6.030	6.170	80.180	100.000	-19.8
Methoxychlor	6.673	6.600	6.740	172.630	250.000	-30.9

Data File: PEM
 PL097358.D Date Acquired 9/23/2025 9:29
 Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.53	203955413.7	221793406.1	17837992.4	8.04
Endrin aldehyde	6.87	3317697.309			
Endrin ketone	7.59	14520295.06			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.70	314600408.7	355434130.5	40833721.7	11.49
Endrin aldehyde #2	6.17	12840138.62			
Endrin ketone #2	6.90	27993583.13			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	6.98	344694710	372233138.8	27538428.8	7.40
4,4'-DDE	6.15	1821810.6			
4,4'-DDD	6.66	25716618.17			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.10	554908257.3	595809407.7	40901150.4	6.86
4,4'-DDE #2	5.30	855418.766			
4,4'-DDD #2	5.85	40045731.62			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097358.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 09:29
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	88546646	122.3E6	19.559	18.594
28)	SA Decachlor...	9.000	7.983	65069354	103.2E6	19.885	18.140
Target Compounds							
2)	A alpha-BHC	3.974	3.328	63757976	88195831	9.125	8.857
3)	MA gamma-BHC...	4.302	3.660	62033294	83769301	9.380	8.993
6)	B beta-BHC	4.489	3.956	26052455	38395150	9.556	9.452
12)	B 4,4'-DDE	6.153	5.297	1821811	855419	0.382m	0.111m#
14)	MA Endrin	6.531	5.702	204.0E6	314.6E6	44.212	41.888
16)	A 4,4'-DDD	6.664	5.849	25716618	40045732	6.820	6.210
17)	MA 4,4'-DDT	6.979	6.101	344.7E6	554.9E6	81.542	80.183
18)	B Endrin al...	6.874	6.173	3317697	12840139	1.076m	2.406 #
20)	A Methoxychlor	7.452	6.673	413.9E6	650.7E6	198.487	172.634
21)	B Endrin ke...	7.586	6.899	14520295	27993583	3.421	3.717

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097358.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:29
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/24/2025
Supervised By : mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Analytical Sequence

Client: Always Available Construction LLC.

SDG No.: Q3157

Project: Yard Soil Cleanup

Instrument ID: ECD_L

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 09/18/2025 09/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	09/18/2025	16:25	PL097268.D	9.00	3.53
PEM	PEM	09/18/2025	16:39	PL097269.D	9.00	3.53
RESCHK	RESCHK	09/18/2025	16:52	PL097270.D	9.00	3.53
PSTDICCC100	PSTDICCC100	09/18/2025	17:06	PL097271.D	9.00	3.53
PSTDICCC075	PSTDICCC075	09/18/2025	17:19	PL097272.D	9.00	3.53
PSTDICCC050	PSTDICCC050	09/18/2025	17:33	PL097273.D	9.00	3.53
PSTDICCC025	PSTDICCC025	09/18/2025	17:47	PL097274.D	9.00	3.53
PSTDICCC005	PSTDICCC005	09/18/2025	18:00	PL097275.D	9.00	3.53
PCHLORICC500	PCHLORICC500	09/18/2025	18:41	PL097278.D	9.00	3.53
PTOXICC500	PTOXICC500	09/18/2025	19:49	PL097283.D	9.00	3.53
PEM	PEM	09/22/2025	11:20	PL097316.D	9.00	3.53
IBLK	IBLK	09/22/2025	17:58	PL097327.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/22/2025	18:11	PL097328.D	9.00	3.53
PB169782BL	PB169782BL	09/22/2025	18:52	PL097331.D	9.00	3.53
291376MS	Q3092-02MS	09/22/2025	19:47	PL097335.D	9.00	3.53
291376MSD	Q3092-02MSD	09/22/2025	20:00	PL097336.D	9.00	3.53
PB169754TB	PB169754TB	09/22/2025	20:28	PL097338.D	9.00	3.53
IBLK	IBLK	09/22/2025	20:41	PL097339.D	9.00	3.53
PEM	PEM	09/22/2025	20:55	PL097340.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/22/2025	21:08	PL097341.D	9.00	3.53
20-DEAD-1	Q3157-02	09/22/2025	22:03	PL097343.D	9.00	3.53
IBLK	IBLK	09/23/2025	00:32	PL097354.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/23/2025	00:46	PL097355.D	9.00	3.53
IBLK	IBLK	09/23/2025	09:16	PL097357.D	9.00	3.53
PEM	PEM	09/23/2025	09:29	PL097358.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/23/2025	10:03	PL097359.D	9.00	3.53
PB169782BS	PB169782BS	09/23/2025	11:47	PL097362.D	9.00	3.53
IBLK	IBLK	09/23/2025	12:09	PL097363.D	9.00	3.53
PSTDCCC050	PSTDCCC050	09/23/2025	12:52	PL097364.D	9.00	3.53

Analytical Sequence

Client: Always Available Construction LLC.

SDG No.: Q3157

Project: Yard Soil Cleanup

Instrument ID: ECD_L

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 09/18/2025 09/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	09/18/2025	16:25	PL097268.D	7.98	2.82
PEM	PEM	09/18/2025	16:39	PL097269.D	7.98	2.82
RESCHK	RESCHK	09/18/2025	16:52	PL097270.D	7.98	2.82
PSTDICCC100	PSTDICCC100	09/18/2025	17:06	PL097271.D	7.98	2.82
PSTDICCC075	PSTDICCC075	09/18/2025	17:19	PL097272.D	7.99	2.82
PSTDICCC050	PSTDICCC050	09/18/2025	17:33	PL097273.D	7.98	2.82
PSTDICCC025	PSTDICCC025	09/18/2025	17:47	PL097274.D	7.98	2.82
PSTDICCC005	PSTDICCC005	09/18/2025	18:00	PL097275.D	7.98	2.82
PCHLORICC500	PCHLORICC500	09/18/2025	18:41	PL097278.D	7.98	2.82
PTOXICC500	PTOXICC500	09/18/2025	19:49	PL097283.D	7.99	2.82
PEM	PEM	09/22/2025	11:20	PL097316.D	7.98	2.82
IBLK	IBLK	09/22/2025	17:58	PL097327.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/22/2025	18:11	PL097328.D	7.98	2.82
PB169782BL	PB169782BL	09/22/2025	18:52	PL097331.D	7.98	2.83
291376MS	Q3092-02MS	09/22/2025	19:47	PL097335.D	7.98	2.83
291376MSD	Q3092-02MSD	09/22/2025	20:00	PL097336.D	7.98	2.82
PB169754TB	PB169754TB	09/22/2025	20:28	PL097338.D	7.98	2.82
IBLK	IBLK	09/22/2025	20:41	PL097339.D	7.98	2.82
PEM	PEM	09/22/2025	20:55	PL097340.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/22/2025	21:08	PL097341.D	7.98	2.82
20-DEAD-1	Q3157-02	09/22/2025	22:03	PL097343.D	7.98	2.83
IBLK	IBLK	09/23/2025	00:32	PL097354.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/23/2025	00:46	PL097355.D	7.98	2.82
IBLK	IBLK	09/23/2025	09:16	PL097357.D	7.98	2.82
PEM	PEM	09/23/2025	09:29	PL097358.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/23/2025	10:03	PL097359.D	7.98	2.82
PB169782BS	PB169782BS	09/23/2025	11:47	PL097362.D	7.98	2.82
IBLK	IBLK	09/23/2025	12:09	PL097363.D	7.98	2.82
PSTDCCC050	PSTDCCC050	09/23/2025	12:52	PL097364.D	7.98	2.82

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

291376MS

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Lab Sample ID: Q3092-02MS

Date(s) Analyzed: 09/22/2025 09/22/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.45	7.40	7.50	4.20	4.9
	2	6.67	6.62	6.72	4.00	
gamma-BHC (Lindane)	1	4.30	4.25	4.35	4.60	2.2
	2	3.66	3.61	3.71	4.70	
Heptachlor	1	4.89	4.84	4.94	4.50	0
	2	4.01	3.96	4.06	4.50	
Heptachlor epoxide	1	5.65	5.60	5.70	4.70	0
	2	4.79	4.74	4.84	4.70	
Endrin	1	6.53	6.48	6.58	4.40	2.2
	2	5.70	5.65	5.75	4.50	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

291376MSD

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Lab Sample ID: Q3092-02MSD

Date(s) Analyzed: 09/22/2025 09/22/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.45	7.40	7.50	4.20	4.9
	2	6.67	6.62	6.72	4.00	
gamma-BHC (Lindane)	1	4.30	4.25	4.35	4.60	0
	2	3.66	3.61	3.71	4.60	
Heptachlor	1	4.89	4.84	4.94	4.50	2.2
	2	4.01	3.96	4.06	4.40	
Heptachlor epoxide	1	5.65	5.60	5.70	4.70	2.2
	2	4.79	4.74	4.84	4.60	
Endrin	1	6.53	6.48	6.58	4.50	0
	2	5.70	5.65	5.75	4.50	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB169782BS

Lab Name: Alliance

Contract: ALWA01

Lab Code: ACE

SDG NO.: Q3157

Lab Sample ID: PB169782BS

Date(s) Analyzed: 09/23/2025 09/23/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin	1	6.53	6.48	6.58	0.45	0.3
	2	5.70	5.65	5.75	0.45	
Methoxychlor	1	7.45	7.40	7.50	0.42	3
	2	6.67	6.62	6.72	0.41	
gamma-BHC (Lindane)	1	4.30	4.25	4.35	0.51	1.8
	2	3.66	3.61	3.71	0.52	
Heptachlor	1	4.89	4.84	4.94	0.48	0.4
	2	4.01	3.96	4.06	0.49	
Heptachlor epoxide	1	5.65	5.60	5.70	0.52	1.4
	2	4.79	4.74	4.84	0.51	



QC SAMPLE DATA

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	
Project:	Yard Soil Cleanup	Date Received:	
Client Sample ID:	PB169782BL	SDG No.:	Q3157
Lab Sample ID:	PB169782BL	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097331.D	1	09/22/25 10:05	09/22/25 18:52	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.5		57 - 171	88%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.9		61 - 148	85%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097331.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 18:52
Operator : AR\AJ
Sample : PB169782BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169782BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.825	75396086	111.2E6	16.654	16.905
28)	SA Decachlor...	9.001	7.983	57319860	88056925	17.517	15.471

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097331.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 18:52
Operator : AR\AJ
Sample : PB169782BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB169782BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 75396086
Conc: 16.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB169782BL




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: 0.001 min
Response: 111158080
Conc: 16.90 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.001 min
Delta R.T.: 0.000 min
Response: 57319860
Conc: 17.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 88056925
Conc: 15.47 ng/ml

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	09/18/25
Project:	Yard Soil Cleanup	Date Received:	09/18/25
Client Sample ID:	PIBLK-PL097268.D	SDG No.:	Q3157
Lab Sample ID:	I.BLK-PL097268.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097268.D	1		09/18/25	PL092025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.3		57 - 171	87%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.0		61 - 148	80%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 16:25
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:42:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	70468568	105.0E6	15.566	15.973
28)	SA Decachlor...	9.001	7.984	56635933	97364753	17.308	17.107

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 16:25
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:42:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 70468568
Conc: 15.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 105031523
Conc: 15.97 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.001 min
Delta R.T.: 0.000 min
Response: 56635933
Conc: 17.31 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.984 min
Delta R.T.: 0.000 min
Response: 97364753
Conc: 17.11 ng/ml

Report of Analysis

Client:	Always Available Construction LLC.		Date Collected:	09/22/25	
Project:	Yard Soil Cleanup		Date Received:	09/22/25	
Client Sample ID:	PIBLK-PL097327.D		SDG No.:	Q3157	
Lab Sample ID:	I.BLK-PL097327.D		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097327.D	1		09/22/25	pl092225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.2		57 - 171	101%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.6		61 - 148	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097327.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 17:58
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:47:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	84311640	122.3E6	18.624	18.599
28)	SA Decachlor...	9.000	7.983	66073819	98546486	20.192	17.314

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097327.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 17:58
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:47:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 84311640
Conc: 18.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 122301082
Conc: 18.60 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.000 min
Delta R.T.: 0.000 min
Response: 66073819
Conc: 20.19 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 98546486
Conc: 17.31 ng/ml

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	09/22/25
Project:	Yard Soil Cleanup	Date Received:	09/22/25
Client Sample ID:	PIBLK-PL097339.D	SDG No.:	Q3157
Lab Sample ID:	I.BLK-PL097339.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097339.D	1		09/22/25	pl092225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.0		57 - 171	105%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.9		61 - 148	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097339.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:41
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	82711967	124.4E6	18.270	18.913
28)	SA Decachlor...	8.999	7.983	68660624	109.3E6	20.983	19.202

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097339.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:41
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:49:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 82711967
Conc: 18.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 124360641
Conc: 18.91 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: -0.002 min
Response: 68660624
Conc: 20.98 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 109291468
Conc: 19.20 ng/ml

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	09/23/25
Project:	Yard Soil Cleanup	Date Received:	09/23/25
Client Sample ID:	PIBLK-PL097354.D	SDG No.:	Q3157
Lab Sample ID:	I.BLK-PL097354.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097354.D	1		09/23/25	pl092225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.8		57 - 171	104%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.6		61 - 148	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097354.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 00:32
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:51:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.824	81964806	122.4E6	18.105	18.621
28)	SA Decachlor...	9.000	7.983	67978796	99090184	20.775	17.410

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097354.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 00:32
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:51:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm




Response_ Signal: PL097354.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 81964806
Conc: 18.11 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




Time

Response_ Signal: PL097354.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.001 min
Response: 122444037
Conc: 18.62 ng/ml




Time

Response_ Signal: PL097354.D\ECD1A.ch

#28 Decachlorobiphenyl

R.T.: 9.000 min
Delta R.T.: -0.001 min
Response: 67978796
Conc: 20.77 ng/ml




Time

Response_ Signal: PL097354.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: -0.001 min
Response: 99090184
Conc: 17.41 ng/ml



Time

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	09/23/25
Project:	Yard Soil Cleanup	Date Received:	09/23/25
Client Sample ID:	PIBLK-PL097357.D	SDG No.:	Q3157
Lab Sample ID:	I.BLK-PL097357.D	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097357.D	1		09/23/25	pl092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.2		57 - 171	111%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.7		61 - 148	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:16
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:25:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.823	89086643	127.1E6	19.678	19.331
28)	SA Decachlor...	8.999	7.982	72637734	114.4E6	22.198	20.091

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:16
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:25:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm




Response_ Signal: PL097357.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 89086643
Conc: 19.68 ng/ml


Instrument :
ECD_L
ClientSampleId :
I.BLK



Time Response_ Signal: PL097357.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.823 min
Delta R.T.: 0.000 min
Response: 127111040
Conc: 19.33 ng/ml



Time Response_ Signal: PL097357.D\ECD1A.ch

#28 Decachlorobiphenyl


R.T.: 8.999 min
Delta R.T.: -0.001 min
Response: 72637734
Conc: 22.20 ng/ml



Time Response_ Signal: PL097357.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 7.982 min
Delta R.T.: -0.002 min
Response: 114350562
Conc: 20.09 ng/ml



Report of Analysis

Client:	Always Available Construction LLC.			Date Collected:	09/23/25	
Project:	Yard Soil Cleanup			Date Received:	09/23/25	
Client Sample ID:	PIBLK-PL097363.D			SDG No.:	Q3157	
Lab Sample ID:	I.BLK-PL097363.D			Matrix:	TCLP	
Analytical Method:	8081B			% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	10000	uL
Soil Aliquot Vol:			uL	Test:	TCLP Pesticide	
Extraction Type:				Injection Volume :		
GPC Factor :	1.0		PH :			
Prep Method :	3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097363.D	1		09/23/25	pl092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.0037	U	0.0037	0.050	ug/L
76-44-8	Heptachlor	0.0027	U	0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.0096	U	0.0096	0.050	ug/L
72-20-8	Endrin	0.0032	U	0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.011	U	0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.4		57 - 171	107%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.6		61 - 148	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:09
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.822	85041873	129.1E6	18.785	19.632
28)	SA Decachlor...	9.001	7.983	70128233	116.0E6	21.431	20.373

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:09
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.527 min
Delta R.T.: 0.000 min
Response: 85041873
Conc: 18.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.822 min
Delta R.T.: -0.002 min
Response: 129088691
Conc: 19.63 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.001 min
Delta R.T.: 0.000 min
Response: 70128233
Conc: 21.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 115954509
Conc: 20.37 ng/ml

Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	
Project:	Yard Soil Cleanup	Date Received:	
Client Sample ID:	PB169782BS	SDG No.:	Q3157
Lab Sample ID:	PB169782BS	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	TCLP Pesticide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097362.D	1	09/22/25 10:05	09/23/25 11:47	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.52		0.0037	0.050	ug/L
76-44-8	Heptachlor	0.49		0.0027	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.52		0.0096	0.050	ug/L
72-20-8	Endrin	0.45		0.0032	0.050	ug/L
72-43-5	Methoxychlor	0.42		0.011	0.050	ug/L
8001-35-2	Toxaphene	0.17	U	0.17	1.00	ug/L
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.1		57 - 171	91%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.4		61 - 148	87%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
 Data File : PL097362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 11:47
 Operator : AR\AJ
 Sample : PB169782BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB169782BS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:26:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.823	77023023	114.6E6	17.014	17.436
28)	SA Decachlor...	8.999	7.982	59276355	98829039	18.115	17.364
Target Compounds							
2)	A alpha-BHC	3.973	3.328	361.5E6	522.2E6	51.737	52.440
3)	MA gamma-BHC...	4.301	3.660	336.8E6	483.0E6	50.928	51.858
4)	MA Heptachlor	4.892	4.007	322.6E6	451.9E6	48.367	48.561
5)	MB Aldrin	5.231	4.289	323.8E6	451.1E6	51.117	51.972
6)	B beta-BHC	4.489	3.956	137.2E6	198.4E6	50.337	48.846
7)	B delta-BHC	4.733	4.188	318.3E6	467.2E6	52.449m	50.823
8)	B Heptachlo...	5.650	4.791	298.1E6	405.6E6	51.871	51.137
9)	A Endosulfan I	6.032	5.162	265.4E6	380.9E6	51.229	49.569
10)	B gamma-Chl...	5.904	5.043	290.8E6	444.0E6	51.441	50.837
11)	B alpha-Chl...	5.984	5.107	288.7E6	426.0E6	50.823	51.327
12)	B 4,4'-DDE	6.154	5.296	252.8E6	396.5E6	53.047	51.593
13)	MA Dieldrin	6.304	5.426	281.0E6	421.9E6	51.252	51.597
14)	MA Endrin	6.530	5.701	209.5E6	340.2E6	45.418	45.300
15)	B Endosulfa...	6.744	5.993	238.4E6	356.2E6	48.474	50.104
16)	A 4,4'-DDD	6.664	5.848	216.2E6	354.9E6	57.338	55.031
17)	MA 4,4'-DDT	6.978	6.100	185.7E6	293.1E6	43.927	42.353
18)	B Endrin al...	6.872	6.171	173.5E6	270.2E6	56.275	50.633
19)	B Endosulfa...	7.106	6.394	208.8E6	337.9E6	51.150	49.324
20)	A Methoxychlor	7.451	6.672	87648949	153.8E6	42.030	40.795
21)	B Endrin ke...	7.585	6.898	232.3E6	408.2E6	54.730	54.206
22)	Mirex	8.063	7.087	166.3E6	284.0E6	46.829	47.498

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092325\
Data File : PL097362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:47
Operator : AR\AJ
Sample : PB169782BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PB169782BS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/24/2025

Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:26:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	09/12/25
Project:	Yard Soil Cleanup	Date Received:	09/12/25
Client Sample ID:	291376MS	SDG No.:	Q3157
Lab Sample ID:	Q3092-02MS	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097335.D	1	09/22/25 10:05	09/22/25 19:47	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	4.70		0.037	0.50	ug/L
76-44-8	Heptachlor	4.50		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	4.70		0.096	0.50	ug/L
72-20-8	Endrin	4.50		0.032	0.50	ug/L
72-43-5	Methoxychlor	4.20		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.9		57 - 171	84%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.4		61 - 148	87%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097335.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 19:47
 Operator : AR\AJ
 Sample : Q3092-02MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 291376MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/23/2025
 Supervised By : mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:48:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.825	74231591	114.3E6	16.397m	17.383
28)	SA Decachlor...	8.999	7.984	55196689	87385123	16.868	15.353
Target Compounds							
2)	A alpha-BHC	3.974	3.330	324.7E6	466.2E6	46.465	46.821
3)	MA gamma-BHC...	4.302	3.661	303.7E6	434.0E6	45.923	46.590
4)	MA Heptachlor	4.893	4.009	301.1E6	416.4E6	45.135	44.749
5)	MB Aldrin	5.232	4.291	286.5E6	399.5E6	45.241	46.027
6)	B beta-BHC	4.489	3.957	124.8E6	176.2E6	45.770	43.372
7)	B delta-BHC	4.734	4.190	290.5E6	421.8E6	47.869m	45.886
8)	B Heptachlo...	5.651	4.793	271.4E6	370.7E6	47.220	46.740
9)	A Endosulfan I	6.033	5.163	238.2E6	351.2E6	45.969	45.701
10)	B gamma-Chl...	5.905	5.045	261.3E6	404.4E6	46.214	46.305
11)	B alpha-Chl...	5.985	5.108	257.9E6	389.1E6	45.399	46.882
12)	B 4,4'-DDE	6.155	5.297	222.3E6	360.2E6	46.650	46.868
13)	MA Dieldrin	6.305	5.428	251.4E6	384.0E6	45.859	46.957
14)	MA Endrin	6.531	5.702	203.3E6	337.7E6	44.076	44.970
15)	B Endosulfa...	6.745	5.994	210.5E6	324.0E6	42.807	45.574
16)	A 4,4'-DDD	6.664	5.849	182.2E6	305.5E6	48.310	47.378
17)	MA 4,4'-DDT	6.978	6.102	188.1E6	292.1E6	44.503	42.203
18)	B Endrin al...	6.873	6.172	153.8E6	243.7E6	49.881	45.660
19)	B Endosulfa...	7.107	6.396	188.0E6	302.2E6	46.057	44.125
20)	A Methoxychlor	7.451	6.674	88194366	152.3E6	42.292	40.398
21)	B Endrin ke...	7.586	6.900	200.8E6	335.5E6	47.304	44.558
22)	Mirex	8.065	7.089	149.9E6	249.0E6	42.217	41.646

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097335.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 19:47
Operator : AR\AJ
Sample : Q3092-02MS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

291376MS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Report of Analysis

Client:	Always Available Construction LLC.	Date Collected:	09/12/25
Project:	Yard Soil Cleanup	Date Received:	09/12/25
Client Sample ID:	291376MSD	SDG No.:	Q3157
Lab Sample ID:	Q3092-02MSD	Matrix:	TCLP
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	TCLP Pesticide
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL097336.D	1	09/22/25 10:05	09/22/25 20:00	PB169782

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	4.60		0.037	0.50	ug/L
76-44-8	Heptachlor	4.50		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	4.70		0.096	0.50	ug/L
72-20-8	Endrin	4.50		0.032	0.50	ug/L
72-43-5	Methoxychlor	4.20		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	16.9		57 - 171	85%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.1		61 - 148	86%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
 Data File : PL097336.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2025 20:00
 Operator : AR\AJ
 Sample : Q3092-02MSD
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 291376MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/23/2025
 Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 01:48:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.824	74636866	112.8E6	16.486m	17.149
28)	SA Decachlor...	8.999	7.984	55476108	85996726	16.954	15.109
Target Compounds							
2)	A alpha-BHC	3.975	3.329	324.2E6	461.3E6	46.399	46.324
3)	MA gamma-BHC...	4.302	3.661	303.3E6	430.2E6	45.865	46.185
4)	MA Heptachlor	4.893	4.009	301.0E6	412.6E6	45.125	44.340
5)	MB Aldrin	5.232	4.291	286.3E6	396.1E6	45.196	45.636
6)	B beta-BHC	4.489	3.957	124.8E6	174.1E6	45.780	42.855
7)	B delta-BHC	4.733	4.190	289.1E6	418.4E6	47.643m	45.514
8)	B Heptachlo...	5.651	4.792	270.4E6	368.5E6	47.053	46.460
9)	A Endosulfan I	6.033	5.163	238.3E6	348.9E6	45.994	45.402
10)	B gamma-Chl...	5.905	5.044	262.2E6	401.5E6	46.375	45.970
11)	B alpha-Chl...	5.985	5.108	257.7E6	386.7E6	45.377	46.588
12)	B 4,4'-DDE	6.156	5.297	222.7E6	357.0E6	46.746	46.451
13)	MA Dieldrin	6.305	5.428	251.0E6	380.7E6	45.781	46.563
14)	MA Endrin	6.531	5.702	207.4E6	335.7E6	44.958	44.692
15)	B Endosulfa...	6.745	5.994	212.6E6	323.9E6	43.219	45.563
16)	A 4,4'-DDD	6.664	5.849	181.7E6	305.3E6	48.197	47.338
17)	MA 4,4'-DDT	6.979	6.102	187.6E6	291.5E6	44.377	42.118
18)	B Endrin al...	6.873	6.172	153.4E6	243.6E6	49.749	45.646
19)	B Endosulfa...	7.107	6.395	187.2E6	302.9E6	45.861	44.222
20)	A Methoxychlor	7.451	6.674	87326449	151.9E6	41.876	40.301
21)	B Endrin ke...	7.586	6.900	197.6E6	335.1E6	46.560	44.502
22)	Mirex	8.064	7.089	149.6E6	247.8E6	42.150	41.443

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092225\
Data File : PL097336.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2025 20:00
Operator : AR\AJ
Sample : Q3092-02MSD
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

291376MSD

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 09/23/2025

Supervised By :mohammad ahmed 09/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 01:48:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm



Manual Integration Report

Sequence:

PL092025

Instrument

ECD_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL097269.D	4,4"-DDD	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	4,4"-DDD #2	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin aldehyde	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin aldehyde #2	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin ketone	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PEM	PL097269.D	Endrin ketone #2	Abdul	9/19/2025 11:23:24 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	alpha-Chlordane #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	delta-BHC	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Endosulfan II	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Endrin aldehyde #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Endrin ketone #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	gamma-Chlordane #2	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICC005	PL097275.D	Heptachlor epoxide	Abdul	9/19/2025 11:23:27 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software

Manual Integration Report

Sequence:	PL092025	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PSTDICV050	PL097286.D	4,4"-DDD	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	4,4"-DDE	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	4,4"-DDE #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	alpha-Chlordane	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	delta-BHC	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Dieldrin #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endosulfan I	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endosulfan II	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endrin	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Endrin ketone #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	gamma-Chlordane	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Heptachlor epoxide	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PSTDICV050	PL097286.D	Heptachlor epoxide #2	Abdul	9/22/2025 9:04:29 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software

Manual Integration Report

Sequence:	PL092025	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PCHLORICV500	PL097287.D	Chlordane-2	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-3	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-4	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-4 #2	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PCHLORICV500	PL097287.D	Chlordane-5	Abdul	9/22/2025 9:04:32 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software
PTOXICV500	PL097288.D	Toxaphene-1 #2	Abdul	9/22/2025 9:04:35 AM	mohammad	9/23/2025 1:17:52	Peak Integrated by Software

Manual Integration Report

Sequence:

pl092225

Instrument

ECD_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL097316.D	Endrin aldehyde	Abdul	9/23/2025 8:41:59 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097317.D	delta-BHC	Abdul	9/23/2025 8:42:03 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097328.D	delta-BHC	Abdul	9/23/2025 8:42:21 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MS	PL097335.D	delta-BHC	Abdul	9/23/2025 8:43:22 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MS	PL097335.D	Tetrachloro-m-xylene	Abdul	9/23/2025 8:43:22 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MSD	PL097336.D	delta-BHC	Abdul	9/23/2025 8:43:24 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
Q3092-02MSD	PL097336.D	Tetrachloro-m-xylene	Abdul	9/23/2025 8:43:24 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PEM	PL097340.D	Endrin aldehyde	Abdul	9/23/2025 8:43:27 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097341.D	4,4"-DDE #2	Abdul	9/23/2025 8:43:30 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097341.D	delta-BHC	Abdul	9/23/2025 8:43:30 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software
PSTDCCC050	PL097355.D	delta-BHC	Abdul	9/23/2025 8:44:17 AM	mohammad	9/24/2025 2:25:05	Peak Integrated by Software

Manual Integration Report

Sequence:

pl092325

Instrument

ECD_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL097358.D	4,4"-DDE	Abdul	9/24/2025 9:40:47 AM	 	 	Peak Integrated by Software
PEM	PL097358.D	4,4"-DDE #2	Abdul	9/24/2025 9:40:47 AM	 	 	Peak Integrated by Software
PEM	PL097358.D	Endrin aldehyde	Abdul	9/24/2025 9:40:47 AM	 	 	Peak Integrated by Software
PSTDCCC050	PL097359.D	delta-BHC	Abdul	9/24/2025 9:39:53 AM	 	 	Peak Integrated by Software
PB169782BS	PL097362.D	delta-BHC	Abdul	9/24/2025 9:40:03 AM	 	 	Peak Integrated by Software
PSTDCCC050	PL097364.D	beta-BHC #2	Abdul	9/24/2025 9:40:07 AM	 	 	Peak Integrated by Software
PSTDCCC050	PL097364.D	delta-BHC	Abdul	9/24/2025 9:40:07 AM	 	 	Peak Integrated by Software
PSTDCCC050	PL097364.D	delta-BHC #2	Abdul	9/24/2025 9:40:07 AM	 	 	Peak Integrated by Software

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097267.D	18 Sep 2025 16:11	AR\AJ	Ok
2	I.BLK	PL097268.D	18 Sep 2025 16:25	AR\AJ	Ok
3	PEM	PL097269.D	18 Sep 2025 16:39	AR\AJ	Ok,M
4	RESCHK	PL097270.D	18 Sep 2025 16:52	AR\AJ	Ok
5	PSTDICC100	PL097271.D	18 Sep 2025 17:06	AR\AJ	Ok
6	PSTDICC075	PL097272.D	18 Sep 2025 17:19	AR\AJ	Ok
7	PSTDICC050	PL097273.D	18 Sep 2025 17:33	AR\AJ	Ok
8	PSTDICC025	PL097274.D	18 Sep 2025 17:47	AR\AJ	Ok
9	PSTDICC005	PL097275.D	18 Sep 2025 18:00	AR\AJ	Ok,M
10	PCHLORICC1000	PL097276.D	18 Sep 2025 18:14	AR\AJ	Ok
11	PCHLORICC750	PL097277.D	18 Sep 2025 18:28	AR\AJ	Ok
12	PCHLORICC500	PL097278.D	18 Sep 2025 18:41	AR\AJ	Ok
13	PCHLORICC250	PL097279.D	18 Sep 2025 18:55	AR\AJ	Ok
14	PCHLORICC050	PL097280.D	18 Sep 2025 19:08	AR\AJ	Ok
15	PTOXICC1000	PL097281.D	18 Sep 2025 19:22	AR\AJ	Ok
16	PTOXICC750	PL097282.D	18 Sep 2025 19:35	AR\AJ	Ok,M
17	PTOXICC500	PL097283.D	18 Sep 2025 19:49	AR\AJ	Ok
18	PTOXICC250	PL097284.D	18 Sep 2025 20:03	AR\AJ	Ok
19	PTOXICC100	PL097285.D	18 Sep 2025 20:16	AR\AJ	Ok,M
20	PSTDICV050	PL097286.D	19 Sep 2025 08:39	AR\AJ	Ok,M
21	PCHLORICV500	PL097287.D	19 Sep 2025 08:52	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL097288.D	19 Sep 2025 09:25	AR\AJ	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097314.D	22 Sep 2025 10:12	AR\AJ	Ok
2	I.BLK	PL097315.D	22 Sep 2025 11:06	AR\AJ	Ok
3	PEM	PL097316.D	22 Sep 2025 11:20	AR\AJ	Ok,M
4	PSTDCCC050	PL097317.D	22 Sep 2025 11:33	AR\AJ	Ok,M
5	PB169749BS	PL097318.D	22 Sep 2025 12:31	AR\AJ	Ok,M
6	PB169772BL	PL097319.D	22 Sep 2025 13:30	AR\AJ	Ok
7	PB169772BS	PL097320.D	22 Sep 2025 13:44	AR\AJ	Ok,M
8	Q3152-01	PL097321.D	22 Sep 2025 14:05	AR\AJ	Ok,M
9	Q3135-01	PL097322.D	22 Sep 2025 14:19	AR\AJ	Ok,M
10	Q3151-02	PL097323.D	22 Sep 2025 14:32	AR\AJ	Ok
11	PB169766BL	PL097324.D	22 Sep 2025 14:46	AR\AJ	Ok
12	Q3153-04	PL097325.D	22 Sep 2025 17:30	AR\AJ	Ok
13	Q3153-08	PL097326.D	22 Sep 2025 17:44	AR\AJ	Ok
14	I.BLK	PL097327.D	22 Sep 2025 17:58	AR\AJ	Ok
15	PSTDCCC050	PL097328.D	22 Sep 2025 18:11	AR\AJ	Ok,M
16	PB169766BS	PL097329.D	22 Sep 2025 18:25	AR\AJ	Not Ok
17	PB169766BSD	PL097330.D	22 Sep 2025 18:39	AR\AJ	Not Ok
18	PB169782BL	PL097331.D	22 Sep 2025 18:52	AR\AJ	Ok
19	PB169782BS	PL097332.D	22 Sep 2025 19:06	AR\AJ	Not Ok
20	Q3065-04	PL097333.D	22 Sep 2025 19:20	AR\AJ	Ok
21	Q3092-02	PL097334.D	22 Sep 2025 19:33	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3092-02MS	PL097335.D	22 Sep 2025 19:47	AR\AJ	Ok,M
23	Q3092-02MSD	PL097336.D	22 Sep 2025 20:00	AR\AJ	Ok,M
24	PB169677TB	PL097337.D	22 Sep 2025 20:14	AR\AJ	Ok
25	PB169754TB	PL097338.D	22 Sep 2025 20:28	AR\AJ	Ok
26	I.BLK	PL097339.D	22 Sep 2025 20:41	AR\AJ	Ok
27	PEM	PL097340.D	22 Sep 2025 20:55	AR\AJ	Ok,M
28	PSTDCCC050	PL097341.D	22 Sep 2025 21:08	AR\AJ	Ok,M
29	Q3092-04	PL097342.D	22 Sep 2025 21:49	AR\AJ	Ok,M
30	Q3157-02	PL097343.D	22 Sep 2025 22:03	AR\AJ	Ok
31	Q3145-01	PL097344.D	22 Sep 2025 22:16	AR\AJ	Ok,M
32	Q3145-03	PL097345.D	22 Sep 2025 22:30	AR\AJ	Ok,M
33	Q3153-01	PL097346.D	22 Sep 2025 22:44	AR\AJ	Ok,M
34	Q3153-01MS	PL097347.D	22 Sep 2025 22:57	AR\AJ	Ok,M
35	Q3153-01MSD	PL097348.D	22 Sep 2025 23:11	AR\AJ	Ok,M
36	Q3153-05	PL097349.D	22 Sep 2025 23:24	AR\AJ	Ok,M
37	Q3155-01	PL097350.D	22 Sep 2025 23:38	AR\AJ	Ok,M
38	Q3155-02	PL097351.D	22 Sep 2025 23:52	AR\AJ	Ok,M
39	Q3155-03	PL097352.D	23 Sep 2025 00:05	AR\AJ	Ok,M
40	Q3157-01	PL097353.D	23 Sep 2025 00:19	AR\AJ	Ok,M
41	I.BLK	PL097354.D	23 Sep 2025 00:32	AR\AJ	Ok
42	PSTDCCC050	PL097355.D	23 Sep 2025 00:46	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092325

Review By	Abdul	Review On	9/24/2025 9:40:38 AM
Supervise By		Supervise On	
SubDirectory	PL092325	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097356.D	23 Sep 2025 09:02	AR\AJ	Ok
2	I.BLK	PL097357.D	23 Sep 2025 09:16	AR\AJ	Ok
3	PEM	PL097358.D	23 Sep 2025 09:29	AR\AJ	Ok,NS
4	PSTDCCC050	PL097359.D	23 Sep 2025 10:03	AR\AJ	Ok,NS
5	PB169766BS	PL097360.D	23 Sep 2025 11:14	AR\AJ	Ok,NS
6	PB169766BSD	PL097361.D	23 Sep 2025 11:34	AR\AJ	Ok,NS
7	PB169782BS	PL097362.D	23 Sep 2025 11:47	AR\AJ	Ok,NS
8	I.BLK	PL097363.D	23 Sep 2025 12:09	AR\AJ	Ok
9	PSTDCCC050	PL097364.D	23 Sep 2025 12:52	AR\AJ	Ok,NS

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method p092025 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL097267.D	18 Sep 2025 16:11		AR\AJ	Ok
2	I.BLK	I.BLK	PL097268.D	18 Sep 2025 16:25		AR\AJ	Ok
3	PEM	PEM	PL097269.D	18 Sep 2025 16:39		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL097270.D	18 Sep 2025 16:52		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL097271.D	18 Sep 2025 17:06		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL097272.D	18 Sep 2025 17:19		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL097273.D	18 Sep 2025 17:33		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL097274.D	18 Sep 2025 17:47		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL097275.D	18 Sep 2025 18:00		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL097276.D	18 Sep 2025 18:14		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL097277.D	18 Sep 2025 18:28		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL097278.D	18 Sep 2025 18:41		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL097279.D	18 Sep 2025 18:55		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL097280.D	18 Sep 2025 19:08		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL097281.D	18 Sep 2025 19:22		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL097282.D	18 Sep 2025 19:35		AR\AJ	Ok,M
17	PTOXICC500	PTOXICC500	PL097283.D	18 Sep 2025 19:49		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL097284.D	18 Sep 2025 20:03		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092025

Review By	Abdul	Review On	9/22/2025 9:04:56 AM
Supervise By	mohammad	Supervise On	9/23/2025 1:17:52 AM
SubDirectory	PL092025	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL097285.D	18 Sep 2025 20:16		AR\AJ	Ok,M
20	PSTDICV050	ICVPL092025	PL097286.D	19 Sep 2025 08:39		AR\AJ	Ok,M
21	PCHLORICV500	ICVPL092025	PL097287.D	19 Sep 2025 08:52		AR\AJ	Ok,M
22	PTOXICV500	ICVPL092025	PL097288.D	19 Sep 2025 09:25		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method p1092025 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL097314.D	22 Sep 2025 10:12		AR\AJ	Ok
2	I.BLK	I.BLK	PL097315.D	22 Sep 2025 11:06		AR\AJ	Ok
3	PEM	PEM	PL097316.D	22 Sep 2025 11:20		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL097317.D	22 Sep 2025 11:33		AR\AJ	Ok,M
5	PB169749BS	PB169749BS	PL097318.D	22 Sep 2025 12:31		AR\AJ	Ok,M
6	PB169772BL	PB169772BL	PL097319.D	22 Sep 2025 13:30		AR\AJ	Ok
7	PB169772BS	PB169772BS	PL097320.D	22 Sep 2025 13:44		AR\AJ	Ok,M
8	Q3152-01	EO-02-091925	PL097321.D	22 Sep 2025 14:05		AR\AJ	Ok,M
9	Q3135-01	MH-121	PL097322.D	22 Sep 2025 14:19	DCB high in 2nd column	AR\AJ	Ok,M
10	Q3151-02	1402	PL097323.D	22 Sep 2025 14:32	DCB high in 2nd column	AR\AJ	Ok
11	PB169766BL	PB169766BL	PL097324.D	22 Sep 2025 14:46		AR\AJ	Ok
12	Q3153-04	OR-500-COMP-39	PL097325.D	22 Sep 2025 17:30		AR\AJ	Ok
13	Q3153-08	OR-500-COMP-40	PL097326.D	22 Sep 2025 17:44		AR\AJ	Ok
14	I.BLK	I.BLK	PL097327.D	22 Sep 2025 17:58		AR\AJ	Ok
15	PSTDCCC050	PSTDCCC050	PL097328.D	22 Sep 2025 18:11		AR\AJ	Ok,M
16	PB169766BS	PB169766BS	PL097329.D	22 Sep 2025 18:25	Almost all compound recovery fail	AR\AJ	Not Ok
17	PB169766BSD	PB169766BSD	PL097330.D	22 Sep 2025 18:39	Almost all compound recovery fail	AR\AJ	Not Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PB169782BL	PB169782BL	PL097331.D	22 Sep 2025 18:52		AR\AJ	Ok
19	PB169782BS	PB169782BS	PL097332.D	22 Sep 2025 19:06	Almost all compound recovery fail	AR\AJ	Not Ok
20	Q3065-04	VNJ-204	PL097333.D	22 Sep 2025 19:20		AR\AJ	Ok
21	Q3092-02	291376	PL097334.D	22 Sep 2025 19:33		AR\AJ	Ok,M
22	Q3092-02MS	291376MS	PL097335.D	22 Sep 2025 19:47		AR\AJ	Ok,M
23	Q3092-02MSD	291376MSD	PL097336.D	22 Sep 2025 20:00		AR\AJ	Ok,M
24	PB169677TB	PB169677TB	PL097337.D	22 Sep 2025 20:14		AR\AJ	Ok
25	PB169754TB	PB169754TB	PL097338.D	22 Sep 2025 20:28		AR\AJ	Ok
26	I.BLK	I.BLK	PL097339.D	22 Sep 2025 20:41		AR\AJ	Ok
27	PEM	PEM	PL097340.D	22 Sep 2025 20:55		AR\AJ	Ok,M
28	PSTDCCC050	PSTDCCC050	PL097341.D	22 Sep 2025 21:08		AR\AJ	Ok,M
29	Q3092-04	279182	PL097342.D	22 Sep 2025 21:49		AR\AJ	Ok,M
30	Q3157-02	20-DEAD-1	PL097343.D	22 Sep 2025 22:03		AR\AJ	Ok
31	Q3145-01	SP-A	PL097344.D	22 Sep 2025 22:16		AR\AJ	Ok,M
32	Q3145-03	SP-B	PL097345.D	22 Sep 2025 22:30		AR\AJ	Ok,M
33	Q3153-01	OR-500-COMP-39	PL097346.D	22 Sep 2025 22:44		AR\AJ	Ok,M
34	Q3153-01MS	OR-500-COMP-39MS	PL097347.D	22 Sep 2025 22:57		AR\AJ	Ok,M
35	Q3153-01MSD	OR-500-COMP-39MSD	PL097348.D	22 Sep 2025 23:11		AR\AJ	Ok,M
36	Q3153-05	OR-500-COMP-40	PL097349.D	22 Sep 2025 23:24		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092225

Review By	Abdul	Review On	9/23/2025 8:44:42 AM
Supervise By	mohammad	Supervise On	9/24/2025 2:25:05 AM
SubDirectory	PL092225	HP Acquire Method	HP Processing Method pl092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q3155-01	COMP-1	PL097350.D	22 Sep 2025 23:38		AR\AJ	Ok,M
38	Q3155-02	COMP-2	PL097351.D	22 Sep 2025 23:52		AR\AJ	Ok,M
39	Q3155-03	COMP-3	PL097352.D	23 Sep 2025 00:05		AR\AJ	Ok,M
40	Q3157-01	20-DEAD-1	PL097353.D	23 Sep 2025 00:19		AR\AJ	Ok,M
41	I.BLK	I.BLK	PL097354.D	23 Sep 2025 00:32		AR\AJ	Ok
42	PSTDCCC050	PSTDCCC050	PL097355.D	23 Sep 2025 00:46		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092325

Review By	Abdul	Review On	9/24/2025 9:40:38 AM
Supervise By		Supervise On	
SubDirectory	PL092325	HP Acquire Method	HP Processing Method p092025 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL097356.D	23 Sep 2025 09:02		AR\AJ	Ok
2	I.BLK	I.BLK	PL097357.D	23 Sep 2025 09:16		AR\AJ	Ok
3	PEM	PEM	PL097358.D	23 Sep 2025 09:29		AR\AJ	Ok,NS
4	PSTDCCC050	PSTDCCC050	PL097359.D	23 Sep 2025 10:03		AR\AJ	Ok,NS
5	PB169766BS	PB169766BS	PL097360.D	23 Sep 2025 11:14		AR\AJ	Ok,NS
6	PB169766BSD	PB169766BSD	PL097361.D	23 Sep 2025 11:34		AR\AJ	Ok,NS
7	PB169782BS	PB169782BS	PL097362.D	23 Sep 2025 11:47		AR\AJ	Ok,NS
8	I.BLK	I.BLK	PL097363.D	23 Sep 2025 12:09		AR\AJ	Ok
9	PSTDCCC050	PSTDCCC050	PL097364.D	23 Sep 2025 12:52		AR\AJ	Ok,NS

M : Manual Integration

SOP ID : M1311-TCLP-16

SDG No : N/A

Weigh By : JP

Balance ID : WC SC-7

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pipette ID : WC

Tumbler ID : T-1 / T-2

TCLP Filter ID : 40819719

Start Prep Date : 09/18/2025 Time : 16:00

End Prep Date : 09/19/2025 Time : 08:35

Combination Ratio : 20

ZHE Cleaning Batch : N/A 10

Initial Room Temperature: 23 °C

Final Room Temperature: 22 °C

TCLP Technician Signature : 18

Supervisor By : 12

Standardized Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP114525
HCL-TCLP,1N	N/A	WP114527
HNO3-TCLP,1N	N/A	WP114528
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP86978

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 / T-2 checked,30 rpm. Q3157-02 IS USED FOR MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
09/22/25 10:00	50 100g Room	SKS. RJ / EXT
	Preparation Group	Analysis Group met dig.

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169754TB	LEB754	17	N/A	2000	N/A	N/A	N/A	4.94	1.0	T-2
Q3065-04	VNJ-204	01	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-1
Q3133-02	VNJ-238	02	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-1
Q3133-04	VNJ-222	03	100.03	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3133-06	361	04	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-1
Q3133-08	VNJ-210	05	100.03	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3136-01	34922	06	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q3136-02	34923	07	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q3136-03	34924	08	100.02	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q3136-04	34925	09	100.01	2000	N/A	N/A	N/A	4.0	1.5	T-1
Q3136-05	34926	10	100.02	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q3138-08	MOO-25-0239-0241	11	100.03	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q3150-01	MER Rd Con Ed	12	100.02	2000	N/A	N/A	N/A	7.2	1.0	T-2
Q3150-03	Mid Site Grid 5-7	13	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q3153-04	OR-500-COMP-39	14	100.02	2000	N/A	N/A	N/A	5.8	1.0	T-2
Q3153-08	OR-500-COMP-40	15	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q3157-02	20-DEAD-1	16	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB169754TB	LEB754	N/A	N/A	N/A	N/A	N/A	N/A
Q3065-04	VNJ-204	N/A	N/A	N/A	N/A	N/A	N/A
Q3133-02	VNJ-238	N/A	N/A	N/A	N/A	100	N/A
Q3133-04	VNJ-222	N/A	N/A	N/A	N/A	100	N/A
Q3133-06	361	N/A	N/A	N/A	N/A	100	N/A
Q3133-08	VNJ-210	N/A	N/A	N/A	N/A	100	N/A
Q3136-01	34922	N/A	N/A	N/A	N/A	100	N/A
Q3136-02	34923	N/A	N/A	N/A	N/A	100	N/A
Q3136-03	34924	N/A	N/A	N/A	N/A	100	N/A
Q3136-04	34925	N/A	N/A	N/A	N/A	100	N/A
Q3136-05	34926	N/A	N/A	N/A	N/A	100	N/A
Q3138-08	MOO-25-0239-0241	N/A	N/A	N/A	N/A	100	N/A
Q3150-01	MER Rd Con Ed	N/A	N/A	N/A	N/A	100	N/A
Q3150-03	Mid Site Grid 5-7	N/A	N/A	N/A	N/A	100	N/A
Q3153-04	OR-500-COMP-39	N/A	N/A	N/A	N/A	100	N/A
Q3153-08	OR-500-COMP-40	N/A	N/A	N/A	N/A	100	N/A
Q3157-02	20-DEAD-1	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2

 Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB169754TB	LEB754	N/A	96.5	N/A	N/A	#1	4.94
Q3065-04	VNJ-204	5.02	96.5	6.6	2.5	#1	4.94
Q3133-02	VNJ-238	5.02	96.5	7.0	2.5	#1	4.94
Q3133-04	VNJ-222	5.03	96.5	7.6	2.5	#1	4.94
Q3133-06	361	5.02	96.5	7.4	2.0	#1	4.94
Q3133-08	VNJ-210	5.03	96.5	7.0	2.5	#1	4.94
Q3136-01	34922	5.02	96.5	5.6	2.5	#1	4.94
Q3136-02	34923	5.01	96.5	5.8	1.5	#1	4.94
Q3136-03	34924	5.02	96.5	5.8	1.5	#1	4.94
Q3136-04	34925	5.03	96.5	5.8	2.0	#1	4.94
Q3136-05	34926	5.02	96.5	6.0	2.0	#1	4.94
Q3138-08	MOO-25-0239-0241	5.03	96.5	6.4	2.0	#1	4.94
Q3150-01	MER Rd Con Ed	5.01	96.5	8.6	3.5	#1	4.94
Q3150-03	Mid Site Grid 5-7	5.02	96.5	8.4	3.0	#1	4.94
Q3153-04	OR-500-COMP-39	5.03	96.5	7.6	2.5	#1	4.94
Q3153-08	OR-500-COMP-40	5.02	96.5	7.6	2.5	#1	4.94
Q3157-02	20-DEAD-1	5.03	96.5	8.6	3.0	#1	4.94

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tcip q3138

WorkList ID : 191956

Department : TCLP Extraction

Date : 09-19-2025 11:59:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3065-04	VNU-204	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/10/2025	1311
Q3133-02	VNU-238	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3133-04	VNU-222	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3133-06	361	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3133-08	VNU-210	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3136-01	34922	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3136-02	34923	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3136-03	34924	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3136-04	34925	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3136-05	34926	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3138-08	MOO-25-0239-0241	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3150-01	MER Rd Con Ed	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/18/2025	1311
Q3150-03	Mid Site Grid 5-7	Solid	TCLP Extraction	Cool 4 deg C	SCAL01	J23	09/18/2025	1311
Q3153-04	OR-500-COMP-39	Solid	TCLP Extraction	Cool 4 deg C	SCAL01	J23	09/18/2025	1311
Q3153-08	OR-500-COMP-40	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/19/2025	1311
Q3157-02	20-DEAD-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	D31	09/19/2025	1311
					ALWA01	D31	09/19/2025	1311

Date/Time 09-19-25 15:00

Raw Sample Received by: sd (COOP)

Raw Sample Relinquished by: sd (COOP)

Date/Time 09-19-25

Raw Sample Received by: sd (COOP)

Raw Sample Relinquished by: sd (COOP)

SOP ID: M3510C,3580A-Extraction Pesticide-17

Clean Up SOP #: Florisil

Matrix : Water

Extraction Start Date : 09/22/2025

Extraction Start Time : 10:05

Weigh By: N/A **Extraction By:** RJ **Extraction End Date :** 09/22/2025

Balance check: N/A **Filter By:** RJ **Extraction End Time :** 14:30

Balance ID: N/A **pH Meter ID:** N/A **Concentration By:** EH

pH Strip Lot#: E3953 **Hood ID:** 4,5,6,7 **Supervisor By :** ritesh

Extraction Method: ☒ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24766
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3973
Baked Na2SO4	N/A	EP2639
Hexane	N/A	E3971
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A **Envap ID:** NEVAP-02


KD Bath Temperature: N/A **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
9/22/25	RJ (Ext-106)	RJ (Ext-106)
14:35	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction Pesticide-17

Concentration Date: 09/22/2025

Sample ID	Client Sample ID	Test	g / ml	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB169677TB	PB169677TB	TCLP Pesticide	100	6	ritesh	Evelyn	10			SEP-1
PB169754TB	PB169754TB	TCLP Pesticide	100	6	ritesh	Evelyn	10			2
PB169782BL	PBLK782	TCLP Pesticide	1000	6	ritesh	Evelyn	10			3
PB169782BS	PLCS782	TCLP Pesticide	1000	6	ritesh	Evelyn	10			4
Q3065-04	VNJ-204	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		5
Q3092-02	291376	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		6
Q3092-02MS	291376MS	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		7
Q3092-02MS D	291376MSD	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		8
Q3092-04	279182	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		9
Q3153-04	OR-500-COMP-39	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		10
Q3153-08	OR-500-COMP-40	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		11
Q3157-02	20-DEAD-1	TCLP Pesticide	100	6	ritesh	Evelyn	10	A		12



RJ
9/22

* Extracts relinquished on the same date as received.

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169677TB	LEB677	11	N/A	2000	N/A	N/A	N/A	4.94	1.5	T-2
Q3082-02	WC-29	01	100.03	2000	N/A	N/A	N/A	4.0	1.0	T-1
Q3082-04	WC-30	02	100.02	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3082-06	WC-31	03	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q3082-08	WC-32	04	100.04	2000	N/A	N/A	N/A	6.2	1.0	T-1
Q3082-10	WC-33	05	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q3085-01	PAULDING-WC1	06	100.02	2000	N/A	N/A	N/A	7.6	1.5	T-1
Q3091-07	12/9C-COMP	07	100.03	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q3091-14	12/9B-COMP	08	100.02	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q3092-02	291376	09	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-1
Q3092-04	279182	10	100.02	2000	N/A	N/A	N/A	5.0	1.5	T-1

09/11/15/125
101.00

TCLP EXTRACTION LOGPAGE

PB169754

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB169754TB	LEB754	17	N/A	2000	N/A	N/A	N/A	4.94	1.0	T-2
Q3065-04	VNJ-204	01	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-1
Q3133-02	VNJ-238	02	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-1
Q3133-04	VNJ-222	03	100.03	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3133-06	361	04	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-1
Q3133-08	VNJ-210	05	100.03	2000	N/A	N/A	N/A	5.5	1.5	T-1
Q3136-01	34922	06	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q3136-02	34923	07	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-1
Q3136-03	34924	08	100.02	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q3136-04	34925	09	100.01	2000	N/A	N/A	N/A	4.0	1.5	T-1
Q3136-05	34926	10	100.02	2000	N/A	N/A	N/A	3.5	1.0	T-1
Q3138-08	MOO-25-0239-0241	11	100.03	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q3150-01	MER Rd Con Ed	12	100.02	2000	N/A	N/A	N/A	7.2	1.0	T-2
Q3150-03	Mid Site Grid 5-7	13	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q3153-04	OR-500-COMP-39	14	100.02	2000	N/A	N/A	N/A	5.8	1.0	T-2
Q3153-08	OR-500-COMP-40	15	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q3157-02	20-DEAD-1	16	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

09/19/25
701.00

Prep Standard - Chemical Standard Summary

Order ID : Q3157

Test : TCLP Pesticide

Prepbatch ID : PB169782,

Sequence ID/Qc Batch ID: pl092225,pl092325,

Standard ID :

EP2639,PP24651,PP24663,PP24738,PP24739,PP24740,PP24741,PP24742,PP24744,PP24745,PP24746,PP24747,P
P24748,PP24749,PP24750,PP24751,PP24752,PP24753,PP24754,PP24755,PP24756,PP24757,PP24758,PP24759,P
P24760,PP24761,PP24762,PP24763,PP24764,PP24766,PP24805,PP24890,

Chemical ID :

E3875,E3941,E3944,E3955,E3956,E3962,E3971,E3973,P12604,P12610,P13038,P13041,P13196,P13246,P13774,P13
786,P13788,P13789,P13862,P9053,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2639	09/12/2025	01/28/2026	Riteshkumar Patel	Extraction_SC ALE_2 (EX-SC-2)	None	Evelyn Huang 09/12/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24651	06/16/2025	12/11/2025	Abdul Mirza	None	None	Yogesh Patel 07/22/2025

FROM 1.00000ml of P13246 + 99.00000ml of E3941 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP24738	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13038 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24739	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13041 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1273	20 PPM Mirex Stock (Primary Source)	PP24740	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P9053 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24741	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13196 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24742	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13788 + 9.00000ml of E3956 = Final Quantity: 10.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP24744	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel 07/24/2025

FROM	TO	REMARKS
98.50000ml of E3956 + 0.50000ml of PP24738 + 0.50000ml of PP24740 + 0.50000ml of PP24742		= Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
80	100/100 PPB Pesticide Working Solution 2nd Source	PP24745	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel 07/24/2025

FROM	TO	REMARKS
98.50000ml of E3956 + 0.50000ml of PP24739 + 0.50000ml of PP24741 + 0.50000ml of PP24742		= Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP24746	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12604 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP24747	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12610 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP24748	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13774 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP24749	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13862 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3631	75 PPB ICAL PEST STD(RESTEK)	PP24750	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP24751	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24744 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3633	25 PPB ICAL PEST STD(RESTEK)	PP24752	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.75000ml of E3956 + 0.25000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3634	5 PPB ICAL PEST STD(RESTEK)	PP24753	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.90000ml of E3956 + 0.10000ml of PP24751 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3988	50 PPB PEST ICV STD(RESTEK)	PP24754	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24745 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
528	CHLOR 750 PPB STD	PP24755	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24746 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP24756	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
530	CHLOR 250 PPB STD	PP24757	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.75000ml of E3956 + 0.25000ml of PP24746 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3408	CHLOR 50 PPB STD	PP24758	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.90000ml of E3956 + 0.10000ml of PP24756 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP24759	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24747 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
533	TOX 750 PPB STD	PP24760	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP24761	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
535	TOX 250 PPB STD	PP24762	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.75000ml of E3956 + 0.25000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2217	TOX 100 PPB STD	PP24763	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.90000ml of E3956 + 0.10000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP24764	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of PP24749 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
79	500 PPB Pesticide Spike Solution	PP24766	07/28/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								08/19/2025

FROM 95.00000ml of E3955 + 2.50000ml of PP24739 + 2.50000ml of PP24741 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24805	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13789 + 9.00000ml of E3962 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
518	Pest/PCB I.BLK 20 PPB	PP24890	08/20/2025	02/18/2026	Abdul Mirza	None	None	Yogesh Patel
								09/17/2025

FROM 99.90000ml of E3962 + 0.10000ml of PP24805 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	12/11/2025	06/11/2025 / Rajesh	06/04/2025 / Rajesh	E3941

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	04/18/2027	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3956

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	09/15/2025 / Evelyn	09/04/2025 / Riteshkumar	E3971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	09/15/2025 / Riteshkumar	09/15/2025 / Riteshkumar	E3973

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12604

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12610

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	01/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13038

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	07/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13041

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	01/21/2026	07/21/2025 / Abdul	01/17/2024 / Abdul	P13196

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	12/17/2025	06/17/2025 / Abdul	02/09/2024 / Abdul	P13246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	01/21/2026	07/21/2025 / Abdul	05/03/2024 / Ankita	P13774

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	01/21/2026	07/21/2025 / Abdul	11/19/2024 / Ankita	P13788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	02/18/2026	08/18/2025 / yogesh	11/19/2024 / Ankita	P13789

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	01/21/2026	07/21/2025 / Abdul	12/09/2024 / Abdul	P13862

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	112018	01/21/2026	07/21/2025 / Abdul	11/01/2019 / Stephen	P9053



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001



Recd. by RI on 6/11/25

E3941

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.
 *Based on suggested storage condition.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E 3955

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**[™]



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E3956



Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/30/25

E3962

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3971

Jamie Croak
Director Quality Operations, Bioscience Production

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4

Batch No.: 25C1262005

Manufactured Date: 2025-01-15

Expiration Date: 2026-04-16

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3973

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 Lot No.: A0199099

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL Pkg Amt: > 1 mL

Expiration Date : June 30, 2027 Storage: 10°C or colder

Ship: Ambient

P130397 5
↓
P13043 1
✓
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1	µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0	µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1	µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0	µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13039
 ↓
 P13043
 5
 1
 JAW
 12/26/23

Quality Confirmation Test

Column:
 30m x .25mm x .2um
 Rtx-CLP II (cat.# 11323)

Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)


Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD

Split Vent:
 Split ratio 50:1

Inj. Vol
 1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
 Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023

Balance Serial # 1128360905

Jennifer Pollino
 Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397

**CERTIFIED WEIGHT REPORT**

Part Number:	<u>79136</u>
Lot Number:	<u>042022</u>
Description:	<u>Milrex</u>

Solvent(s):	Lot#
Acetone	81025

Expiration Date: 04/20/27

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration ($\mu\text{g/mL}$):

NIST Test ID#:

Weight(s) shown below were combined and diluted to (mL):

5E-05	Balance Uncertainty
0.006	Flask Uncertainty

Formulated By:		042022Z
	Prashant Chauhan	DATE
Reviewed By:		042022Z
	Pedro L. Rentes	DATE

Reviewed By:


Pedro L. Rentas

DATE _____

SDS Information

[illegible][illegible]

Method GC7MSD-1.M: Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B = 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 * Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
 * Standards are certified ($\pm$) 0.5% of the stated value, unless otherwise stated.

- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.

- Uncertainty Reference: Taylor, B.N., and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

$P_{13, 195}$
 $\downarrow$
 $P_{13, 199}$
 $\downarrow$
 $\textcircled{5}$
 $\downarrow$

01/17/2024
JALIN

5



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993

Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2029 **Storage:** 10°C or colder

Ship: Ambient

P 12603
↓
P 12605
[3]
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician


Date Mixed: 11-May-2023

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 12603 } (3)
↓
P 12605

7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32021 Lot No.: A0197993
Description: Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size: 2 mL Pkg Amt: > 1 mL
Expiration Date: August 31, 2029 Storage: 10°C or colder
Ship: Ambient

Handwritten: P 12606, P 12610, 5 Five, 7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. { hold 10 min.}

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 11-May-2023

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: 19161
Lot Number: 013124
Description: CIP Pesticides & PCBs Resolution Check Standard
Expiration Date: 9 components
Recommended Storage: 013129
Nominal Concentration (µg/mL): Varied
NIST Test ID#: 6UTB
Volume(s) shown below were combined and diluted to (mL): 100.0
Balance Uncertainty: 5E-05
Pipette Uncertainty: 0.021

<i>Lawrence Barry</i>	
Formulated By:	Lawrence Barry
DATE	013124
<i>Pedro L. Rentes</i>	
Reviewed By:	Pedro L. Rentes
DATE	013124

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty (mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (±) µg/mL	(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	or-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	or-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	or-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	or-rat 36300µg/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	or-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	or-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

P13243
P13244
P13245
P13246
P13247
P13248
P13249
P13250
P13251
P13252
P13253
P13254
P13255
P13256
P13257
P13258
P13259
P13260
P13261
P13262
P13263
P13264
P13265
P13266
P13267
P13268
P13269
P13270
P13271
P13272
P13273
P13274
P13275
P13276
P13277
P13278
P13279
P13280
P13281
P13282
P13283
P13284
P13285
P13286
P13287
P13288
P13289
P13290
P13291
P13292
P13293
P13294
P13295
P13296
P13297
P13298
P13299
P13300
P13301
P13302
P13303
P13304
P13305
P13306
P13307
P13308
P13309
P13310
P13311
P13312
P13313
P13314
P13315
P13316
P13317
P13318
P13319
P13320
P13321
P13322
P13323
P13324
P13325
P13326
P13327
P13328
P13329
P13330
P13331
P13332
P13333
P13334
P13335
P13336
P13337
P13338
P13339
P13340
P13341
P13342
P13343
P13344
P13345
P13346
P13347
P13348
P13349
P13350
P13351
P13352
P13353
P13354
P13355
P13356
P13357
P13358
P13359
P13360
P13361
P13362
P13363
P13364
P13365
P13366
P13367
P13368
P13369
P13370
P13371
P13372
P13373
P13374
P13375
P13376
P13377
P13378
P13379
P13380
P13381
P13382
P13383
P13384
P13385
P13386
P13387
P13388
P13389
P13390
P13391
P13392
P13393
P13394
P13395
P13396
P13397
P13398
P13399
P13400
P13401
P13402
P13403
P13404
P13405
P13406
P13407
P13408
P13409
P13410
P13411
P13412
P13413
P13414
P13415
P13416
P13417
P13418
P13419
P13420
P13421
P13422
P13423
P13424
P13425
P13426
P13427
P13428
P13429
P13430
P13431
P13432
P13433
P13434
P13435
P13436
P13437
P13438
P13439
P13440
P13441
P13442
P13443
P13444
P13445
P13446
P13447
P13448
P13449
P13450
P13451
P13452
P13453
P13454
P13455
P13456
P13457
P13458
P13459
P13460
P13461
P13462
P13463
P13464
P13465
P13466
P13467
P13468
P13469
P13470
P13471
P13472
P13473
P13474
P13475
P13476
P13477
P13478
P13479
P13480
P13481
P13482
P13483
P13484
P13485
P13486
P13487
P13488
P13489
P13490
P13491
P13492
P13493
P13494
P13495
P13496
P13497
P13498
P13499
P13500
P13501
P13502
P13503
P13504
P13505
P13506
P13507
P13508
P13509
P13510
P13511
P13512
P13513
P13514
P13515
P13516
P13517
P13518
P13519
P13520
P13521
P13522
P13523
P13524
P13525
P13526
P13527
P13528
P13529
P13530
P13531
P13532
P13533
P13534
P13535
P13536
P13537
P13538
P13539
P13540
P13541
P13542
P13543
P13544
P13545
P13546
P13547
P13548
P13549
P13550
P13551
P13552
P13553
P13554
P13555
P13556
P13557
P13558
P13559
P13560
P13561
P13562
P13563
P13564
P13565
P13566
P13567
P13568
P13569
P13570
P13571
P13572
P13573
P13574
P13575
P13576
P13577
P13578
P13579
P13580
P13581
P13582
P13583
P13584
P13585
P13586
P13587
P13588
P13589
P13590
P13591
P13592
P13593
P13594
P13595
P13596
P13597
P13598
P13599
P13600
P13601
P13602
P13603
P13604
P13605
P13606
P13607
P13608
P13609
P13610
P13611
P13612
P13613
P13614
P13615
P13616
P13617
P13618
P13619
P13620
P13621
P13622
P13623
P13624
P13625
P13626
P13627
P13628
P13629
P13630
P13631
P13632
P13633
P13634
P13635
P13636
P13637
P13638
P13639
P13640
P13641
P13642
P13643
P13644
P13645
P13646
P13647
P13648
P13649
P13650
P13651
P13652
P13653
P13654
P13655
P13656
P13657
P13658
P13659
P13660
P13661
P13662
P13663
P13664
P13665
P13666
P13667
P13668
P13669
P13670
P13671
P13672
P13673
P13674
P13675
P13676
P13677
P13678
P13679
P13680
P13681
P13682
P13683
P13684
P13685
P13686
P13687
P13688
P13689
P13690
P13691
P13692
P13693
P13694
P13695
P13696
P13697
P13698
P13699
P13700
P13701
P13702
P13703
P13704
P13705
P13706
P13707
P13708
P13709
P13710
P13711
P13712
P13713
P13714
P13715
P13716
P13717
P13718
P13719
P13720
P13721
P13722
P13723
P13724
P13725
P13726
P13727
P13728
P13729
P13730
P13731
P13732
P13733
P13734
P13735
P13736
P13737
P13738
P13739
P13740
P13741
P13742
P13743
P13744
P13745
P13746
P13747
P13748
P13749
P13750
P13751
P13752
P13753
P13754
P13755
P13756
P13757
P13758
P13759
P13760
P13761
P13762
P13763
P13764
P13765
P13766
P13767
P13768
P13769
P13770
P13771
P13772
P13773
P13774
P13775
P13776
P13777
P13778
P13779
P13780
P13781
P13782
P13783
P13784
P13785
P13786
P13787
P13788
P13789
P13790
P13791
P13792
P13793
P13794
P13795
P13796
P13797
P13798
P13799
P13800
P13801
P13802
P13803
P13804
P13805
P13806
P13807
P13808
P13809
P13810
P13811
P13812
P13813
P13814
P13815
P13816
P13817
P13818
P13819
P13820
P13821
P13822
P13823
P13824
P13825
P13826
P13827
P13828
P13829
P13830
P13831
P13832
P13833
P13834
P13835
P13836
P13837
P13838
P13839
P13840
P13841
P13842
P13843
P13844
P13845
P13846
P13847
P13848
P13849
P13850
P13851
P13852
P13853
P13854
P13855
P13856
P13857
P13858
P13859
P13860
P13861
P13862
P13863
P13864
P13865
P13866
P13867
P13868
P13869
P13870
P13871
P13872
P13873
P13874
P13875
P13876
P13877
P13878
P13879
P13880
P13881
P13882
P13883
P13884
P13885
P13886
P13887
P13888
P13889
P13890
P13891
P13892
P13893
P13894
P13895
P13896
P13897
P13898
P13899
P13900
P13901
P13902
P13903
P13904
P13905
P13906
P13907
P13908
P13909
P13910
P13911
P13912
P13913
P13914
P13915
P13916
P13917
P13918
P13919
P13920
P13921
P13922
P13923
P13924
P13925
P13926
P13927
P13928
P13929
P13930
P13931
P13932
P13933
P13934
P13935
P13936
P13937
P13938
P13939
P13940
P13941
P13942
P13943
P13944
P13945
P13946
P13947
P13948
P13949
P13950
P13951
P13952
P13953
P13954
P13955
P13956
P13957
P13958
P13959
P13960
P13961
P13962
P13963
P13964
P13965
P13966
P13967
P13968
P13969
P13970
P13971
P13972
P13973
P13974
P13975
P13976
P13977
P13978
P13979
P13980
P13981
P13982
P13983
P13984
P13985
P13986
P13987
P13988
P13989
P13990
P13991
P13992
P13993
P13994
P13995
P13996
P13997
P13998
P13999
P14000
P14001
P14002
P14003
P14004
P14005
P14006
P14007
P14008
P14009
P14010
P14011
P14012
P14013
P14014
P14015
P14016
P14017
P14018
P14019
P14020
P14021
P14022
P14023
P14024
P14025
P14026
P14027
P14028
P14029
P14030
P14031
P14032
P14033
P14034
P14035
P14036
P14037
P14038
P14039
P14040
P14041
P14042
P14043
P14044
P14045
P14046
P14047
P14048
P14049
P14050
P14051
P14052
P14053
P14054
P14055
P14056
P14057
P14058
P14059
P14060
P14061
P14062
P14063
P14064
P14065
P14066
P14067
P14068
P14069
P14070
P14071
P14072
P14073
P14074
P14075
P14076
P14077
P14078
P14079
P14080
P14081
P14082
P14083
P14084
P14085
P14086
P14087
P14088
P14089
P14090
P14091
P14092
P14093
P14094
P14095
P14096
P14097
P14098
P14099
P14100
P14101
P14102
P14103
P14104
P14105
P14106
P14107
P14108
P14109
P14110
P14111
P14112
P14113
P14114
P14115
P14116
P14117
P14118
P14119
P14120
P14121
P14122
P14123
P14124
P14125
P14126
P14127
P14128
P14129
P14130
P14131
P14132
P14133
P14134
P14135
P14136
P14137
P14138
P14139
P14140
P14141
P14142
P14143
P14144
P14145
P14146
P14147
P14148
P14149
P14150
P14151
P14152
P14153
P14154
P14155
P14156
P14157
P14158
P14159
P14160
P14161
P14162
P14163
P14164
P14165
P14166
P14167
P14168
P14169
P14170
P14171
P14172
P14173
P14174
P14175
P14176
P14177
P14178
P14179
P14180
P14181
P14182
P14183
P14184
P14185
P14186
P14187
P14188
P14189
P14190
P14191
P14192
P14193
P14194
P14195
P14196
P14197
P14198
P14199
P14200
P14201
P14202
P14203
P14204
P14205
P14206
P14207
P14208
P14209
P14210
P14211
P14212
P14213
P14214
P14215
P14216
P14217
P14218
P14219
P14220
P14221
P14222
P14223
P14224
P14225
P14226
P14227
P14228
P14229
P14230
P14231
P14232
P14233
P14234
P14235
P14236
P14237
P14238
P14239
P14240
P14241
P14242
P14243
P14244
P14245
P14246
P14247
P14248
P14249
P14250
P14251
P14252
P14253
P14254
P14255
P14256
P14257
P14258
P14259
P14260
P14261
P14262
P14263
P14264
P14265
P14266
P14267
P14268
P14269
P14270
P14271
P14272
P14273
P14274
P14275
P14276
P14277
P14278
P14279
P14280
P14281
P14282
P14283
P14284
P14285
P14286
P14287
P14288
P14289
P14290
P14291
P14292
P14293
P14294
P14295
P14296
P14297
P14298
P14299
P14300
P14301
P14302
P14303
P14304
P14305
P14306
P14307
P14308
P14309
P14310
P14311
P14312
P14313
P14314
P14315
P14316
P14317
P14318
P14319
P14320
P14321
P14322
P14323
P14324
P14325
P14326
P14327
P14328
P14329
P14330
P14331
P14332
P14333
P14334
P14335
P14336
P14337
P14338
P14339
P14340
P14341
P14342
P14343
P14344
P14345
P14346
P14347
P14348
P14349
P14350
P14351
P14352
P14353
P14354
P14355
P14356
P14357
P14358
P14359
P14360
P14361
P14362
P14363
P14364
P14365
P14366
P14367
P14368
P14369
P14370
P14371
P14372
P14373
P14374
P14375
P14376
P14377
P14378
P14379
P14380
P14381
P14382
P14383
P14384
P14385
P14386
P14387
P14388
P14389
P14390
P14391
P14392
P14393
P14394
P14395
P14396
P14397
P14398
P14399
P14400
P14401
P14402
P14403
P14404
P14405
P14406
P14407
P14408
P14409
P14410
P14411
P14412
P14413
P14414
P14415
P14416
P14417
P14418
P14419
P14420
P14421
P14422
P14423
P14424
P14425
P14426
P14427
P14428
P14429
P14430
P14431
P14432
P14433
P14434
P14435
P14436
P14437
P14438
P14439
P14440
P14441
P14442
P14443
P14444
P14445
P14446
P14447
P14448
P14449
P14450
P14451
P14452
P14453
P14454
P14455
P14456
P14457
P14458
P14459
P14460
P14461
P14462
P14463
P14464
P14465
P14466
P14467
P14468
P14469
P14470
P14471
P14472
P14473
P14474
P14475
P14476
P14477
P14478
P14479
P14480
P14481
P14482
P14483
P14484
P14485
P14486
P14487
P14488
P14489
P14490
P14491
P14492
P14493
P14494
P14495
P14496
P14497
P14498
P14499
P14500
P14501
P14502
P14503
P14504
P14505
P14506
P14507
P14508
P14509
P14510
P14511
P14512
P14513
P14514
P14515
P14516
P14517
P14518
P14519
P14520
P14521
P14522
P14523
P14524
P14525
P14526




110 Benner Circle
Belleville, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	—%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Handwritten notes and signatures:

- Signature: *[Signature]*
- Date: *11/19/24*
- Text: *P13779*
- Text: *P13773*

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

J.P.

Dakota Parson - Operations Technician I

Date Mixed: 10-Oct-2023

Balance Serial # 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397

(P)
P13773
P19779
A7
11/19/24
10-50




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024

Balance Serial # B442140311



Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }


12/9/2024



CERTIFIED WEIGHT REPORT

Part Number: **72072**
Lot Number: **112018**
Description: **n-Tetracosane-d50**

Expiration Date: **112028**
Recommended Storage: **Ambient (20 °C)**
Nominal Concentration (µg/mL): **1000**
NIST Test ID#: **2684186**

Weight(s) shown below were combined and diluted to (mL):


Solvent(s): **Methylene chloride**
Lot#: **102669**

Received by
SG on 11/11/19
p9044-p9053
5E-05 Balance Uncertainty
0.058 Flask Uncertainty


<i>Prashant Chauhan</i>	
Formulated By:	Prashant Chauhan
DATE	112018
<i>Pedro Rentas</i>	
Reviewed By:	Pedro Rentas
DATE	112018

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)
1. n-Tetracosane-d50	2072	PR-17753/09216TC1	1000	98	0.2	0.20411	0.20415	1000.2	4.2	18416-32-3 N/A
Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 250°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Candice Warren.										

TC 720720



The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
• Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Run 40, "P72072 L112018 [1000µg/mL in MeCl2]"

Run Length: 35.00 min, 20999 points at 10 points/second.

Created: Thu, Nov 22, 2018 at 7:23:18 AM.

Sampled: Sequence "112018-GC4M1", Method "GC4-M1".

Analyzed using Method "GC4-M1".

Comments

GC4-M1 Analysis by Melissa Stonier


Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5µm Film Thickness

Flow rates: Total Flow = 300 mL/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL, Hydrogen (detector) = 30 mL, Air (detector) = 360 mL

Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.

Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDAQ Channel 1.

Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 µL, Range = 3





SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Always Available Construction
ADDRESS: 20 Dead River Rd-
CITY: WARREN STATE: NJ ZIP: 07059
ATTENTION: Michael Roth
PHONE: (973) 699-4453 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Yard Soil Cleanup
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*

*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1: 2: 3: 4: 5: 6: 7: 8: 9:
VOCs TCLP VOCs GND, DRO SDOC, PEST. PCBs MET. TOB-TAL Heavy Duty Cuv. Full TCLP

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		E/F	E	E	E	E	E	E				
1.	20 DEAD-1	S.	X		9/19/25	1:45pm	11	X	X	X	X	X	X	X				
2.	20 DEAD-1	W		X	9/19/25	1:45pm	4	X	X									
3.																		
4.																		
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>Michael Roth</u>	DATE/TIME: <u>9/19/25</u>	RECEIVED BY: 1. <u>CR</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>26.1</u> °C Comments: <u>IRB #1</u>
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	

Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3157	ALWA01	Order Date : 9/19/2025 2:22:59 PM	Project Mgr :
Client Name : Always Available Construct		Project Name : Yard Soil Cleanup	Report Type : Level 1
Client Contact : Michael Roth		Receive DateTime : 9/19/2025 2:23:00 PM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Always Available Construct		Purchase Order :	Hard Copy Date :
Invoice Contact : Michael Roth			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3157-01	20-DEAD-1	Solid	09/19/2025	13:45					
					VOC-TCLVOA-10	TCL+30/TAL	8260D	10 Bus. Days	
Q3157-03	TB	Water	09/19/2025	13:45					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q3157-04	TB	Water	09/19/2025	13:45					
					TCLP VOA		8260D	10 Bus. Days	

Relinquished By : OP

Date / Time : 09/19/25 14:46

Received By : Sam

Date / Time : 09/19/25 14:46

Storage Area : VOA Refridgerator Room

Reg # 6
F2 2
Reg 4