

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Always Available Construction LLC.</u>	DataFile :	<u>PL097335.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High
Lab Sample ID:	Q3092-02MS	Client Sample ID:		291376MS							
	(Column 1)										
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92				60	152
	Heptachlor	5	0	4.50	ug/L	90				56	147
	Heptachlor epoxide	5	0	4.70	ug/L	94				77	143
	Endrin	5	0	4.40	ug/L	88				76	144
	Methoxychlor	5	0	4.20	ug/L	84				70	142
Lab Sample ID:	Q3092-02MS	Client Sample ID:		291376MS							
	(Column 2)										
	gamma-BHC (Lindane)	5	0	4.70	ug/L	94				60	152
	Heptachlor	5	0	4.50	ug/L	90				56	147
	Heptachlor epoxide	5	0	4.70	ug/L	94				77	143
	Endrin	5	0	4.50	ug/L	90				76	144
	Methoxychlor	5	0	4.00	ug/L	80				70	142