

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Always Available Construction LLC.</u>	DataFile :	<u>PL097336.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits		
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3092-02MSD	Client Sample ID:		291376MSD								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92		0		60	152	20
	Heptachlor	5	0	4.50	ug/L	90		0		56	147	20
	Heptachlor epoxide	5	0	4.70	ug/L	94		0		77	143	20
	Endrin	5	0	4.50	ug/L	90		2		76	144	20
	Methoxychlor	5	0	4.20	ug/L	84		0		70	142	20
Lab Sample ID:	Q3092-02MSD	Client Sample ID:		291376MSD								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	4.60	ug/L	92		2		60	152	20
	Heptachlor	5	0	4.40	ug/L	88		2		56	147	20
	Heptachlor epoxide	5	0	4.60	ug/L	92		2		77	143	20
	Endrin	5	0	4.50	ug/L	90		0		76	144	20
	Methoxychlor	5	0	4.00	ug/L	80		0		70	142	20