

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG082825\  
 Data File : BG064218.D  
 Acq On : 28 Aug 2025 10:51  
 Operator : RC/JU  
 Sample : SSTDICC2.5  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDICC2.5

Quant Time: Aug 28 18:39:59 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG082825.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 28 17:41:58 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min)    |
|-------------------------------|--------|------|----------|--------|-------|-------------|
| -----                         |        |      |          |        |       |             |
| Internal Standards            |        |      |          |        |       |             |
| 1) 1,4-Dichlorobenzene-d4     | 7.855  | 152  | 23937    | 20.000 | ng    | 0.00        |
| 21) Naphthalene-d8            | 10.645 | 136  | 108069   | 20.000 | ng    | 0.00        |
| 39) Acenaphthene-d10          | 14.482 | 164  | 80336    | 20.000 | ng    | 0.00        |
| 64) Phenanthrene-d10          | 17.220 | 188  | 205650   | 20.000 | ng    | 0.00        |
| 76) Chrysene-d12              | 21.450 | 240  | 221496   | 20.000 | ng    | 0.00        |
| 86) Perylene-d12              | 24.453 | 264  | 230662   | 20.000 | ng    | -0.01       |
| System Monitoring Compounds   |        |      |          |        |       |             |
| 5) 2-Fluorophenol             | 0.000  | 112  | 0d       | 0.000  | ng    |             |
| 7) Phenol-d6                  | 0.000  | 99   | 0d       | 0.000  | ng    |             |
| 23) Nitrobenzene-d5           | 0.000  | 82   | 0d       | 0.000  | ng    |             |
| 42) 2,4,6-Tribromophenol      | 0.000  | 330  | 0d       | 0.000  | ng    |             |
| 45) 2-Fluorobiphenyl          | 0.000  | 172  | 0d       | 0.000  | ng    |             |
| 79) Terphenyl-d14             | 0.000  | 244  | 0d       | 0.000  | ng    |             |
| Target Compounds              |        |      |          |        |       |             |
| 19) n-Nitroso-di-n-propyla... | 8.654  | 70   | 3017     | 2.151  | ng    | Qvalue # 73 |
| -----                         |        |      |          |        |       |             |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG082825\  
Data File : BG064218.D  
Acq On : 28 Aug 2025 10:51  
Operator : RC/JU  
Sample : SSTDICC2.5  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDICC2.5

Quant Time: Aug 28 18:39:59 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\8270-BG082825.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Aug 28 17:41:58 2025  
Response via : Initial Calibration

